

Continuing Calibration Report for SequenceID : FD062525AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049513.D

Aromatic C10-C12	5665606.000	40.000	4.429	6.172	141640.150	125184.053	-13.146
Aromatic C12-C16	9434962.000	60.000	6.173	8.800	157249.367	149149.572	-5.431
Aromatic C16-C21	13095172.000	80.000	8.801	13.090	163689.650	157103.395	-4.192
Aromatic C21-C36	27968135.000	180.000	13.091	18.510	155378.528	138149.872	-12.471
Aromatic EPH	56163875.000	360.000	4.429	18.510	156010.764	142754.403	-9.286