

Continuing Calibration Report for SequenceID : FD062525AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049522.D

Aromatic C10-C12	5218863.000	40.000	4.429	6.172	130471.575	125184.053	-4.224
Aromatic C12-C16	8755996.000	60.000	6.173	8.800	145933.267	149149.572	2.156
Aromatic C16-C21	12149667.000	80.000	8.801	13.090	151870.838	157103.395	3.331
Aromatic C21-C36	26102751.000	180.000	13.091	18.510	145015.283	138149.872	-4.970
Aromatic EPH	52227277.000	360.000	4.429	18.510	145075.769	142754.403	-1.626