

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: EARTH03

Lab Code: CHEM Case No.: Q2392 SAS No.: Q2392 SDG NO.: Q2392

Instrument ID: ECD_P Date(s) Analyzed: 06/17/2025 06/17/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.69	4.59	4.79	24659200
		2	4.78	4.68	4.88	20179000
		3	4.85	4.75	4.95	61551400
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.86	4.76	4.96	46956200
		2	5.38	5.28	5.48	22369800
		3	5.67	5.57	5.77	51597600
		4	5.83	5.73	5.93	25810400
		5	5.92	5.82	6.02	35670200
Aroclor-1262	500	1	8.07	7.97	8.17	148023000
		2	8.39	8.29	8.49	316466000
		3	8.70	8.60	8.80	213972000
		4	8.79	8.69	8.89	157593000
		5	9.44	9.34	9.54	111864000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: EARTH03

Lab Code: CHEM **Case No.:** Q2392 **SAS No.:** Q2392 **SDG NO.:** Q2392

Instrument ID: ECD_P **Date(s) Analyzed:** 06/17/2025 06/17/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.00	3.90	4.10	26872400
		2	4.09	3.99	4.19	20185400
		3	4.16	4.06	4.26	58732000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.16	4.06	4.26	45564800
		2	4.89	4.79	4.99	46259800
		3	5.06	4.96	5.16	24558000
		4	5.15	5.05	5.25	21625200
		5	5.32	5.22	5.42	22858800
Aroclor-1262	500	1	6.90	6.80	7.00	144467000
		2	7.16	7.06	7.26	122231000
		3	7.68	7.58	7.78	112703000
		4	7.75	7.65	7.85	183356000
		5	8.24	8.14	8.34	86585200