

## **Report of Analysis**

Client:	Earth Engineeri	ng Inc.			Date Collected:	06/23/25	
Project: Thomas McGove		/en			Date Received:	06/23/25	
Client Sample ID:S-1MSDLab Sample ID:Q2392-01MSDAnalytical Method:8082A		SD			SDG No.: Matrix: % Solid:	Q2392 SOIL	
Sample Wt/Vol:	30.08 Unit	ts: g			Final Vol:	10000	uL
Soil Aliquot Vol:		uL			Test:	PCB	
Extraction Type:					Injection Volume :		
GPC Factor :	1.0	PH :					
Prep Method :	SW3541B						
File ID/Qc Batch: Dilution:		Prep Date			Date Analyzed	Prep Batch ID	
PP073240.D 1		06/25/25 08:45			06/25/25 14:04	PB168607	
CAS Number	Parameter	Conc.	Qualifier	MDL		LOQ / CRQL	Units(Dry Weigh
TARGETS							
12674-11-2	Aroclor-1016	165		4.30		18.4	ug/kg
11104-28-2	Aroclor-1221	4.40	U	4.40		18.4	ug/kg
11141-16-5	Aroclor-1232	4.00	U	4.00		18.4	ug/kg
53469-21-9	Aroclor-1242	4.30	U	4.30		18.4	ug/kg
12672-29-6	Aroclor-1248	6.40	U	6.40		18.4	ug/kg
11097-69-1	Aroclor-1254	3.50	U	3.50		18.4	ug/kg
37324-23-5	Aroclor-1262	5.40	U	5.40		18.4	ug/kg
11100-14-4	Aroclor-1268	3.90	U	3.90		18.4	ug/kg
11096-82-5	Aroclor-1260	164		3.50		18.4	ug/kg
SURROGATES				32 - 144			
877-09-8	Tetrachloro-m-xylene	19.2	19.2			96%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.6		32 - 175		98%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates > 25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank
- N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

 $\mathbf{S}=\mathbf{Indicates}$  estimated value where valid five-point calibration

was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit