

Surrogate Summary

SDG No.: Q2392

Client: Earth Engineering Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP072990.D	PIBLK-PP072990.D	Tetrachloro-m-xylene	1	20	17.3	86		60	140
		Decachlorobiphenyl	1	20	18.3	91		60	140
		Tetrachloro-m-xylene	2	20	18.0	90		60	140
		Decachlorobiphenyl	2	20	17.2	86		60	140
I.BLK-PP073234.D	PIBLK-PP073234.D	Tetrachloro-m-xylene	1	20	19.6	98		60	140
		Decachlorobiphenyl	1	20	19.6	98		60	140
		Tetrachloro-m-xylene	2	20	19.2	96		60	140
		Decachlorobiphenyl	2	20	21.1	105		60	140
PB168607BL	PB168607BL	Tetrachloro-m-xylene	1	20	19.8	99		32	144
		Decachlorobiphenyl	1	20	19.6	98		32	175
		Tetrachloro-m-xylene	2	20	19.6	98		32	144
		Decachlorobiphenyl	2	20	20.5	103		32	175
PB168607BS	PB168607BS	Tetrachloro-m-xylene	1	20	20.2	101		32	144
		Decachlorobiphenyl	1	20	20.4	102		32	175
		Tetrachloro-m-xylene	2	20	19.1	96		32	144
		Decachlorobiphenyl	2	20	21.6	108		32	175
Q2392-01	S-1	Tetrachloro-m-xylene	1	20	19.1	96		32	144
		Decachlorobiphenyl	1	20	18.2	91		32	175
		Tetrachloro-m-xylene	2	20	18.2	91		32	144
		Decachlorobiphenyl	2	20	20.0	100		32	175
Q2392-01MS	S-1MS	Tetrachloro-m-xylene	1	20	20.8	104		32	144
		Decachlorobiphenyl	1	20	20.4	102		32	175
		Tetrachloro-m-xylene	2	20	19.0	95		32	144
		Decachlorobiphenyl	2	20	22.0	110		32	175
Q2392-01MSD	S-1MSD	Tetrachloro-m-xylene	1	20	19.2	96		32	144
		Decachlorobiphenyl	1	20	18.4	92		32	175
		Tetrachloro-m-xylene	2	20	17.3	86		32	144
		Decachlorobiphenyl	2	20	19.6	98		32	175
I.BLK-PP073249.D	PIBLK-PP073249.D	Tetrachloro-m-xylene	1	20	18.5	93		60	140
		Decachlorobiphenyl	1	20	18.9	95		60	140
		Tetrachloro-m-xylene	2	20	18.5	93		60	140
		Decachlorobiphenyl	2	20	20.2	101		60	140