

Surrogate Summary

SDG No.: **Q2392**

Client: **Earth Engineering Inc.**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088990.D	PIBLK-PD088990.D	Decachlorobiphenyl	1	20	18.2	91		57	171
		Tetrachloro-m-xylene	1	20	16.1	80		61	148
		Decachlorobiphenyl	2	20	18.6	93		57	171
		Tetrachloro-m-xylene	2	20	17.7	88		61	148
I.BLK-PD089133.D	PIBLK-PD089133.D	Decachlorobiphenyl	1	20	19.9	100		57	171
		Tetrachloro-m-xylene	1	20	19.2	96		61	148
		Decachlorobiphenyl	2	20	20.8	104		57	171
		Tetrachloro-m-xylene	2	20	21.0	105		61	148
PB168608BL	PB168608BL	Decachlorobiphenyl	1	20	18.8	94		20	144
		Tetrachloro-m-xylene	1	20	17.1	86		19	148
		Decachlorobiphenyl	2	20	19.3	96		20	144
		Tetrachloro-m-xylene	2	20	18.8	94		19	148
PB168608BS	PB168608BS	Decachlorobiphenyl	1	20	23.6	118		20	144
		Tetrachloro-m-xylene	1	20	21.2	106		19	148
		Decachlorobiphenyl	2	20	24.3	121		20	144
		Tetrachloro-m-xylene	2	20	23.2	116		19	148
Q2392-01	S-1	Decachlorobiphenyl	1	20	17.1	86		20	144
		Tetrachloro-m-xylene	1	20	16.6	83		19	148
		Decachlorobiphenyl	2	20	17.5	87		20	144
		Tetrachloro-m-xylene	2	20	17.2	86		19	148
Q2393-01MS	PARK AVE -1MS	Decachlorobiphenyl	1	20	16.5	82		20	144
		Tetrachloro-m-xylene	1	20	15.3	77		19	148
		Decachlorobiphenyl	2	20	15.6	78		20	144
		Tetrachloro-m-xylene	2	20	17.1	86		19	148
Q2393-01MSD	PARK AVE -1MSD	Decachlorobiphenyl	1	20	16.4	82		20	144
		Tetrachloro-m-xylene	1	20	15.5	77		19	148
		Decachlorobiphenyl	2	20	16.0	80		20	144
		Tetrachloro-m-xylene	2	20	17.2	86		19	148
I.BLK-PD089144.D	PIBLK-PD089144.D	Decachlorobiphenyl	1	20	20.7	104		57	171
		Tetrachloro-m-xylene	1	20	19.6	98		61	148
		Decachlorobiphenyl	2	20	20.8	104		57	171
		Tetrachloro-m-xylene	2	20	21.0	105		61	148