

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: EARTH03

Lab Code: CHEM Case No.: Q2392 SAS No.: Q2392 SDG No.: Q2392

Instrument ID: BNA_F Calibration Date/Time: 06/24/2025 10:16

Lab File ID: BF142811.D Init. Calib. Date(s): 06/19/2025 06/19/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 17:07 20:40

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.201	1.291		7.5	
Benzaldehyde	0.841	0.794		-5.6	
Phenol-d6	1.417	1.525		7.6	
Phenol	1.575	1.704		8.2	20.0
bis(2-Chloroethyl)ether	1.126	1.218		8.2	
2-Chlorophenol	1.296	1.398		7.9	
2-Methylphenol	0.982	1.064		8.4	
2,2-oxybis(1-Chloropropane)	1.809	1.882		4.0	
Acetophenone	0.441	0.464		5.2	
3+4-Methylphenols	1.224	1.330		8.7	
n-Nitroso-di-n-propylamine	0.824	0.878	0.050	6.6	
Nitrobenzene-d5	0.365	0.392		7.4	
Hexachloroethane	0.511	0.538		5.3	
Nitrobenzene	0.330	0.354		7.3	
Isophorone	0.585	0.629		7.5	
2-Nitrophenol	0.180	0.200		11.1	20.0
2,4-Dimethylphenol	0.311	0.334		7.4	
bis(2-Chloroethoxy)methane	0.372	0.395		6.2	
2,4-Dichlorophenol	0.282	0.303		7.4	20.0
Naphthalene	0.979	1.037		5.9	
4-Chloroaniline	0.398	0.430		8.0	
Hexachlorobutadiene	0.190	0.198		4.2	20.0
Caprolactam	0.075	0.089		18.7	
4-Chloro-3-methylphenol	0.281	0.306		8.9	20.0
2-Methylnaphthalene	0.607	0.642		5.8	
Hexachlorocyclopentadiene	0.335	0.363	0.050	8.4	
2,4,6-Trichlorophenol	0.385	0.402		4.4	20.0
2-Fluorobiphenyl	1.522	1.513		-0.6	
2,4,5-Trichlorophenol	0.405	0.431		6.4	
1,1-Biphenyl	1.580	1.627		3.0	
2-Chloronaphthalene	1.186	1.226		3.4	
2-Nitroaniline	0.332	0.365		9.9	
Dimethylphthalate	1.295	1.370		5.8	
Acenaphthylene	1.952	2.032		4.1	
2,6-Dinitrotoluene	0.284	0.307		8.1	
3-Nitroaniline	0.313	0.349		11.5	
Acenaphthene	1.191	1.256		5.5	20.0
2,4-Dinitrophenol	0.142	0.166	0.050	16.9	
4-Nitrophenol	0.214	0.246	0.050	15.0	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: EARTH03

Lab Code: CHEM Case No.: Q2392 SAS No.: Q2392 SDG No.: Q2392

Instrument ID: BNA_F Calibration Date/Time: 06/24/2025 10:16

Lab File ID: BF142811.D Init. Calib. Date(s): 06/19/2025 06/19/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 17:07 20:40

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.708	1.783		4.4	
2,4-Dinitrotoluene	0.378	0.420		11.1	
Diethylphthalate	1.269	1.361		7.3	
4-Chlorophenyl-phenylether	0.636	0.648		1.9	
Fluorene	1.334	1.373		2.9	
4-Nitroaniline	0.286	0.329		15.0	
4,6-Dinitro-2-methylphenol	0.121	0.138		14.1	
n-Nitrosodiphenylamine	0.693	0.711		2.6	20.0
2,4,6-Tribromophenol	0.206	0.220		6.8	
4-Bromophenyl-phenylether	0.228	0.240		5.3	
Hexachlorobenzene	0.253	0.264		4.3	
Atrazine	0.179	0.199		11.2	
Pentachlorophenol	0.126	0.142		12.7	20.0
Phenanthrene	1.083	1.122		3.6	
Anthracene	1.111	1.145		3.1	
Carbazole	0.959	1.009		5.2	
Di-n-butylphthalate	0.999	1.075		7.6	
Fluoranthene	1.028	1.044		1.6	20.0
Pyrene	1.875	1.902		1.4	
Terphenyl-d14	1.447	1.406		-2.8	
Butylbenzylphthalate	0.544	0.587		7.9	
3,3-Dichlorobenzidine	0.438	0.472		7.8	
Benzo (a) anthracene	1.349	1.414		4.8	
Chrysene	1.204	1.284		6.6	
Bis(2-ethylhexyl)phthalate	0.782	0.826		5.6	
Di-n-octyl phthalate	1.475	1.519		3.0	20.0
Benzo (b) fluoranthene	1.157	1.271		9.9	
Benzo (k) fluoranthene	1.083	1.127		4.1	
Benzo (a) pyrene	1.118	1.198		7.2	20.0
Indeno (1,2,3-cd) pyrene	1.523	1.575		3.4	
Dibenzo (a,h) anthracene	1.238	1.269		2.5	
Benzo (g,h,i) perylene	1.234	1.268		2.8	
1,2,4,5-Tetrachlorobenzene	0.590	0.606		2.7	
1,4-Dioxane	0.480	0.520		8.3	20.0
2,3,4,6-Tetrachlorophenol	0.326	0.344		5.5	

All other compounds must meet a minimum RRF of 0.010.