

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: EARTH03
 Lab Code: CHEM Case No.: Q2392 SAS No.: Q2392 SDG NO.: Q2392
 Lab File ID: VY022785.D Date Analyzed: 06/24/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	340335	13.34				
UPPER LIMIT	680670	13.84				
LOWER LIMIT	170168	12.84				
EPA SAMPLE NO.						
S-1A	269981	13.34				
VY0624SBL01	171409	13.34				
VY0624SBS01	311846	13.34				
VY0624SBSD01	303566	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.