



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Union Paving & Construction Company

SDG No.: Q2393

Project: 190 Park Ave Hanover Twp NJ - PO 2556-001

RunNo.: LB136256

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	06/25/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	06/25/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	06/25/2025

Initial and Continuing Calibration Verification

Client: Union Paving & Construction Company

SDG No.: Q2393

Project: 190 Park Ave Hanover Twp NJ - PO 2556-001

RunNo.: LB136259

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.096	0.099	97	90-110	06/25/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	06/25/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	06/25/2025
Sample ID: CCV3 Cyanide	mg/L	0.24	0.25	96	90-110	06/25/2025