

Initial and Continuing Calibration Blank Summary

Client: Union Paving & Construction Company

SDG No.: Q2393

Project: 190 Park Ave Hanover Twp NJ - PO 2556-001

RunNo.: LB136259

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.0016	0.0025	J	0.00096	0.005	06/25/2025
Sample ID: CCB1 Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	06/25/2025
Sample ID: CCB2 Cyanide	mg/L	0.0012	0.0025	J	0.00096	0.005	06/25/2025
Sample ID: CCB3 Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	06/25/2025