

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Union Paving & Construction Company **SDG No.:** Q2393
Contract: UNIO02 **Lab Code:** CHEM **Case No.:** Q2393 **SAS No.:** Q2393
Matrix: Solid **Level:** LOW **Client ID:** PARK AVE -6A
Sample ID: Q2393-11 **Spiked ID:** Q2393-11A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	mg/Kg	75 - 125	31.9		2.40	J	41.1	72	N	P
Beryllium	mg/Kg	75 - 125	8.54		0.96		10.3	74	N	P
Chromium	mg/Kg	75 - 125	31.4		17.9		20.6	66	N	P
Cobalt	mg/Kg	75 - 125	21.9		12.7		10.3	89		P
Copper	mg/Kg	75 - 125	40.5		29.5		15.4	71	N	P
Nickel	mg/Kg	75 - 125	44.4		21.3		25.7	90		P
Potassium	mg/Kg	75 - 125	1120		692		510	83		P
Selenium	mg/Kg	75 - 125	83.8		11.5		100	72	N	P
Silver	mg/Kg	75 - 125	3.87		1.34		3.90	65	N	P
Sodium	mg/Kg	75 - 125	466		363		150	69	N	P
Vanadium	mg/Kg	75 - 125	44.4		34.8		15.4	62	N	P
Zinc	mg/Kg	75 - 125	44.0		38.0		10.3	58	N	P