

### CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP073245.D Time Analyzed: 15:59

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.644  | 5.546     | 5.746  | 491.050            | 500.000           | -1.8 |
| Aroclor-1016-2       | 5.666  | 5.567     | 5.767  | 517.830            | 500.000           | 3.6  |
| Aroclor-1016-3       | 5.728  | 5.629     | 5.829  | 515.440            | 500.000           | 3.1  |
| Aroclor-1016-4       | 5.826  | 5.727     | 5.927  | 548.500            | 500.000           | 9.7  |
| Aroclor-1016-5       | 6.117  | 6.019     | 6.219  | 510.890            | 500.000           | 2.2  |
| Aroclor-1260-1       | 7.233  | 7.136     | 7.336  | 549.470            | 500.000           | 9.9  |
| Aroclor-1260-2       | 7.487  | 7.389     | 7.589  | 535.750            | 500.000           | 7.2  |
| Aroclor-1260-3       | 7.845  | 7.747     | 7.947  | 541.180            | 500.000           | 8.2  |
| Aroclor-1260-4       | 8.069  | 7.971     | 8.171  | 540.740            | 500.000           | 8.1  |
| Aroclor-1260-5       | 8.386  | 8.288     | 8.488  | 517.600            | 500.000           | 3.5  |
| Decachlorobiphenyl   | 10.183 | 10.087    | 10.287 | 50.590             | 50.000            | 1.2  |
| Tetrachloro-m-xylene | 4.493  | 4.394     | 4.594  | 50.850             | 50.000            | 1.7  |

### CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP073245.D Time Analyzed: 15:59

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.866 | 4.768     | 4.968 | 493.090            | 500.000           | -1.4 |
| Aroclor-1016-2       | 4.883 | 4.786     | 4.986 | 489.210            | 500.000           | -2.2 |
| Aroclor-1016-3       | 5.060 | 4.962     | 5.162 | 496.280            | 500.000           | -0.7 |
| Aroclor-1016-4       | 5.102 | 5.004     | 5.204 | 491.290            | 500.000           | -1.7 |
| Aroclor-1016-5       | 5.315 | 5.218     | 5.418 | 539.600            | 500.000           | 7.9  |
| Aroclor-1260-1       | 6.346 | 6.249     | 6.449 | 499.230            | 500.000           | -0.2 |
| Aroclor-1260-2       | 6.535 | 6.438     | 6.638 | 517.590            | 500.000           | 3.5  |
| Aroclor-1260-3       | 6.686 | 6.589     | 6.789 | 495.890            | 500.000           | -0.8 |
| Aroclor-1260-4       | 7.156 | 7.060     | 7.260 | 493.520            | 500.000           | -1.3 |
| Aroclor-1260-5       | 7.398 | 7.302     | 7.502 | 497.660            | 500.000           | -0.5 |
| Decachlorobiphenyl   | 8.792 | 8.697     | 8.897 | 52.400             | 50.000            | 4.8  |
| Tetrachloro-m-xylene | 3.786 | 3.688     | 3.888 | 49.630             | 50.000            | -0.7 |