

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP073260.D Time Analyzed: 20:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.643	5.546	5.746	467.170	500.000	-6.6
Aroclor-1016-2	5.665	5.567	5.767	485.650	500.000	-2.9
Aroclor-1016-3	5.727	5.629	5.829	486.920	500.000	-2.6
Aroclor-1016-4	5.824	5.727	5.927	512.820	500.000	2.6
Aroclor-1016-5	6.116	6.019	6.219	480.300	500.000	-3.9
Aroclor-1260-1	7.233	7.136	7.336	501.350	500.000	0.3
Aroclor-1260-2	7.486	7.389	7.589	484.800	500.000	-3.0
Aroclor-1260-3	7.844	7.747	7.947	476.370	500.000	-4.7
Aroclor-1260-4	8.068	7.971	8.171	484.000	500.000	-3.2
Aroclor-1260-5	8.386	8.288	8.488	457.790	500.000	-8.4
Decachlorobiphenyl	10.184	10.087	10.287	46.170	50.000	-7.7
Tetrachloro-m-xylene	4.492	4.394	4.594	47.050	50.000	-5.9

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP073260.D Time Analyzed: 20:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.866	4.768	4.968	483.320	500.000	-3.3
Aroclor-1016-2	4.884	4.786	4.986	493.020	500.000	-1.4
Aroclor-1016-3	5.060	4.962	5.162	492.960	500.000	-1.4
Aroclor-1016-4	5.102	5.004	5.204	489.880	500.000	-2.0
Aroclor-1016-5	5.315	5.218	5.418	542.220	500.000	8.4
Aroclor-1260-1	6.347	6.249	6.449	492.220	500.000	-1.6
Aroclor-1260-2	6.535	6.438	6.638	544.410	500.000	8.9
Aroclor-1260-3	6.687	6.589	6.789	491.150	500.000	-1.8
Aroclor-1260-4	7.157	7.060	7.260	476.840	500.000	-4.6
Aroclor-1260-5	7.398	7.302	7.502	477.200	500.000	-4.6
Decachlorobiphenyl	8.793	8.697	8.897	50.930	50.000	1.9
Tetrachloro-m-xylene	3.786	3.688	3.888	49.440	50.000	-1.1