

Surrogate Summary

SDG No.: Q2393

Client: Union Paving & Construction Company

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP072990.D	PIBLK-PP072990.D	Tetrachloro-m-xylene	1	20	17.3	86		60	140
		Decachlorobiphenyl	1	20	18.3	91		60	140
		Tetrachloro-m-xylene	2	20	18.0	90		60	140
		Decachlorobiphenyl	2	20	17.2	86		60	140
I.BLK-PP073234.D	PIBLK-PP073234.D	Tetrachloro-m-xylene	1	20	19.6	98		60	140
		Decachlorobiphenyl	1	20	19.6	98		60	140
		Tetrachloro-m-xylene	2	20	19.2	96		60	140
		Decachlorobiphenyl	2	20	21.1	105		60	140
PB168607BL	PB168607BL	Tetrachloro-m-xylene	1	20	19.8	99		32	144
		Decachlorobiphenyl	1	20	19.6	98		32	175
		Tetrachloro-m-xylene	2	20	19.6	98		32	144
		Decachlorobiphenyl	2	20	20.5	103		32	175
PB168607BS	PB168607BS	Tetrachloro-m-xylene	1	20	20.2	101		32	144
		Decachlorobiphenyl	1	20	20.4	102		32	175
		Tetrachloro-m-xylene	2	20	19.1	96		32	144
		Decachlorobiphenyl	2	20	21.6	108		32	175
Q2392-01MS	S-1MS	Tetrachloro-m-xylene	1	20	20.8	104		32	144
		Decachlorobiphenyl	1	20	20.4	102		32	175
		Tetrachloro-m-xylene	2	20	19.0	95		32	144
		Decachlorobiphenyl	2	20	22.0	110		32	175
Q2392-01MSD	S-1MSD	Tetrachloro-m-xylene	1	20	19.2	96		32	144
		Decachlorobiphenyl	1	20	18.4	92		32	175
		Tetrachloro-m-xylene	2	20	17.3	86		32	144
		Decachlorobiphenyl	2	20	19.6	98		32	175
Q2393-01	PARK AVE -1	Tetrachloro-m-xylene	1	20	20.1	101		32	144
		Decachlorobiphenyl	1	20	19.5	98		32	175
		Tetrachloro-m-xylene	2	20	19.9	99		32	144
		Decachlorobiphenyl	2	20	20.3	101		32	175
Q2393-03	PARK AVE -2	Tetrachloro-m-xylene	1	20	18.8	94		32	144
		Decachlorobiphenyl	1	20	18.1	90		32	175
		Tetrachloro-m-xylene	2	20	19.1	95		32	144
		Decachlorobiphenyl	2	20	18.0	90		32	175
Q2393-05	PARK AVE -3	Tetrachloro-m-xylene	1	20	17.0	85		32	144
		Decachlorobiphenyl	1	20	15.8	79		32	175
		Tetrachloro-m-xylene	2	20	17.5	88		32	144
		Decachlorobiphenyl	2	20	17.4	87		32	175
Q2393-07	PARK AVE -4	Tetrachloro-m-xylene	1	20	18.4	92		32	144
		Decachlorobiphenyl	1	20	16.0	80		32	175
		Tetrachloro-m-xylene	2	20	17.5	88		32	144
		Decachlorobiphenyl	2	20	17.0	85		32	175
I.BLK-PP073249.D	PIBLK-PP073249.D	Tetrachloro-m-xylene	1	20	18.5	93		60	140

Surrogate Summary

SDG No.: Q2393

Client: Union Paving & Construction Company

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP073249.D	PIBLK-PP073249.D	Decachlorobiphenyl	1	20	18.9	95		60	140
		Tetrachloro-m-xylene	2	20	18.5	93		60	140
Q2393-09	PARK AVE -5	Decachlorobiphenyl	2	20	20.2	101		60	140
		Tetrachloro-m-xylene	1	20	17.3	87		32	144
		Decachlorobiphenyl	1	20	14.4	72		32	175
		Tetrachloro-m-xylene	2	20	17.2	86		32	144
		Decachlorobiphenyl	2	20	15.4	77		32	175
		Tetrachloro-m-xylene	1	20	16.8	84		32	144
Q2393-11	PARK AVE -6	Decachlorobiphenyl	1	20	14.5	72		32	175
		Tetrachloro-m-xylene	2	20	16.6	83		32	144
		Decachlorobiphenyl	2	20	15.2	76		32	175
		Tetrachloro-m-xylene	1	20	17.7	89		60	140
I.BLK-PP073264.D	PIBLK-PP073264.D	Decachlorobiphenyl	1	20	17.2	86		60	140
		Tetrachloro-m-xylene	2	20	18.4	92		60	140
		Decachlorobiphenyl	2	20	20.1	101		60	140
		Tetrachloro-m-xylene	1	20	17.7	89		60	140