

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2393</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Union Paving & Construction Company</u>	DataFile :	<u>PP073239.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2392-01MS (Column 1)	S-IMS AR1016	180.6	0	166	ug/kg	92				55	146	
	AR1260	180.6	0	164	ug/kg	91				54	119	
Client Sample ID: Q2392-01MS (Column 2)	S-IMS AR1016	180.6	0	152	ug/kg	84				55	146	
	AR1260	180.6	0	155	ug/kg	86				54	119	