

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2393</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Union Paving & Construction Company</u>	DataFile :	<u>PP073240.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2392-01MSD (Column 1)	S-1MSD AR1016	180.3	0	165	ug/kg	92		0		55	146	15
	AR1260	180.3	0	164	ug/kg	91		0		54	119	15
Client Sample ID: Q2392-01MSD (Column 2)	S-1MSD AR1016	180.3	0	150	ug/kg	83		1		55	146	15
	AR1260	180.3	0	150	ug/kg	83		4		54	119	15