

Hit Summary Sheet
SW-846

SDG No.: Q2393

Order ID: Q2393

Client: Union Paving & Construction Company

Project ID: 190 Park Ave Hanover Twp NJ - PO 24

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : PARK AVE -3								
Q2393-05	PARK AVE -3	SOIL	4,4-DDE	0.45	JP	0.16	2.00	ug/kg
Q2393-05	PARK AVE -3	SOIL	4,4-DDD	0.70	JP	0.17	2.00	ug/kg
Q2393-05	PARK AVE -3	SOIL	4,4-DDT	0.53	J	0.16	2.00	ug/kg
Q2393-05	PARK AVE -3	SOIL	alpha-Chlordane	1.10	JP	0.14	2.00	ug/kg
Q2393-05	PARK AVE -3	SOIL	gamma-Chlordane	0.73	J	0.17	2.00	ug/kg
Total Concentration:				3.510				
Client ID : PARK AVE -4								
Q2393-07	PARK AVE -4	SOIL	4,4-DDE	0.27	J	0.16	1.90	ug/kg
Q2393-07	PARK AVE -4	SOIL	4,4-DDT	0.28	J	0.16	1.90	ug/kg
Q2393-07	PARK AVE -4	SOIL	alpha-Chlordane	0.55	JP	0.14	1.90	ug/kg
Q2393-07	PARK AVE -4	SOIL	gamma-Chlordane	0.25	J	0.17	1.90	ug/kg
Total Concentration:				1.350				
Client ID : PARK AVE -5								
Q2393-09	PARK AVE -5	SOIL	Aldrin	0.19	J	0.13	1.90	ug/kg
Q2393-09	PARK AVE -5	SOIL	4,4-DDE	0.36	J	0.15	1.90	ug/kg
Q2393-09	PARK AVE -5	SOIL	4,4-DDT	0.28	JP	0.15	1.90	ug/kg
Q2393-09	PARK AVE -5	SOIL	alpha-Chlordane	0.56	J	0.13	1.90	ug/kg
Q2393-09	PARK AVE -5	SOIL	gamma-Chlordane	0.35	J	0.16	1.90	ug/kg
Total Concentration:				1.740				
Client ID : PARK AVE -6								
Q2393-11	PARK AVE -6	SOIL	4,4-DDE	0.29	JP	0.16	2.00	ug/kg
Q2393-11	PARK AVE -6	SOIL	4,4-DDT	0.26	J	0.16	2.00	ug/kg
Q2393-11	PARK AVE -6	SOIL	alpha-Chlordane	0.61	J	0.14	2.00	ug/kg
Q2393-11	PARK AVE -6	SOIL	gamma-Chlordane	0.38	JP	0.17	2.00	ug/kg
Total Concentration:				1.540				



SAMPLE DATA

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/23/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/23/25	
Client Sample ID:	PARK AVE -1		SDG No.:	Q2393	
Lab Sample ID:	Q2393-01		Matrix:	SOIL	
Analytical Method:	8081B		% Solid:	83.3	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089140.D	1	06/25/25 08:45	06/25/25 17:54	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.16	U	0.16	2.00	ug/kg
319-85-7	beta-BHC	0.22	U	0.22	2.00	ug/kg
319-86-8	delta-BHC	0.47	U	0.47	2.00	ug/kg
58-89-9	gamma-BHC (Lindane)	0.17	U	0.17	2.00	ug/kg
76-44-8	Heptachlor	0.14	U	0.14	2.00	ug/kg
309-00-2	Aldrin	0.14	U	0.14	2.00	ug/kg
1024-57-3	Heptachlor epoxide	0.23	U	0.23	2.00	ug/kg
959-98-8	Endosulfan I	0.17	U	0.17	2.00	ug/kg
60-57-1	Dieldrin	0.17	U	0.17	2.00	ug/kg
72-55-9	4,4-DDE	0.17	U	0.17	2.00	ug/kg
72-20-8	Endrin	0.17	U	0.17	2.00	ug/kg
33213-65-9	Endosulfan II	0.35	U	0.35	2.00	ug/kg
72-54-8	4,4-DDD	0.18	U	0.18	2.00	ug/kg
1031-07-8	Endosulfan Sulfate	0.16	U	0.16	2.00	ug/kg
50-29-3	4,4-DDT	0.17	U	0.17	2.00	ug/kg
72-43-5	Methoxychlor	0.44	U	0.44	2.00	ug/kg
53494-70-5	Endrin ketone	0.23	U	0.23	2.00	ug/kg
7421-93-4	Endrin aldehyde	0.44	U	0.44	2.00	ug/kg
5103-71-9	alpha-Chlordane	0.14	U	0.14	2.00	ug/kg
5103-74-2	gamma-Chlordane	0.18	U	0.18	2.00	ug/kg
8001-35-2	Toxaphene	6.50	U	6.50	39.5	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.6		20 - 144	88%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.9		19 - 148	95%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	06/23/25
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	06/23/25
Client Sample ID:	PARK AVE -1	SDG No.:	Q2393
Lab Sample ID:	Q2393-01	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	83.3 Decanted:
Sample Wt/Vol:	30.05 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089140.D	1	06/25/25 08:45	06/25/25 17:54	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	06/23/25
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	06/23/25
Client Sample ID:	PARK AVE -2	SDG No.:	Q2393
Lab Sample ID:	Q2393-03	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	83.5 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089143.D	1	06/25/25 08:45	06/25/25 18:35	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.16	U	0.16	2.00	ug/kg
319-85-7	beta-BHC	0.22	U	0.22	2.00	ug/kg
319-86-8	delta-BHC	0.47	U	0.47	2.00	ug/kg
58-89-9	gamma-BHC (Lindane)	0.17	U	0.17	2.00	ug/kg
76-44-8	Heptachlor	0.14	U	0.14	2.00	ug/kg
309-00-2	Aldrin	0.14	U	0.14	2.00	ug/kg
1024-57-3	Heptachlor epoxide	0.23	U	0.23	2.00	ug/kg
959-98-8	Endosulfan I	0.17	U	0.17	2.00	ug/kg
60-57-1	Dieldrin	0.17	U	0.17	2.00	ug/kg
72-55-9	4,4-DDE	0.17	U	0.17	2.00	ug/kg
72-20-8	Endrin	0.17	U	0.17	2.00	ug/kg
33213-65-9	Endosulfan II	0.35	U	0.35	2.00	ug/kg
72-54-8	4,4-DDD	0.18	U	0.18	2.00	ug/kg
1031-07-8	Endosulfan Sulfate	0.16	U	0.16	2.00	ug/kg
50-29-3	4,4-DDT	0.17	U	0.17	2.00	ug/kg
72-43-5	Methoxychlor	0.44	U	0.44	2.00	ug/kg
53494-70-5	Endrin ketone	0.23	U	0.23	2.00	ug/kg
7421-93-4	Endrin aldehyde	0.44	U	0.44	2.00	ug/kg
5103-71-9	alpha-Chlordane	0.14	U	0.14	2.00	ug/kg
5103-74-2	gamma-Chlordane	0.18	U	0.18	2.00	ug/kg
8001-35-2	Toxaphene	6.50	U	6.50	39.5	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	15.8		20 - 144	79%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.5		19 - 148	87%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	06/23/25
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	06/23/25
Client Sample ID:	PARK AVE -2	SDG No.:	Q2393
Lab Sample ID:	Q2393-03	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	83.5 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089143.D	1	06/25/25 08:45	06/25/25 18:35	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/23/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/23/25	
Client Sample ID:	PARK AVE -3		SDG No.:	Q2393	
Lab Sample ID:	Q2393-05		Matrix:	SOIL	
Analytical Method:	8081B		% Solid:	86.6	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089147.D	1	06/25/25 08:45	06/25/25 19:57	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.15	U	0.15	2.00	ug/kg
319-85-7	beta-BHC	0.21	U	0.21	2.00	ug/kg
319-86-8	delta-BHC	0.45	U	0.45	2.00	ug/kg
58-89-9	gamma-BHC (Lindane)	0.16	U	0.16	2.00	ug/kg
76-44-8	Heptachlor	0.14	U	0.14	2.00	ug/kg
309-00-2	Aldrin	0.14	U	0.14	2.00	ug/kg
1024-57-3	Heptachlor epoxide	0.22	U	0.22	2.00	ug/kg
959-98-8	Endosulfan I	0.16	U	0.16	2.00	ug/kg
60-57-1	Dieldrin	0.16	U	0.16	2.00	ug/kg
72-55-9	4,4-DDE	0.45	JP	0.16	2.00	ug/kg
72-20-8	Endrin	0.16	U	0.16	2.00	ug/kg
33213-65-9	Endosulfan II	0.33	U	0.33	2.00	ug/kg
72-54-8	4,4-DDD	0.70	JP	0.17	2.00	ug/kg
1031-07-8	Endosulfan Sulfate	0.15	U	0.15	2.00	ug/kg
50-29-3	4,4-DDT	0.53	J	0.16	2.00	ug/kg
72-43-5	Methoxychlor	0.43	U	0.43	2.00	ug/kg
53494-70-5	Endrin ketone	0.22	U	0.22	2.00	ug/kg
7421-93-4	Endrin aldehyde	0.43	U	0.43	2.00	ug/kg
5103-71-9	alpha-Chlordane	1.10	JP	0.14	2.00	ug/kg
5103-74-2	gamma-Chlordane	0.73	J	0.17	2.00	ug/kg
8001-35-2	Toxaphene	6.20	U	6.20	38.0	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	14.5		20 - 144	72%	SPK: 20
877-09-8	Tetrachloro-m-xylene	15.9		19 - 148	79%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/23/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/23/25	
Client Sample ID:	PARK AVE -3		SDG No.:	Q2393	
Lab Sample ID:	Q2393-05		Matrix:	SOIL	
Analytical Method:	8081B		% Solid:	86.6	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089147.D	1	06/25/25 08:45	06/25/25 19:57	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	06/23/25
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	06/23/25
Client Sample ID:	PARK AVE -4	SDG No.:	Q2393
Lab Sample ID:	Q2393-07	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	87.7 Decanted:
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089148.D	1	06/25/25 08:45	06/25/25 20:10	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.15	U	0.15	1.90	ug/kg
319-85-7	beta-BHC	0.21	U	0.21	1.90	ug/kg
319-86-8	delta-BHC	0.44	U	0.44	1.90	ug/kg
58-89-9	gamma-BHC (Lindane)	0.16	U	0.16	1.90	ug/kg
76-44-8	Heptachlor	0.14	U	0.14	1.90	ug/kg
309-00-2	Aldrin	0.14	U	0.14	1.90	ug/kg
1024-57-3	Heptachlor epoxide	0.22	U	0.22	1.90	ug/kg
959-98-8	Endosulfan I	0.16	U	0.16	1.90	ug/kg
60-57-1	Dieldrin	0.16	U	0.16	1.90	ug/kg
72-55-9	4,4-DDE	0.27	J	0.16	1.90	ug/kg
72-20-8	Endrin	0.16	U	0.16	1.90	ug/kg
33213-65-9	Endosulfan II	0.33	U	0.33	1.90	ug/kg
72-54-8	4,4-DDD	0.17	U	0.17	1.90	ug/kg
1031-07-8	Endosulfan Sulfate	0.15	U	0.15	1.90	ug/kg
50-29-3	4,4-DDT	0.28	J	0.16	1.90	ug/kg
72-43-5	Methoxychlor	0.42	U	0.42	1.90	ug/kg
53494-70-5	Endrin ketone	0.22	U	0.22	1.90	ug/kg
7421-93-4	Endrin aldehyde	0.42	U	0.42	1.90	ug/kg
5103-71-9	alpha-Chlordane	0.55	JP	0.14	1.90	ug/kg
5103-74-2	gamma-Chlordane	0.25	J	0.17	1.90	ug/kg
8001-35-2	Toxaphene	6.20	U	6.20	37.6	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	11.1		20 - 144	56%	SPK: 20
877-09-8	Tetrachloro-m-xylene	12.5		19 - 148	63%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	06/23/25
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	06/23/25
Client Sample ID:	PARK AVE -4	SDG No.:	Q2393
Lab Sample ID:	Q2393-07	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	87.7 Decanted:
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089148.D	1	06/25/25 08:45	06/25/25 20:10	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	06/23/25
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	06/23/25
Client Sample ID:	PARK AVE -5	SDG No.:	Q2393
Lab Sample ID:	Q2393-09	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	91.3 Decanted:
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089149.D	1	06/25/25 08:45	06/25/25 20:24	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.14	U	0.14	1.90	ug/kg
319-85-7	beta-BHC	0.20	U	0.20	1.90	ug/kg
319-86-8	delta-BHC	0.43	U	0.43	1.90	ug/kg
58-89-9	gamma-BHC (Lindane)	0.15	U	0.15	1.90	ug/kg
76-44-8	Heptachlor	0.13	U	0.13	1.90	ug/kg
309-00-2	Aldrin	0.19	J	0.13	1.90	ug/kg
1024-57-3	Heptachlor epoxide	0.21	U	0.21	1.90	ug/kg
959-98-8	Endosulfan I	0.15	U	0.15	1.90	ug/kg
60-57-1	Dieldrin	0.15	U	0.15	1.90	ug/kg
72-55-9	4,4-DDE	0.36	J	0.15	1.90	ug/kg
72-20-8	Endrin	0.15	U	0.15	1.90	ug/kg
33213-65-9	Endosulfan II	0.32	U	0.32	1.90	ug/kg
72-54-8	4,4-DDD	0.16	U	0.16	1.90	ug/kg
1031-07-8	Endosulfan Sulfate	0.14	U	0.14	1.90	ug/kg
50-29-3	4,4-DDT	0.28	JP	0.15	1.90	ug/kg
72-43-5	Methoxychlor	0.41	U	0.41	1.90	ug/kg
53494-70-5	Endrin ketone	0.21	U	0.21	1.90	ug/kg
7421-93-4	Endrin aldehyde	0.41	U	0.41	1.90	ug/kg
5103-71-9	alpha-Chlordane	0.56	J	0.13	1.90	ug/kg
5103-74-2	gamma-Chlordane	0.35	J	0.16	1.90	ug/kg
8001-35-2	Toxaphene	5.90	U	5.90	36.1	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	9.08		20 - 144	45%	SPK: 20
877-09-8	Tetrachloro-m-xylene	10.7		19 - 148	54%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	06/23/25
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	06/23/25
Client Sample ID:	PARK AVE -5	SDG No.:	Q2393
Lab Sample ID:	Q2393-09	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	91.3 Decanted:
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089149.D	1	06/25/25 08:45	06/25/25 20:24	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	06/23/25
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	06/23/25
Client Sample ID:	PARK AVE -6	SDG No.:	Q2393
Lab Sample ID:	Q2393-11	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	86.8 Decanted:
Sample Wt/Vol:	30.08 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089150.D	1	06/25/25 08:45	06/25/25 20:37	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.15	U	0.15	2.00	ug/kg
319-85-7	beta-BHC	0.21	U	0.21	2.00	ug/kg
319-86-8	delta-BHC	0.45	U	0.45	2.00	ug/kg
58-89-9	gamma-BHC (Lindane)	0.16	U	0.16	2.00	ug/kg
76-44-8	Heptachlor	0.14	U	0.14	2.00	ug/kg
309-00-2	Aldrin	0.14	U	0.14	2.00	ug/kg
1024-57-3	Heptachlor epoxide	0.22	U	0.22	2.00	ug/kg
959-98-8	Endosulfan I	0.16	U	0.16	2.00	ug/kg
60-57-1	Dieldrin	0.16	U	0.16	2.00	ug/kg
72-55-9	4,4-DDE	0.29	JP	0.16	2.00	ug/kg
72-20-8	Endrin	0.16	U	0.16	2.00	ug/kg
33213-65-9	Endosulfan II	0.33	U	0.33	2.00	ug/kg
72-54-8	4,4-DDD	0.17	U	0.17	2.00	ug/kg
1031-07-8	Endosulfan Sulfate	0.15	U	0.15	2.00	ug/kg
50-29-3	4,4-DDT	0.26	J	0.16	2.00	ug/kg
72-43-5	Methoxychlor	0.43	U	0.43	2.00	ug/kg
53494-70-5	Endrin ketone	0.22	U	0.22	2.00	ug/kg
7421-93-4	Endrin aldehyde	0.43	U	0.43	2.00	ug/kg
5103-71-9	alpha-Chlordane	0.61	J	0.14	2.00	ug/kg
5103-74-2	gamma-Chlordane	0.38	JP	0.17	2.00	ug/kg
8001-35-2	Toxaphene	6.20	U	6.20	37.9	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	10.4		20 - 144	52%	SPK: 20
877-09-8	Tetrachloro-m-xylene	11.1		19 - 148	55%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/23/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/23/25	
Client Sample ID:	PARK AVE -6		SDG No.:	Q2393	
Lab Sample ID:	Q2393-11		Matrix:	SOIL	
Analytical Method:	8081B		% Solid:	86.8	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089150.D	1	06/25/25 08:45	06/25/25 20:37	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: Q2393

Client: Union Paving & Construction Company

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088990.D	PIBLK-PD088990.D	Decachlorobiphenyl	1	20	18.2	91		57	171
		Tetrachloro-m-xylene	1	20	16.1	80		61	148
		Decachlorobiphenyl	2	20	18.6	93		57	171
		Tetrachloro-m-xylene	2	20	17.7	88		61	148
I.BLK-PD089133.D	PIBLK-PD089133.D	Decachlorobiphenyl	1	20	19.9	100		57	171
		Tetrachloro-m-xylene	1	20	19.2	96		61	148
		Decachlorobiphenyl	2	20	20.8	104		57	171
		Tetrachloro-m-xylene	2	20	21.0	105		61	148
PB168608BL	PB168608BL	Decachlorobiphenyl	1	20	18.8	94		20	144
		Tetrachloro-m-xylene	1	20	17.1	86		19	148
		Decachlorobiphenyl	2	20	19.3	96		20	144
		Tetrachloro-m-xylene	2	20	18.8	94		19	148
PB168608BS	PB168608BS	Decachlorobiphenyl	1	20	23.6	118		20	144
		Tetrachloro-m-xylene	1	20	21.2	106		19	148
		Decachlorobiphenyl	2	20	24.3	121		20	144
		Tetrachloro-m-xylene	2	20	23.2	116		19	148
Q2393-01	PARK AVE -1	Decachlorobiphenyl	1	20	17.6	88		20	144
		Tetrachloro-m-xylene	1	20	17.8	89		19	148
		Decachlorobiphenyl	2	20	16.4	82		20	144
		Tetrachloro-m-xylene	2	20	18.9	95		19	148
Q2393-01MS	PARK AVE -1MS	Decachlorobiphenyl	1	20	16.5	82		20	144
		Tetrachloro-m-xylene	1	20	15.3	77		19	148
		Decachlorobiphenyl	2	20	15.6	78		20	144
		Tetrachloro-m-xylene	2	20	17.1	86		19	148
Q2393-01MSD	PARK AVE -1MSD	Decachlorobiphenyl	1	20	16.4	82		20	144
		Tetrachloro-m-xylene	1	20	15.5	77		19	148
		Decachlorobiphenyl	2	20	16.0	80		20	144
		Tetrachloro-m-xylene	2	20	17.2	86		19	148
Q2393-03	PARK AVE -2	Decachlorobiphenyl	1	20	15.8	79		20	144
		Tetrachloro-m-xylene	1	20	16.6	83		19	148
		Decachlorobiphenyl	2	20	15.8	79		20	144
		Tetrachloro-m-xylene	2	20	17.5	87		19	148
I.BLK-PD089144.D	PIBLK-PD089144.D	Decachlorobiphenyl	1	20	20.7	104		57	171
		Tetrachloro-m-xylene	1	20	19.6	98		61	148
		Decachlorobiphenyl	2	20	20.8	104		57	171
		Tetrachloro-m-xylene	2	20	21.0	105		61	148
Q2393-05	PARK AVE -3	Decachlorobiphenyl	1	20	14.5	72		20	144
		Tetrachloro-m-xylene	1	20	15.1	76		19	148
		Decachlorobiphenyl	2	20	13.2	66		20	144
		Tetrachloro-m-xylene	2	20	15.9	79		19	148
Q2393-07	PARK AVE -4	Decachlorobiphenyl	1	20	11.1	56		20	144

Surrogate Summary

SDG No.: Q2393

Client: Union Paving & Construction Company

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2393-07	PARK AVE -4	Tetrachloro-m-xylene	1	20	11.8	59		19	148
		Decachlorobiphenyl	2	20	9.59	48		20	144
Q2393-09	PARK AVE -5	Tetrachloro-m-xylene	2	20	12.5	63		19	148
		Decachlorobiphenyl	1	20	9.08	45		20	144
		Tetrachloro-m-xylene	1	20	10.2	51		19	148
		Decachlorobiphenyl	2	20	8.44	42		20	144
Q2393-11	PARK AVE -6	Tetrachloro-m-xylene	2	20	10.7	54		19	148
		Decachlorobiphenyl	1	20	10.4	52		20	144
		Tetrachloro-m-xylene	1	20	10.5	52		19	148
		Decachlorobiphenyl	2	20	8.52	43		20	144
I.BLK-PD089154.D	PIBLK-PD089154.D	Tetrachloro-m-xylene	2	20	11.1	55		19	148
		Decachlorobiphenyl	1	20	15.6	78		57	171
		Tetrachloro-m-xylene	1	20	17.6	88		61	148
		Decachlorobiphenyl	2	20	10.8	54	*	57	171
		Tetrachloro-m-xylene	2	20	21.2	106		61	148

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2393</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Union Paving & Construction Company</u>	DataFile :	<u>PD089141.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2393-01MS (Column 1)	PARK AVE -1MS											
	alpha-BHC	19.98	0	18.0	ug/kg	90				60	144	
	beta-BHC	19.98	0	17.3	ug/kg	87				54	143	
	delta-BHC	19.98	0	18.8	ug/kg	94				29	151	
	gamma-BHC (Lindane)	19.98	0	18.0	ug/kg	90				61	140	
	Heptachlor	19.98	0	17.3	ug/kg	87				63	135	
	Aldrin	19.98	0	17.8	ug/kg	89				49	139	
	Heptachlor epoxide	19.98	0	17.4	ug/kg	87				41	156	
	Endosulfan I	19.98	0	17.3	ug/kg	87				56	142	
	Dieldrin	19.98	0	17.4	ug/kg	87				47	161	
	4,4'-DDE	19.98	0	17.1	ug/kg	86				55	136	
	Endrin	19.98	0	16.6	ug/kg	83				57	139	
	Endosulfan II	19.98	0	16.6	ug/kg	83				40	163	
	4,4'-DDD	19.98	0	17.6	ug/kg	88				47	163	
	Endosulfan sulfate	19.98	0	16.7	ug/kg	84				62	139	
	4,4'-DDT	19.98	0	15.8	ug/kg	79				51	146	
	Methoxychlor	19.98	0	15.6	ug/kg	78				54	136	
	Endrin ketone	19.98	0	17.0	ug/kg	85				60	129	
	Endrin aldehyde	19.98	0	17.2	ug/kg	86				59	132	
	alpha-Chlordane	19.98	0	17.3	ug/kg	87				39	166	
	gamma-Chlordane	19.98	0	17.1	ug/kg	86				44	175	
Client Sample ID: Q2393-01MS (Column 2)	PARK AVE -1MS											
	alpha-BHC	19.98	0	18.0	ug/kg	90				60	144	
	beta-BHC	19.98	0	17.7	ug/kg	89				54	143	
	delta-BHC	19.98	0	18.0	ug/kg	90				29	151	
	gamma-BHC (Lindane)	19.98	0	18.0	ug/kg	90				61	140	
	Heptachlor	19.98	0	17.4	ug/kg	87				63	135	
	Aldrin	19.98	0	17.9	ug/kg	90				49	139	
	Heptachlor epoxide	19.98	0	17.6	ug/kg	88				41	156	
	Endosulfan I	19.98	0	17.6	ug/kg	88				56	142	
	Dieldrin	19.98	0	17.5	ug/kg	88				47	161	
	4,4'-DDE	19.98	0	17.3	ug/kg	87				55	136	
	Endrin	19.98	0	17.4	ug/kg	87				57	139	
	Endosulfan II	19.98	0	16.9	ug/kg	85				40	163	
	4,4'-DDD	19.98	0	17.3	ug/kg	87				47	163	
	Endosulfan sulfate	19.98	0	16.6	ug/kg	83				62	139	
	4,4'-DDT	19.98	0	16.0	ug/kg	80				51	146	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2393</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Union Paving & Construction Company</u>	DataFile :	<u>PD089141.D</u>

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Q2393-01MS (Column 2)	Methoxychlor	19.98	0	15.4	ug/kg	77				54	136	
	Endrin ketone	19.98	0	16.4	ug/kg	82				60	129	
	Endrin aldehyde	19.98	0	16.8	ug/kg	84				59	132	
	alpha-Chlordane	19.98	0	17.6	ug/kg	88				39	166	
	gamma-Chlordane	19.98	0	17.6	ug/kg	88				44	175	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2393</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Union Paving & Construction Company</u>	DataFile :	<u>PD089142.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2393-01MSD (Column 1)	PARK AVE -1MSD											
	alpha-BHC	19.94	0	17.9	ug/kg	90		0		60	144	20
	beta-BHC	19.94	0	17.5	ug/kg	88		1		54	143	20
	delta-BHC	19.94	0	18.8	ug/kg	94		0		29	151	20
	gamma-BHC (Lindane)	19.94	0	17.9	ug/kg	90		0		61	140	20
	Heptachlor	19.94	0	17.4	ug/kg	87		0		63	135	20
	Aldrin	19.94	0	17.8	ug/kg	89		0		49	139	20
	Heptachlor epoxide	19.94	0	17.4	ug/kg	87		0		41	156	20
	Endosulfan I	19.94	0	17.6	ug/kg	88		1		56	142	20
	Dieldrin	19.94	0	17.4	ug/kg	87		0		47	161	20
	4,4'-DDE	19.94	0	17.5	ug/kg	88		2		55	136	20
	Endrin	19.94	0	17.2	ug/kg	86		4		57	139	20
	Endosulfan II	19.94	0	16.7	ug/kg	84		1		40	163	20
	4,4'-DDD	19.94	0	17.8	ug/kg	89		1		47	163	20
	Endosulfan sulfate	19.94	0	16.9	ug/kg	85		1		62	139	20
	4,4'-DDT	19.94	0	16.1	ug/kg	81		3		51	146	20
	Methoxychlor	19.94	0	15.7	ug/kg	79		1		54	136	20
	Endrin ketone	19.94	0	17.1	ug/kg	86		1		60	129	20
	Endrin aldehyde	19.94	0	17.4	ug/kg	87		1		59	132	20
	alpha-Chlordane	19.94	0	17.5	ug/kg	88		1		39	166	20
	gamma-Chlordane	19.94	0	17.2	ug/kg	86		0		44	175	20
Client Sample ID: Q2393-01MSD (Column 2)	PARK AVE -1MSD											
	alpha-BHC	19.94	0	18.0	ug/kg	90		0		60	144	20
	beta-BHC	19.94	0	17.8	ug/kg	89		0		54	143	20
	delta-BHC	19.94	0	17.9	ug/kg	90		0		29	151	20
	gamma-BHC (Lindane)	19.94	0	18.0	ug/kg	90		0		61	140	20
	Heptachlor	19.94	0	17.4	ug/kg	87		0		63	135	20
	Aldrin	19.94	0	17.9	ug/kg	90		0		49	139	20
	Heptachlor epoxide	19.94	0	17.7	ug/kg	89		1		41	156	20
	Endosulfan I	19.94	0	17.7	ug/kg	89		1		56	142	20
	Dieldrin	19.94	0	17.5	ug/kg	88		0		47	161	20
	4,4'-DDE	19.94	0	17.4	ug/kg	87		0		55	136	20
	Endrin	19.94	0	17.0	ug/kg	85		2		57	139	20
	Endosulfan II	19.94	0	17.1	ug/kg	86		1		40	163	20
	4,4'-DDD	19.94	0	16.7	ug/kg	84		4		47	163	20
	Endosulfan sulfate	19.94	0	16.8	ug/kg	84		1		62	139	20
	4,4'-DDT	19.94	0	16.3	ug/kg	82		2		51	146	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2393</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Union Paving & Construction Company</u>	DataFile :	<u>PD089142.D</u>

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Q2393-01MSD (Column 2)	Methoxychlor	19.94	0	15.7	ug/kg	79		3		54	136	20
	Endrin ketone	19.94	0	16.7	ug/kg	84		2		60	129	20
	Endrin aldehyde	19.94	0	17.1	ug/kg	86		2		59	132	20
	alpha-Chlordane	19.94	0	17.6	ug/kg	88		0		39	166	20
	gamma-Chlordane	19.94	0	17.7	ug/kg	89		1		44	175	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2393</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Union Paving & Construction Company</u>	Datafile :	<u>PD089137.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168608BS (Column 1)	alpha-BHC	16.65	16.1	ug/kg	97				84	123	
	beta-BHC	16.65	15.5	ug/kg	93				82	123	
	delta-BHC	16.65	17.0	ug/kg	102				83	126	
	gamma-BHC (Lindane)	16.65	16.0	ug/kg	96				83	125	
	Heptachlor	16.65	15.8	ug/kg	95				83	122	
	Aldrin	16.65	16.3	ug/kg	98				82	124	
	Heptachlor epoxide	16.65	16.0	ug/kg	96				83	120	
	Endosulfan I	16.65	16.1	ug/kg	97				81	124	
	Dieldrin	16.65	16.2	ug/kg	97				85	121	
	4,4'-DDE	16.65	16.2	ug/kg	97				81	123	
	Endrin	16.65	15.2	ug/kg	91				76	130	
	Endosulfan II	16.65	16.3	ug/kg	98				80	125	
	4,4'-DDD	16.65	16.6	ug/kg	100				80	131	
	Endosulfan sulfate	16.65	16.0	ug/kg	96				81	122	
	4,4'-DDT	16.65	15.1	ug/kg	91				70	129	
	Methoxychlor	16.65	14.3	ug/kg	86				60	119	
	Endrin ketone	16.65	16.4	ug/kg	98				77	132	
	Endrin aldehyde	16.65	15.9	ug/kg	95				79	124	
	alpha-Chlordane	16.65	16.0	ug/kg	96				84	120	
	gamma-Chlordane	16.65	16.1	ug/kg	97				83	122	
	alpha-BHC	16.65	16.2	ug/kg	97				84	123	
PB168608BS (Column 2)	beta-BHC	16.65	16.0	ug/kg	96				82	123	
	delta-BHC	16.65	16.3	ug/kg	98				83	126	
	gamma-BHC (Lindane)	16.65	16.2	ug/kg	97				83	125	
	Heptachlor	16.65	15.7	ug/kg	94				83	122	
	Aldrin	16.65	16.3	ug/kg	98				82	124	
	Heptachlor epoxide	16.65	16.2	ug/kg	97				83	120	
	Endosulfan I	16.65	16.3	ug/kg	98				81	124	
	Dieldrin	16.65	16.1	ug/kg	97				85	121	
	4,4'-DDE	16.65	16.1	ug/kg	97				81	123	
	Endrin	16.65	15.0	ug/kg	90				76	130	
	Endosulfan II	16.65	16.0	ug/kg	96				80	125	
	4,4'-DDD	16.65	16.6	ug/kg	100				80	131	
	Endosulfan sulfate	16.65	15.9	ug/kg	95				81	122	
	4,4'-DDT	16.65	15.0	ug/kg	90				70	129	
	Methoxychlor	16.65	14.7	ug/kg	88				60	119	
	Endrin ketone	16.65	16.3	ug/kg	98				77	132	
	Endrin aldehyde	16.65	15.8	ug/kg	95				79	124	
	alpha-Chlordane	16.65	16.1	ug/kg	97				84	120	
	gamma-Chlordane	16.65	16.3	ug/kg	98				83	122	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168608BL

Lab Name: CHEMTECH

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393

SAS No.: Q2393 SDG NO.: Q2393

Lab Sample ID: PB168608BL

Lab File ID: PD089136.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 06/25/2025

Date Analyzed (1): 06/25/2025

Date Analyzed (2): 06/25/2025

Time Analyzed (1): 16:59

Time Analyzed (2): 16:59

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB168608BS	PB168608BS	PD089137.D	06/25/2025	06/25/2025
PARK AVE -1	Q2393-01	PD089140.D	06/25/2025	06/25/2025
PARK AVE -1MS	Q2393-01MS	PD089141.D	06/25/2025	06/25/2025
PARK AVE -1MSD	Q2393-01MSD	PD089142.D	06/25/2025	06/25/2025
PARK AVE -2	Q2393-03	PD089143.D	06/25/2025	06/25/2025
PARK AVE -3	Q2393-05	PD089147.D	06/25/2025	06/25/2025
PARK AVE -4	Q2393-07	PD089148.D	06/25/2025	06/25/2025
PARK AVE -5	Q2393-09	PD089149.D	06/25/2025	06/25/2025
PARK AVE -6	Q2393-11	PD089150.D	06/25/2025	06/25/2025

COMMENTS: _____



QC SAMPLE DATA

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	
Client Sample ID:	PB168608BL	SDG No.:	Q2393
Lab Sample ID:	PB168608BL	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	100 Decanted:
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089136.D	1	06/25/25 08:45	06/25/25 16:59	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.13	U	0.13	1.70	ug/kg
319-85-7	beta-BHC	0.18	U	0.18	1.70	ug/kg
319-86-8	delta-BHC	0.39	U	0.39	1.70	ug/kg
58-89-9	gamma-BHC (Lindane)	0.14	U	0.14	1.70	ug/kg
76-44-8	Heptachlor	0.12	U	0.12	1.70	ug/kg
309-00-2	Aldrin	0.12	U	0.12	1.70	ug/kg
1024-57-3	Heptachlor epoxide	0.19	U	0.19	1.70	ug/kg
959-98-8	Endosulfan I	0.14	U	0.14	1.70	ug/kg
60-57-1	Dieldrin	0.14	U	0.14	1.70	ug/kg
72-55-9	4,4-DDE	0.14	U	0.14	1.70	ug/kg
72-20-8	Endrin	0.14	U	0.14	1.70	ug/kg
33213-65-9	Endosulfan II	0.29	U	0.29	1.70	ug/kg
72-54-8	4,4-DDD	0.15	U	0.15	1.70	ug/kg
1031-07-8	Endosulfan Sulfate	0.13	U	0.13	1.70	ug/kg
50-29-3	4,4-DDT	0.14	U	0.14	1.70	ug/kg
72-43-5	Methoxychlor	0.37	U	0.37	1.70	ug/kg
53494-70-5	Endrin ketone	0.19	U	0.19	1.70	ug/kg
7421-93-4	Endrin aldehyde	0.37	U	0.37	1.70	ug/kg
5103-71-9	alpha-Chlordane	0.12	U	0.12	1.70	ug/kg
5103-74-2	gamma-Chlordane	0.15	U	0.15	1.70	ug/kg
8001-35-2	Toxaphene	5.40	U	5.40	33.0	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	19.3		20 - 144	96%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.8		19 - 148	94%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	
Client Sample ID:	PB168608BL		SDG No.:	Q2393
Lab Sample ID:	PB168608BL		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	100 Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL
Extraction Type:			Injection Volume :	
GPC Factor :	1.0	PH :		
Prep Method :	SW3541B			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089136.D	1	06/25/25 08:45	06/25/25 16:59	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/17/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/17/25	
Client Sample ID:	PIBLK-PD088990.D		SDG No.:	Q2393	
Lab Sample ID:	I.BLK-PD088990.D		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088990.D	1		06/17/25	PD061825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.0039	U	0.0039	0.050	ug/L
319-85-7	beta-BHC	0.0049	U	0.0049	0.050	ug/L
319-86-8	delta-BHC	0.011	U	0.011	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
309-00-2	Aldrin	0.0036	U	0.0036	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
959-98-8	Endosulfan I	0.0031	U	0.0031	0.050	ug/L
60-57-1	Dieldrin	0.0036	U	0.0036	0.050	ug/L
72-55-9	4,4-DDE	0.0037	U	0.0037	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
33213-65-9	Endosulfan II	0.0079	U	0.0079	0.050	ug/L
72-54-8	4,4-DDD	0.0071	U	0.0071	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.0037	U	0.0037	0.050	ug/L
50-29-3	4,4-DDT	0.0035	U	0.0035	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.0093	U	0.0093	0.050	ug/L
7421-93-4	Endrin aldehyde	0.011	U	0.011	0.050	ug/L
5103-71-9	alpha-Chlordane	0.0035	U	0.0035	0.050	ug/L
5103-74-2	gamma-Chlordane	0.0039	U	0.0039	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.6		57 - 171	93%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.7		61 - 148	88%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/17/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/17/25	
Client Sample ID:	PIBLK-PD088990.D		SDG No.:	Q2393	
Lab Sample ID:	I.BLK-PD088990.D		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088990.D	1		06/17/25	PD061825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/25/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/25/25	
Client Sample ID:	PIBLK-PD089133.D		SDG No.:	Q2393	
Lab Sample ID:	I.BLK-PD089133.D		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089133.D	1		06/25/25	pd062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.0039	U	0.0039	0.050	ug/L
319-85-7	beta-BHC	0.0049	U	0.0049	0.050	ug/L
319-86-8	delta-BHC	0.011	U	0.011	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
309-00-2	Aldrin	0.0036	U	0.0036	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
959-98-8	Endosulfan I	0.0031	U	0.0031	0.050	ug/L
60-57-1	Dieldrin	0.0036	U	0.0036	0.050	ug/L
72-55-9	4,4-DDE	0.0037	U	0.0037	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
33213-65-9	Endosulfan II	0.0079	U	0.0079	0.050	ug/L
72-54-8	4,4-DDD	0.0071	U	0.0071	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.0037	U	0.0037	0.050	ug/L
50-29-3	4,4-DDT	0.0035	U	0.0035	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.0093	U	0.0093	0.050	ug/L
7421-93-4	Endrin aldehyde	0.011	U	0.011	0.050	ug/L
5103-71-9	alpha-Chlordane	0.0035	U	0.0035	0.050	ug/L
5103-74-2	gamma-Chlordane	0.0039	U	0.0039	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.8		57 - 171	104%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.0		61 - 148	105%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/25/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/25/25	
Client Sample ID:	PIBLK-PD089133.D		SDG No.:	Q2393	
Lab Sample ID:	I.BLK-PD089133.D		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089133.D	1		06/25/25	pd062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/25/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/25/25	
Client Sample ID:	PIBLK-PD089144.D		SDG No.:	Q2393	
Lab Sample ID:	I.BLK-PD089144.D		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089144.D	1		06/25/25	pd062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.0039	U	0.0039	0.050	ug/L
319-85-7	beta-BHC	0.0049	U	0.0049	0.050	ug/L
319-86-8	delta-BHC	0.011	U	0.011	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
309-00-2	Aldrin	0.0036	U	0.0036	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
959-98-8	Endosulfan I	0.0031	U	0.0031	0.050	ug/L
60-57-1	Dieldrin	0.0036	U	0.0036	0.050	ug/L
72-55-9	4,4-DDE	0.0037	U	0.0037	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
33213-65-9	Endosulfan II	0.0079	U	0.0079	0.050	ug/L
72-54-8	4,4-DDD	0.0071	U	0.0071	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.0037	U	0.0037	0.050	ug/L
50-29-3	4,4-DDT	0.0035	U	0.0035	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.0093	U	0.0093	0.050	ug/L
7421-93-4	Endrin aldehyde	0.011	U	0.011	0.050	ug/L
5103-71-9	alpha-Chlordane	0.0035	U	0.0035	0.050	ug/L
5103-74-2	gamma-Chlordane	0.0039	U	0.0039	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.8		57 - 171	104%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.0		61 - 148	105%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/25/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/25/25	
Client Sample ID:	PIBLK-PD089144.D		SDG No.:	Q2393	
Lab Sample ID:	I.BLK-PD089144.D		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089144.D	1		06/25/25	pd062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/25/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/25/25	
Client Sample ID:	PIBLK-PD089154.D		SDG No.:	Q2393	
Lab Sample ID:	I.BLK-PD089154.D		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089154.D	1		06/25/25	pd062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.0039	U	0.0039	0.050	ug/L
319-85-7	beta-BHC	0.0049	U	0.0049	0.050	ug/L
319-86-8	delta-BHC	0.011	U	0.011	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
309-00-2	Aldrin	0.0036	U	0.0036	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
959-98-8	Endosulfan I	0.0031	U	0.0031	0.050	ug/L
60-57-1	Dieldrin	0.0036	U	0.0036	0.050	ug/L
72-55-9	4,4-DDE	0.0037	U	0.0037	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
33213-65-9	Endosulfan II	0.0079	U	0.0079	0.050	ug/L
72-54-8	4,4-DDD	0.0071	U	0.0071	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.0037	U	0.0037	0.050	ug/L
50-29-3	4,4-DDT	0.0035	U	0.0035	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.0093	U	0.0093	0.050	ug/L
7421-93-4	Endrin aldehyde	0.011	U	0.011	0.050	ug/L
5103-71-9	alpha-Chlordane	0.0035	U	0.0035	0.050	ug/L
5103-74-2	gamma-Chlordane	0.0039	U	0.0039	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	15.6		57 - 171	78%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.2		61 - 148	106%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/25/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/25/25	
Client Sample ID:	PIBLK-PD089154.D		SDG No.:	Q2393	
Lab Sample ID:	I.BLK-PD089154.D		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089154.D	1		06/25/25	pd062525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	
Client Sample ID:	PB168608BS	SDG No.:	Q2393
Lab Sample ID:	PB168608BS	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	100 Decanted:
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089137.D	1	06/25/25 08:45	06/25/25 17:13	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	16.2		0.13	1.70	ug/kg
319-85-7	beta-BHC	16.0		0.18	1.70	ug/kg
319-86-8	delta-BHC	17.0		0.39	1.70	ug/kg
58-89-9	gamma-BHC (Lindane)	16.2		0.14	1.70	ug/kg
76-44-8	Heptachlor	15.8		0.12	1.70	ug/kg
309-00-2	Aldrin	16.3		0.12	1.70	ug/kg
1024-57-3	Heptachlor epoxide	16.2		0.19	1.70	ug/kg
959-98-8	Endosulfan I	16.3		0.14	1.70	ug/kg
60-57-1	Dieldrin	16.2		0.14	1.70	ug/kg
72-55-9	4,4-DDE	16.2		0.14	1.70	ug/kg
72-20-8	Endrin	15.2		0.14	1.70	ug/kg
33213-65-9	Endosulfan II	16.3		0.29	1.70	ug/kg
72-54-8	4,4-DDD	16.6		0.15	1.70	ug/kg
1031-07-8	Endosulfan Sulfate	16.0		0.13	1.70	ug/kg
50-29-3	4,4-DDT	15.1		0.14	1.70	ug/kg
72-43-5	Methoxychlor	14.7		0.37	1.70	ug/kg
53494-70-5	Endrin ketone	16.4		0.19	1.70	ug/kg
7421-93-4	Endrin aldehyde	15.9		0.37	1.70	ug/kg
5103-71-9	alpha-Chlordane	16.1		0.12	1.70	ug/kg
5103-74-2	gamma-Chlordane	16.3		0.15	1.70	ug/kg
8001-35-2	Toxaphene	5.40	U	5.40	33.0	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	24.3		20 - 144	121%	SPK: 20
877-09-8	Tetrachloro-m-xylene	23.2		19 - 148	116%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	
Client Sample ID:	PB168608BS		SDG No.:	Q2393
Lab Sample ID:	PB168608BS		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	100
Sample Wt/Vol:	30.03	Units: g	Decanted:	
Soil Aliquot Vol:		uL	Final Vol:	10000 uL
Extraction Type:			Test:	Pesticide-TCL
GPC Factor :	1.0	PH :	Injection Volume :	
Prep Method :	SW3541B			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089137.D	1	06/25/25 08:45	06/25/25 17:13	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/23/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/23/25	
Client Sample ID:	PARK AVE -1MS		SDG No.:	Q2393	
Lab Sample ID:	Q2393-01MS		Matrix:	SOIL	
Analytical Method:	8081B		% Solid:	83.3	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089141.D	1	06/25/25 08:45	06/25/25 18:07	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	18.0		0.16	2.00	ug/kg
319-85-7	beta-BHC	17.7		0.22	2.00	ug/kg
319-86-8	delta-BHC	18.8		0.47	2.00	ug/kg
58-89-9	gamma-BHC (Lindane)	18.0		0.17	2.00	ug/kg
76-44-8	Heptachlor	17.4		0.14	2.00	ug/kg
309-00-2	Aldrin	17.9		0.14	2.00	ug/kg
1024-57-3	Heptachlor epoxide	17.6		0.23	2.00	ug/kg
959-98-8	Endosulfan I	17.6		0.17	2.00	ug/kg
60-57-1	Dieldrin	17.5		0.17	2.00	ug/kg
72-55-9	4,4-DDE	17.3		0.17	2.00	ug/kg
72-20-8	Endrin	17.4		0.17	2.00	ug/kg
33213-65-9	Endosulfan II	16.9		0.35	2.00	ug/kg
72-54-8	4,4-DDD	17.6		0.18	2.00	ug/kg
1031-07-8	Endosulfan Sulfate	16.7		0.16	2.00	ug/kg
50-29-3	4,4-DDT	16.0		0.17	2.00	ug/kg
72-43-5	Methoxychlor	15.6		0.44	2.00	ug/kg
53494-70-5	Endrin ketone	17.0		0.23	2.00	ug/kg
7421-93-4	Endrin aldehyde	17.2		0.44	2.00	ug/kg
5103-71-9	alpha-Chlordane	17.6		0.14	2.00	ug/kg
5103-74-2	gamma-Chlordane	17.6		0.18	2.00	ug/kg
8001-35-2	Toxaphene	6.50	U	6.50	39.6	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.5		20 - 144	82%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.1		19 - 148	86%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company	Date Collected:	06/23/25
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001	Date Received:	06/23/25
Client Sample ID:	PARK AVE -1MS	SDG No.:	Q2393
Lab Sample ID:	Q2393-01MS	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	83.3 Decanted:
Sample Wt/Vol:	30.04 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089141.D	1	06/25/25 08:45	06/25/25 18:07	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/23/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/23/25	
Client Sample ID:	PARK AVE -1MSD		SDG No.:	Q2393	
Lab Sample ID:	Q2393-01MSD		Matrix:	SOIL	
Analytical Method:	8081B		% Solid:	83.3	Decanted:
Sample Wt/Vol:	30.1	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089142.D	1	06/25/25 08:45	06/25/25 18:21	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	18.0		0.16	2.00	ug/kg
319-85-7	beta-BHC	17.8		0.22	2.00	ug/kg
319-86-8	delta-BHC	18.8		0.47	2.00	ug/kg
58-89-9	gamma-BHC (Lindane)	18.0		0.17	2.00	ug/kg
76-44-8	Heptachlor	17.4		0.14	2.00	ug/kg
309-00-2	Aldrin	17.9		0.14	2.00	ug/kg
1024-57-3	Heptachlor epoxide	17.7		0.23	2.00	ug/kg
959-98-8	Endosulfan I	17.7		0.17	2.00	ug/kg
60-57-1	Dieldrin	17.5		0.17	2.00	ug/kg
72-55-9	4,4-DDE	17.5		0.17	2.00	ug/kg
72-20-8	Endrin	17.2		0.17	2.00	ug/kg
33213-65-9	Endosulfan II	17.1		0.35	2.00	ug/kg
72-54-8	4,4-DDD	17.8		0.18	2.00	ug/kg
1031-07-8	Endosulfan Sulfate	16.9		0.16	2.00	ug/kg
50-29-3	4,4-DDT	16.3		0.17	2.00	ug/kg
72-43-5	Methoxychlor	15.7		0.44	2.00	ug/kg
53494-70-5	Endrin ketone	17.1		0.23	2.00	ug/kg
7421-93-4	Endrin aldehyde	17.4		0.44	2.00	ug/kg
5103-71-9	alpha-Chlordane	17.6		0.14	2.00	ug/kg
5103-74-2	gamma-Chlordane	17.7		0.18	2.00	ug/kg
8001-35-2	Toxaphene	6.50	U	6.50	39.5	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.4		20 - 144	82%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.2		19 - 148	86%	SPK: 20

Report of Analysis

Client:	Union Paving & Construction Company		Date Collected:	06/23/25	
Project:	190 Park Ave Hanover Twp NJ - PO 2556-001		Date Received:	06/23/25	
Client Sample ID:	PARK AVE -1MSD		SDG No.:	Q2393	
Lab Sample ID:	Q2393-01MSD		Matrix:	SOIL	
Analytical Method:	8081B		% Solid:	83.3	Decanted:
Sample Wt/Vol:	30.1	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD089142.D	1	06/25/25 08:45	06/25/25 18:21	PB168608

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



CALIBRATION SUMMARY

A
B
C
D
E
F
G
H
I
J
K
L

GC Column: ZB-MR1 ID: 0.32 (mm)

[illegible]

A
B
C
D
E
F
G
H
I
J
K
L

GC Column: ZB-MR2 ID: 0.32 (mm)

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: UNIO02

Lab Code: CHEM **Case No.:** Q2393 **SAS No.:** Q2393 **SDG NO.:** Q2393

Instrument ID: ECD_D

Calibration Date(s): 06/17/2025 06/17/2025

Calibration Times: 15:52 16:47

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PD088993.D</u>		CF 075 = <u>PD088994.D</u>			
CF 050 = <u>PD088995.D</u>		CF 025 = <u>PD088996.D</u>		CF 005 = <u>PD088997.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3557450000	3439230000	3414120000	3228670000	3520780000	3432050000	4
4,4'-DDE	4564130000	4401240000	4312380000	4092310000	4435110000	4361030000	4
4,4'-DDT	3928430000	3802000000	3763430000	3580670000	3866490000	3788210000	3
Aldrin	5504150000	5312640000	5254800000	4989300000	5504000000	5312980000	4
alpha-BHC	6295370000	6031790000	5874750000	5417910000	5525530000	5829070000	6
alpha-Chlordane	4885060000	4758900000	4734630000	4584390000	5229700000	4838540000	5
beta-BHC	2136750000	2100860000	2137680000	2144670000	2515730000	2207140000	8
Decachlorobiphenyl	3598890000	3628150000	3773200000	3897960000	4739940000	3927630000	12
delta-BHC	5546280000	5292010000	5182350000	4751190000	4892690000	5132900000	6
Dieldrin	4933650000	4797080000	4750270000	4527510000	4978810000	4797470000	4
Endosulfan I	4511200000	4405770000	4412600000	4289200000	4918470000	4507450000	5
Endosulfan II	3829260000	3982180000	4002080000	3816490000	4553000000	4036600000	7
Endosulfan sulfate	3786900000	3718160000	3741600000	3667440000	4271990000	3837220000	6
Endrin	4238140000	4109540000	4096430000	3878120000	4305700000	4125590000	4
Endrin aldehyde	2964260000	2935670000	2964230000	2966830000	3488720000	3063940000	8
Endrin ketone	4050100000	3970810000	4004900000	3895960000	4443590000	4073070000	5
gamma-BHC (Lindane)	5900830000	5686510000	5566590000	5221690000	5532800000	5581680000	4
gamma-Chlordane	4869330000	4801300000	4716110000	4508810000	5045020000	4788110000	4
Heptachlor	5621840000	5429540000	5358570000	5116310000	5733550000	5451960000	4
Heptachlor epoxide	4794850000	4673610000	4749650000	4554690000	5346870000	4823940000	6
Methoxychlor	1926110000	1918890000	1971480000	1982080000	2304580000	2020630000	8
Tetrachloro-m-xylene	2812300000	2762040000	2767280000	2751820000	3172230000	2853130000	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: UNIO02

Lab Code: CHEM **Case No.:** Q2393 **SAS No.:** Q2393 **SDG NO.:** Q2393

Instrument ID: ECD_D

Calibration Date(s): 06/17/2025 06/17/2025

Calibration Times: 15:52 16:47

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PD088993.D</u>		CF 075 = <u>PD088994.D</u>			
CF 050 = <u>PD088995.D</u>		CF 025 = <u>PD088996.D</u>		CF 005 = <u>PD088997.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	17754100000	17899700000	18191100000	18805800000	22996200000	19129400000	11
4,4'-DDE	21343500000	21440200000	21823600000	22564100000	27409000000	22916100000	11
4,4'-DDT	19391000000	19327800000	19594100000	20005200000	23078500000	20279300000	8
Aldrin	22719200000	22747700000	23159400000	23852800000	28607200000	24217300000	10
alpha-BHC	25649000000	25438700000	25849200000	26153100000	30895500000	26797100000	9
alpha-Chlordane	21191300000	21145300000	21570900000	22517200000	28059300000	22896800000	13
beta-BHC	9903630000	9918550000	10198200000	10624100000	13086800000	10746200000	12
Decachlorobiphenyl	17946500000	17939700000	18428600000	19427500000	25107500000	19770000000	15
delta-BHC	23614400000	23533700000	23918700000	24298300000	29003000000	24873600000	9
Dieldrin	21438700000	21622000000	22120600000	22941800000	27919200000	23208500000	12
Endosulfan I	19170300000	19201000000	19315500000	20752900000	25772600000	20842500000	14
Endosulfan II	18473200000	18573800000	19087100000	20002300000	24846800000	20196600000	13
Endosulfan sulfate	17925600000	18043500000	18492500000	19409200000	24234500000	19621100000	13
Endrin	20535400000	19906600000	20644700000	21276500000	26206500000	21713900000	12
Endrin aldehyde	13773600000	13920200000	14379400000	15195500000	19287900000	15311300000	15
Endrin ketone	19587400000	19824900000	20446500000	21535300000	26614600000	21601700000	13
gamma-BHC (Lindane)	23622800000	23442000000	23749900000	24221500000	28748400000	24756900000	9
gamma-Chlordane	22229100000	22022500000	22519500000	23493700000	29348800000	23922700000	13
Heptachlor	23291600000	23365200000	23922200000	24716000000	29896800000	25038400000	11
Heptachlor epoxide	20080600000	20251900000	20827300000	21695300000	26635200000	21898100000	12
Methoxychlor	9712450000	9897880000	10259100000	10830900000	13140100000	10768100000	13
Tetrachloro-m-xylene	15879800000	15866400000	15978000000	16730900000	20310700000	16953200000	11

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

Instrument ID: ECD_D Date(s) Analyzed: 06/17/2025 06/17/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	33601800
		2	6.44	6.34	6.54	47528600
		3	7.15	7.05	7.25	87292700
		4	7.56	7.46	7.66	111357000
		5	7.93	7.83	8.03	63112600

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

Instrument ID: ECD_D Date(s) Analyzed: 06/17/2025 06/17/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.47	5.37	5.57	158160000
		2	5.65	5.55	5.75	108820000
		3	6.76	6.66	6.86	511856000
		4	7.20	7.10	7.30	350145000
		5	7.33	7.23	7.43	253920000

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

Continuing Calib Date: 06/25/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 16:12 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
alpha-BHC	4.00	4.00	3.90	4.10	0.00
beta-BHC	4.52	4.52	4.42	4.62	0.01
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Aldrin	5.27	5.27	5.17	5.37	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endosulfan I	6.07	6.08	5.98	6.18	0.01
Dieldrin	6.35	6.35	6.25	6.45	0.00
4,4'-DDE	6.20	6.20	6.10	6.30	0.00
Endrin	6.58	6.58	6.48	6.68	0.01
Endosulfan II	6.79	6.79	6.69	6.89	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.01
Endosulfan sulfate	7.15	7.15	7.05	7.25	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.63	7.53	7.73	0.00
Endrin aldehyde	6.92	6.92	6.82	7.02	0.01
alpha-Chlordane	6.03	6.03	5.93	6.13	0.00
gamma-Chlordane	5.95	5.95	5.85	6.05	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

Continuing Calib Date: 06/25/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 16:12 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.07	8.07	7.97	8.17	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.03	4.03	3.93	4.13	0.00
delta-BHC	4.26	4.26	4.16	4.36	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.08	4.08	3.98	4.18	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Heptachlor epoxide	4.87	4.87	4.77	4.97	0.00
Endosulfan I	5.25	5.25	5.15	5.35	0.00
Dieldrin	5.51	5.51	5.41	5.61	0.00
4,4'-DDE	5.38	5.38	5.28	5.48	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Endosulfan II	6.08	6.08	5.98	6.18	0.00
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
Endosulfan sulfate	6.48	6.48	6.38	6.58	0.00
4,4'-DDT	6.18	6.18	6.08	6.28	0.00
Methoxychlor	6.75	6.76	6.66	6.86	0.01
Endrin ketone	6.99	6.99	6.89	7.09	0.00
Endrin aldehyde	6.26	6.26	6.16	6.36	0.00
alpha-Chlordane	5.19	5.19	5.09	5.29	0.00
gamma-Chlordane	5.13	5.13	5.03	5.23	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/25/2025

Lab Sample No.: PSTDCCC050 Data File : PD089134.D Time Analyzed: 16:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.605	6.805	52.370	50.000	4.7
4,4'-DDE	6.195	6.096	6.296	50.910	50.000	1.8
4,4'-DDT	7.021	6.921	7.121	48.180	50.000	-3.6
Aldrin	5.271	5.171	5.371	52.530	50.000	5.1
alpha-BHC	3.999	3.899	4.099	53.570	50.000	7.1
alpha-Chlordane	6.027	5.927	6.127	51.430	50.000	2.9
beta-BHC	4.515	4.415	4.615	51.190	50.000	2.4
Decachlorobiphenyl	9.073	8.972	9.172	48.720	50.000	-2.6
delta-BHC	4.764	4.664	4.864	55.600	50.000	11.2
Dieldrin	6.347	6.247	6.447	51.740	50.000	3.5
Endosulfan I	6.074	5.975	6.175	51.320	50.000	2.6
Endosulfan II	6.786	6.686	6.886	51.720	50.000	3.4
Endosulfan sulfate	7.149	7.049	7.249	50.030	50.000	0.1
Endrin	6.575	6.475	6.675	50.280	50.000	0.6
Endrin aldehyde	6.915	6.815	7.015	49.430	50.000	-1.1
Endrin ketone	7.630	7.530	7.730	50.650	50.000	1.3
gamma-BHC (Lindane)	4.330	4.230	4.430	53.110	50.000	6.2
gamma-Chlordane	5.945	5.846	6.046	51.930	50.000	3.9
Heptachlor	4.929	4.829	5.029	51.830	50.000	3.7
Heptachlor epoxide	5.691	5.591	5.791	50.860	50.000	1.7
Methoxychlor	7.493	7.393	7.593	45.860	50.000	-8.3
Tetrachloro-m-xylene	3.550	3.450	3.650	51.690	50.000	3.4

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/25/2025

Lab Sample No.: PSTDCCC050 Data File : PD089134.D Time Analyzed: 16:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	50.660	50.000	1.3
4,4'-DDE	5.375	5.275	5.475	50.160	50.000	0.3
4,4'-DDT	6.183	6.084	6.284	47.880	50.000	-4.2
Aldrin	4.368	4.269	4.469	51.110	50.000	2.2
alpha-BHC	3.393	3.293	3.493	51.690	50.000	3.4
alpha-Chlordane	5.190	5.091	5.291	50.080	50.000	0.2
beta-BHC	4.025	3.926	4.126	50.170	50.000	0.3
Decachlorobiphenyl	8.072	7.972	8.172	49.080	50.000	-1.8
delta-BHC	4.261	4.162	4.362	51.300	50.000	2.6
Dieldrin	5.512	5.413	5.613	50.380	50.000	0.8
Endosulfan I	5.246	5.147	5.347	48.960	50.000	-2.1
Endosulfan II	6.080	5.981	6.181	49.710	50.000	-0.6
Endosulfan sulfate	6.481	6.383	6.583	49.190	50.000	-1.6
Endrin	5.789	5.689	5.889	49.670	50.000	-0.7
Endrin aldehyde	6.258	6.159	6.359	48.480	50.000	-3.0
Endrin ketone	6.991	6.892	7.092	49.700	50.000	-0.6
gamma-BHC (Lindane)	3.729	3.630	3.830	51.520	50.000	3.0
gamma-Chlordane	5.125	5.026	5.226	50.180	50.000	0.4
Heptachlor	4.083	3.983	4.183	49.870	50.000	-0.3
Heptachlor epoxide	4.872	4.773	4.973	50.580	50.000	1.2
Methoxychlor	6.754	6.655	6.855	46.450	50.000	-7.1
Tetrachloro-m-xylene	2.881	2.780	2.980	51.570	50.000	3.1

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

Continuing Calib Date: 06/25/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 19:15 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
alpha-BHC	4.00	4.00	3.90	4.10	0.00
beta-BHC	4.52	4.52	4.42	4.62	0.01
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Aldrin	5.27	5.27	5.17	5.37	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endosulfan I	6.07	6.08	5.98	6.18	0.01
Dieldrin	6.35	6.35	6.25	6.45	0.00
4,4'-DDE	6.20	6.20	6.10	6.30	0.00
Endrin	6.57	6.58	6.48	6.68	0.01
Endosulfan II	6.79	6.79	6.69	6.89	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.01
Endosulfan sulfate	7.15	7.15	7.05	7.25	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.63	7.53	7.73	0.00
Endrin aldehyde	6.92	6.92	6.82	7.02	0.01
alpha-Chlordane	6.03	6.03	5.93	6.13	0.00
gamma-Chlordane	5.95	5.95	5.85	6.05	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

Continuing Calib Date: 06/25/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 19:15 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.07	8.07	7.97	8.17	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.03	4.03	3.93	4.13	0.00
delta-BHC	4.26	4.26	4.16	4.36	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.08	4.08	3.98	4.18	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Heptachlor epoxide	4.87	4.87	4.77	4.97	0.00
Endosulfan I	5.25	5.25	5.15	5.35	0.00
Dieldrin	5.51	5.51	5.41	5.61	0.00
4,4'-DDE	5.38	5.38	5.28	5.48	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Endosulfan II	6.08	6.08	5.98	6.18	0.00
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
Endosulfan sulfate	6.48	6.48	6.38	6.58	0.00
4,4'-DDT	6.18	6.18	6.08	6.28	0.00
Methoxychlor	6.75	6.76	6.66	6.86	0.01
Endrin ketone	6.99	6.99	6.89	7.09	0.00
Endrin aldehyde	6.26	6.26	6.16	6.36	0.00
alpha-Chlordane	5.19	5.19	5.09	5.29	0.00
gamma-Chlordane	5.13	5.13	5.03	5.23	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/25/2025

Lab Sample No.: PSTDCCC050 Data File : PD089146.D Time Analyzed: 19:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.605	6.805	52.530	50.000	5.1
4,4'-DDE	6.196	6.096	6.296	51.170	50.000	2.3
4,4'-DDT	7.021	6.921	7.121	48.420	50.000	-3.2
Aldrin	5.271	5.171	5.371	52.830	50.000	5.7
alpha-BHC	3.999	3.899	4.099	54.380	50.000	8.8
alpha-Chlordane	6.026	5.927	6.127	51.910	50.000	3.8
beta-BHC	4.515	4.415	4.615	51.800	50.000	3.6
Decachlorobiphenyl	9.073	8.972	9.172	48.940	50.000	-2.1
delta-BHC	4.764	4.664	4.864	56.290	50.000	12.6
Dieldrin	6.346	6.247	6.447	51.830	50.000	3.7
Endosulfan I	6.074	5.975	6.175	51.790	50.000	3.6
Endosulfan II	6.786	6.686	6.886	51.980	50.000	4.0
Endosulfan sulfate	7.149	7.049	7.249	50.040	50.000	0.1
Endrin	6.574	6.475	6.675	50.290	50.000	0.6
Endrin aldehyde	6.915	6.815	7.015	49.890	50.000	-0.2
Endrin ketone	7.630	7.530	7.730	50.970	50.000	1.9
gamma-BHC (Lindane)	4.330	4.230	4.430	53.840	50.000	7.7
gamma-Chlordane	5.945	5.846	6.046	52.630	50.000	5.3
Heptachlor	4.929	4.829	5.029	52.360	50.000	4.7
Heptachlor epoxide	5.690	5.591	5.791	51.690	50.000	3.4
Methoxychlor	7.493	7.393	7.593	45.660	50.000	-8.7
Tetrachloro-m-xylene	3.550	3.450	3.650	52.350	50.000	4.7

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/25/2025

Lab Sample No.: PSTDCCC050 Data File : PD089146.D Time Analyzed: 19:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	51.340	50.000	2.7
4,4'-DDE	5.375	5.275	5.475	50.750	50.000	1.5
4,4'-DDT	6.183	6.084	6.284	48.290	50.000	-3.4
Aldrin	4.369	4.269	4.469	51.750	50.000	3.5
alpha-BHC	3.393	3.293	3.493	52.130	50.000	4.3
alpha-Chlordane	5.190	5.091	5.291	50.620	50.000	1.2
beta-BHC	4.025	3.926	4.126	50.200	50.000	0.4
Decachlorobiphenyl	8.071	7.972	8.172	48.540	50.000	-2.9
delta-BHC	4.262	4.162	4.362	51.710	50.000	3.4
Dieldrin	5.513	5.413	5.613	50.760	50.000	1.5
Endosulfan I	5.246	5.147	5.347	48.710	50.000	-2.6
Endosulfan II	6.080	5.981	6.181	50.180	50.000	0.4
Endosulfan sulfate	6.482	6.383	6.583	49.170	50.000	-1.7
Endrin	5.789	5.689	5.889	50.400	50.000	0.8
Endrin aldehyde	6.258	6.159	6.359	48.810	50.000	-2.4
Endrin ketone	6.991	6.892	7.092	49.570	50.000	-0.9
gamma-BHC (Lindane)	3.729	3.630	3.830	51.880	50.000	3.8
gamma-Chlordane	5.125	5.026	5.226	50.860	50.000	1.7
Heptachlor	4.083	3.983	4.183	50.530	50.000	1.1
Heptachlor epoxide	4.872	4.773	4.973	51.260	50.000	2.5
Methoxychlor	6.754	6.655	6.855	46.630	50.000	-6.7
Tetrachloro-m-xylene	2.881	2.780	2.980	51.950	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

Continuing Calib Date: 06/25/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 22:40 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
alpha-BHC	4.00	4.00	3.90	4.10	0.00
beta-BHC	4.52	4.52	4.42	4.62	0.01
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Aldrin	5.27	5.27	5.17	5.37	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endosulfan I	6.07	6.08	5.98	6.18	0.01
Dieldrin	6.35	6.35	6.25	6.45	0.00
4,4'-DDE	6.20	6.20	6.10	6.30	0.00
Endrin	6.57	6.58	6.48	6.68	0.01
Endosulfan II	6.79	6.79	6.69	6.89	0.01
4,4'-DDD	6.70	6.71	6.61	6.81	0.01
Endosulfan sulfate	7.15	7.15	7.05	7.25	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.63	7.53	7.73	0.00
Endrin aldehyde	6.91	6.92	6.82	7.02	0.01
alpha-Chlordane	6.03	6.03	5.93	6.13	0.00
gamma-Chlordane	5.95	5.95	5.85	6.05	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

Continuing Calib Date: 06/25/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 22:40 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.07	8.07	7.97	8.17	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.03	4.03	3.93	4.13	0.00
delta-BHC	4.26	4.26	4.16	4.36	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.08	4.08	3.98	4.18	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Heptachlor epoxide	4.87	4.87	4.77	4.97	0.00
Endosulfan I	5.25	5.25	5.15	5.35	0.00
Dieldrin	5.51	5.51	5.41	5.61	0.00
4,4'-DDE	5.38	5.38	5.28	5.48	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Endosulfan II	6.08	6.08	5.98	6.18	0.00
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
Endosulfan sulfate	6.48	6.48	6.38	6.58	0.00
4,4'-DDT	6.18	6.18	6.08	6.28	0.00
Methoxychlor	6.75	6.76	6.66	6.86	0.01
Endrin ketone	6.99	6.99	6.89	7.09	0.00
Endrin aldehyde	6.26	6.26	6.16	6.36	0.00
alpha-Chlordane	5.19	5.19	5.09	5.29	0.00
gamma-Chlordane	5.13	5.13	5.03	5.23	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/25/2025

Lab Sample No.: PSTDCCC050 Data File : PD089155.D Time Analyzed: 22:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	45.490	50.000	-9.0
4,4'-DDE	6.195	6.096	6.296	46.090	50.000	-7.8
4,4'-DDT	7.020	6.921	7.121	40.060	50.000	-19.9
Aldrin	5.271	5.171	5.371	50.000	50.000	0.0
alpha-BHC	3.999	3.899	4.099	54.860	50.000	9.7
alpha-Chlordane	6.027	5.927	6.127	48.640	50.000	-2.7
beta-BHC	4.515	4.415	4.615	52.340	50.000	4.7
Decachlorobiphenyl	9.072	8.972	9.172	41.670	50.000	-16.7
delta-BHC	4.764	4.664	4.864	56.380	50.000	12.8
Dieldrin	6.347	6.247	6.447	45.400	50.000	-9.2
Endosulfan I	6.074	5.975	6.175	48.330	50.000	-3.3
Endosulfan II	6.785	6.686	6.886	45.580	50.000	-8.8
Endosulfan sulfate	7.148	7.049	7.249	42.630	50.000	-14.7
Endrin	6.572	6.475	6.675	43.610	50.000	-12.8
Endrin aldehyde	6.914	6.815	7.015	43.020	50.000	-14.0
Endrin ketone	7.630	7.530	7.730	42.260	50.000	-15.5
gamma-BHC (Lindane)	4.330	4.230	4.430	54.290	50.000	8.6
gamma-Chlordane	5.946	5.846	6.046	49.610	50.000	-0.8
Heptachlor	4.929	4.829	5.029	51.480	50.000	3.0
Heptachlor epoxide	5.690	5.591	5.791	49.080	50.000	-1.8
Methoxychlor	7.491	7.393	7.593	38.160	50.000	-23.7
Tetrachloro-m-xylene	3.551	3.450	3.650	52.830	50.000	5.7

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/25/2025

Lab Sample No.: PSTDCCC050 Data File : PD089155.D Time Analyzed: 22:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	49.150	50.000	-1.7
4,4'-DDE	5.375	5.275	5.475	49.570	50.000	-0.9
4,4'-DDT	6.183	6.084	6.284	44.960	50.000	-10.1
Aldrin	4.368	4.269	4.469	51.690	50.000	3.4
alpha-BHC	3.393	3.293	3.493	52.300	50.000	4.6
alpha-Chlordane	5.190	5.091	5.291	49.730	50.000	-0.5
beta-BHC	4.025	3.926	4.126	49.430	50.000	-1.1
Decachlorobiphenyl	8.071	7.972	8.172	40.090	50.000	-19.8
delta-BHC	4.262	4.162	4.362	51.750	50.000	3.5
Dieldrin	5.512	5.413	5.613	49.360	50.000	-1.3
Endosulfan I	5.246	5.147	5.347	46.120	50.000	-7.8
Endosulfan II	6.079	5.981	6.181	47.570	50.000	-4.9
Endosulfan sulfate	6.481	6.383	6.583	45.570	50.000	-8.9
Endrin	5.789	5.689	5.889	50.010	50.000	0.0
Endrin aldehyde	6.258	6.159	6.359	45.240	50.000	-9.5
Endrin ketone	6.990	6.892	7.092	43.850	50.000	-12.3
gamma-BHC (Lindane)	3.730	3.630	3.830	52.050	50.000	4.1
gamma-Chlordane	5.125	5.026	5.226	49.880	50.000	-0.2
Heptachlor	4.082	3.983	4.183	50.510	50.000	1.0
Heptachlor epoxide	4.872	4.773	4.973	50.800	50.000	1.6
Methoxychlor	6.752	6.655	6.855	42.310	50.000	-15.4
Tetrachloro-m-xylene	2.881	2.780	2.980	52.150	50.000	4.3

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No. (PEM): PEM - PD088991.D Date Analyzed: 06/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 15:25

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.071	8.970	9.170	21.440	20.000	7.2
Tetrachloro-m-xylene	3.550	3.500	3.600	19.860	20.000	-0.7
alpha-BHC	3.999	3.950	4.050	8.970	10.000	-10.3
beta-BHC	4.514	4.460	4.560	9.720	10.000	-2.8
gamma-BHC (Lindane)	4.330	4.280	4.380	9.320	10.000	-6.8
Endrin	6.574	6.500	6.640	50.260	50.000	0.5
4,4'-DDT	7.020	6.950	7.090	100.820	100.000	0.8
Methoxychlor	7.493	7.420	7.560	229.880	250.000	-8.0

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No. (PEM): PEM - PD088991.D Date Analyzed: 06/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 15:25

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.072	7.970	8.170	21.290	20.000	6.5
Tetrachloro-m-xylene	2.880	2.830	2.930	20.440	20.000	2.2
alpha-BHC	3.393	3.340	3.440	10.570	10.000	5.7
beta-BHC	4.025	3.970	4.080	10.750	10.000	7.5
gamma-BHC (Lindane)	3.729	3.680	3.780	10.580	10.000	5.8
Endrin	5.789	5.720	5.860	50.490	50.000	1.0
4,4'-DDT	6.183	6.110	6.250	96.080	100.000	-3.9
Methoxychlor	6.754	6.680	6.820	198.860	250.000	-20.5

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No. (PEM): PEM - PD089118.D Date Analyzed: 06/25/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:25

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.073	8.970	9.170	20.340	20.000	1.7
Tetrachloro-m-xylene	3.551	3.500	3.600	20.150	20.000	0.8
alpha-BHC	3.999	3.950	4.050	9.340	10.000	-6.6
beta-BHC	4.516	4.470	4.570	10.310	10.000	3.1
gamma-BHC (Lindane)	4.330	4.280	4.380	9.690	10.000	-3.1
Endrin	6.574	6.500	6.640	47.680	50.000	-4.6
4,4'-DDT	7.021	6.950	7.090	91.830	100.000	-8.2
Methoxychlor	7.493	7.420	7.560	207.080	250.000	-17.2

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No. (PEM): PEM - PD089118.D Date Analyzed: 06/25/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:25

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.072	7.970	8.170	21.190	20.000	6.0
Tetrachloro-m-xylene	2.881	2.830	2.930	21.190	20.000	6.0
alpha-BHC	3.393	3.340	3.440	10.920	10.000	9.2
beta-BHC	4.025	3.970	4.080	11.040	10.000	10.4
gamma-BHC (Lindane)	3.729	3.680	3.780	10.990	10.000	9.9
Endrin	5.789	5.720	5.860	46.590	50.000	-6.8
4,4'-DDT	6.183	6.110	6.250	89.560	100.000	-10.4
Methoxychlor	6.753	6.680	6.820	175.380	250.000	-29.8

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No. (PEM): PEM - PD089145.D Date Analyzed: 06/25/2025

Lab Sample No.(PEM): PEM Time Analyzed: 19:02

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.071	8.970	9.170	20.450	20.000	2.3
Tetrachloro-m-xylene	3.550	3.500	3.600	20.140	20.000	0.7
alpha-BHC	3.999	3.950	4.050	9.260	10.000	-7.4
beta-BHC	4.515	4.460	4.570	10.200	10.000	2.0
gamma-BHC (Lindane)	4.330	4.280	4.380	9.720	10.000	-2.8
Endrin	6.574	6.500	6.640	47.960	50.000	-4.1
4,4'-DDT	7.021	6.950	7.090	91.290	100.000	-8.7
Methoxychlor	7.492	7.420	7.560	206.430	250.000	-17.4

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No. (PEM): PEM - PD089145.D Date Analyzed: 06/25/2025

Lab Sample No.(PEM): PEM Time Analyzed: 19:02

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.072	7.970	8.170	20.280	20.000	1.4
Tetrachloro-m-xylene	2.881	2.830	2.930	20.900	20.000	4.5
alpha-BHC	3.393	3.340	3.440	10.770	10.000	7.7
beta-BHC	4.025	3.970	4.080	10.780	10.000	7.8
gamma-BHC (Lindane)	3.729	3.680	3.780	10.840	10.000	8.4
Endrin	5.789	5.720	5.860	47.420	50.000	-5.2
4,4'-DDT	6.183	6.110	6.250	88.830	100.000	-11.2
Methoxychlor	6.754	6.680	6.820	180.220	250.000	-27.9

Analytical Sequence

Client: Union Paving & Construction Company	SDG No.: Q2393
Project: 190 Park Ave Hanover Twp NJ - PO 2556-00	Instrument ID: ECD_D
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 06/17/2025 06/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	06/17/2025	15:11	PD088990.D	9.07	3.55
PEM	PEM	06/17/2025	15:25	PD088991.D	9.07	3.55
RESCHK	RESCHK	06/17/2025	15:39	PD088992.D	9.07	3.55
PSTDICC100	PSTDICC100	06/17/2025	15:52	PD088993.D	9.07	3.55
PSTDICC075	PSTDICC075	06/17/2025	16:06	PD088994.D	9.07	3.55
PSTDICC050	PSTDICC050	06/17/2025	16:20	PD088995.D	9.07	3.55
PSTDICC025	PSTDICC025	06/17/2025	16:33	PD088996.D	9.07	3.55
PSTDICC005	PSTDICC005	06/17/2025	16:47	PD088997.D	9.07	3.55
PCHLORICC500	PCHLORICC500	06/17/2025	17:28	PD089000.D	9.07	3.55
PTOXICC500	PTOXICC500	06/17/2025	18:36	PD089005.D	9.07	3.55
PEM	PEM	06/25/2025	09:25	PD089118.D	9.07	3.55
IBLK	IBLK	06/25/2025	15:58	PD089133.D	9.08	3.55
PSTDCCC050	PSTDCCC050	06/25/2025	16:12	PD089134.D	9.07	3.55
PB168608BL	PB168608BL	06/25/2025	16:59	PD089136.D	9.07	3.55
PB168608BS	PB168608BS	06/25/2025	17:13	PD089137.D	9.07	3.55
PARK AVE -1	Q2393-01	06/25/2025	17:54	PD089140.D	9.07	3.55
PARK AVE -1MS	Q2393-01MS	06/25/2025	18:07	PD089141.D	9.07	3.55
PARK AVE -1MSD	Q2393-01MSD	06/25/2025	18:21	PD089142.D	9.07	3.55
PARK AVE -2	Q2393-03	06/25/2025	18:35	PD089143.D	9.07	3.55
IBLK	IBLK	06/25/2025	18:48	PD089144.D	9.07	3.55
PEM	PEM	06/25/2025	19:02	PD089145.D	9.07	3.55
PSTDCCC050	PSTDCCC050	06/25/2025	19:15	PD089146.D	9.07	3.55
PARK AVE -3	Q2393-05	06/25/2025	19:57	PD089147.D	9.07	3.55
PARK AVE -4	Q2393-07	06/25/2025	20:10	PD089148.D	9.07	3.55
PARK AVE -5	Q2393-09	06/25/2025	20:24	PD089149.D	9.07	3.55
PARK AVE -6	Q2393-11	06/25/2025	20:37	PD089150.D	9.07	3.55
IBLK	IBLK	06/25/2025	21:32	PD089154.D	9.07	3.55
PSTDCCC050	PSTDCCC050	06/25/2025	22:40	PD089155.D	9.07	3.55

Analytical Sequence

Client: Union Paving & Construction Company	SDG No.: Q2393
Project: 190 Park Ave Hanover Twp NJ - PO 2556-00	Instrument ID: ECD_D
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 06/17/2025 06/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	06/17/2025	15:11	PD088990.D	8.07	2.88
PEM	PEM	06/17/2025	15:25	PD088991.D	8.07	2.88
RESCHK	RESCHK	06/17/2025	15:39	PD088992.D	8.07	2.88
PSTDICC100	PSTDICC100	06/17/2025	15:52	PD088993.D	8.07	2.88
PSTDICC075	PSTDICC075	06/17/2025	16:06	PD088994.D	8.07	2.88
PSTDICC050	PSTDICC050	06/17/2025	16:20	PD088995.D	8.07	2.88
PSTDICC025	PSTDICC025	06/17/2025	16:33	PD088996.D	8.07	2.88
PSTDICC005	PSTDICC005	06/17/2025	16:47	PD088997.D	8.07	2.88
PCHLORICC500	PCHLORICC500	06/17/2025	17:28	PD089000.D	8.07	2.88
PTOXICC500	PTOXICC500	06/17/2025	18:36	PD089005.D	8.07	2.88
PEM	PEM	06/25/2025	09:25	PD089118.D	8.07	2.88
IBLK	IBLK	06/25/2025	15:58	PD089133.D	8.07	2.88
PSTDCCC050	PSTDCCC050	06/25/2025	16:12	PD089134.D	8.07	2.88
PB168608BL	PB168608BL	06/25/2025	16:59	PD089136.D	8.07	2.88
PB168608BS	PB168608BS	06/25/2025	17:13	PD089137.D	8.07	2.88
PARK AVE -1	Q2393-01	06/25/2025	17:54	PD089140.D	8.07	2.88
PARK AVE -1MS	Q2393-01MS	06/25/2025	18:07	PD089141.D	8.07	2.88
PARK AVE -1MSD	Q2393-01MSD	06/25/2025	18:21	PD089142.D	8.07	2.88
PARK AVE -2	Q2393-03	06/25/2025	18:35	PD089143.D	8.07	2.88
IBLK	IBLK	06/25/2025	18:48	PD089144.D	8.07	2.88
PEM	PEM	06/25/2025	19:02	PD089145.D	8.07	2.88
PSTDCCC050	PSTDCCC050	06/25/2025	19:15	PD089146.D	8.07	2.88
PARK AVE -3	Q2393-05	06/25/2025	19:57	PD089147.D	8.07	2.88
PARK AVE -4	Q2393-07	06/25/2025	20:10	PD089148.D	8.07	2.88
PARK AVE -5	Q2393-09	06/25/2025	20:24	PD089149.D	8.07	2.88
PARK AVE -6	Q2393-11	06/25/2025	20:37	PD089150.D	8.07	2.88
IBLK	IBLK	06/25/2025	21:32	PD089154.D	8.07	2.88
PSTDCCC050	PSTDCCC050	06/25/2025	22:40	PD089155.D	8.07	2.88

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PARK AVE -1MS

Contract: UNIO02

Lab Code: CHEM

Case No.: Q2393

SAS No.: Q2393

SDG NO.: Q2393

Lab Sample ID: Q2393-01MS

Date(s) Analyzed: 06/25/2025

06/25/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.79	6.74	6.84	16.6	1.8
	2	6.08	6.03	6.13	16.9	
4,4'-DDD	1	6.71	6.66	6.76	17.6	1.7
	2	5.93	5.88	5.98	17.3	
4,4'-DDT	1	7.02	6.97	7.07	15.8	1.3
	2	6.18	6.13	6.23	16.0	
alpha-BHC	1	4.00	3.95	4.05	18.0	0
	2	3.39	3.34	3.44	18.0	
Aldrin	1	5.27	5.22	5.32	17.8	0.6
	2	4.37	4.32	4.42	17.9	
beta-BHC	1	4.52	4.47	4.57	17.3	2.3
	2	4.03	3.98	4.08	17.7	
delta-BHC	1	4.76	4.71	4.81	18.8	4.3
	2	4.26	4.21	4.31	18.0	
Endosulfan I	1	6.07	6.02	6.12	17.3	1.7
	2	5.25	5.20	5.30	17.6	
alpha-Chlordane	1	6.03	5.98	6.08	17.3	1.7
	2	5.19	5.14	5.24	17.6	
4,4'-DDE	1	6.20	6.15	6.25	17.1	1.2
	2	5.38	5.33	5.43	17.3	
Dieldrin	1	6.35	6.30	6.40	17.4	0.6
	2	5.51	5.46	5.56	17.5	
Endrin aldehyde	1	6.92	6.87	6.97	17.2	2.4
	2	6.26	6.21	6.31	16.8	
Endosulfan sulfate	1	7.15	7.10	7.20	16.7	0.6
	2	6.48	6.43	6.53	16.6	
Methoxychlor	1	7.49	7.44	7.54	15.6	1.3
	2	6.75	6.70	6.80	15.4	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PARK AVE -1MS

Contract: UNIO02

Lab Code: CHEM

Case No.: Q2393

SAS No.: Q2393

SDG NO.: Q2393

Lab Sample ID: Q2393-01MS

Date(s) Analyzed: 06/25/2025

06/25/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin ketone	1	7.63	7.58	7.68	17.0	3.6
	2	6.99	6.94	7.04	16.4	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	18.0	0
	2	3.73	3.68	3.78	18.0	
Heptachlor	1	4.93	4.88	4.98	17.3	0.6
	2	4.08	4.03	4.13	17.4	
Heptachlor epoxide	1	5.69	5.64	5.74	17.4	1.1
	2	4.87	4.82	4.92	17.6	
gamma-Chlordane	1	5.95	5.90	6.00	17.1	2.9
	2	5.13	5.08	5.18	17.6	
Endrin	1	6.57	6.52	6.62	16.6	4.7
	2	5.79	5.74	5.84	17.4	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PARK AVE -1MSD

Contract: UNIO02

Lab Code: CHEM

Case No.: Q2393

SAS No.: Q2393

SDG NO.: Q2393

Lab Sample ID: Q2393-01MSD

Date(s) Analyzed: 06/25/2025

06/25/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.79	6.74	6.84	16.7	2.4
	2	6.08	6.03	6.13	17.1	
4,4'-DDD	1	6.70	6.65	6.75	17.8	6.4
	2	5.93	5.88	5.98	16.7	
4,4'-DDT	1	7.02	6.97	7.07	16.1	1.2
	2	6.18	6.13	6.23	16.3	
Endrin aldehyde	1	6.91	6.86	6.96	17.4	1.7
	2	6.26	6.21	6.31	17.1	
Endosulfan sulfate	1	7.15	7.10	7.20	16.9	0.6
	2	6.48	6.43	6.53	16.8	
Methoxychlor	1	7.49	7.44	7.54	15.7	0
	2	6.75	6.70	6.80	15.7	
Endrin ketone	1	7.63	7.58	7.68	17.1	2.4
	2	6.99	6.94	7.04	16.7	
alpha-BHC	1	4.00	3.95	4.05	17.9	0.6
	2	3.39	3.34	3.44	18.0	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	17.9	0.6
	2	3.73	3.68	3.78	18.0	
Heptachlor	1	4.93	4.88	4.98	17.4	0
	2	4.08	4.03	4.13	17.4	
Aldrin	1	5.27	5.22	5.32	17.8	0.6
	2	4.37	4.32	4.42	17.9	
beta-BHC	1	4.52	4.47	4.57	17.5	1.7
	2	4.03	3.98	4.08	17.8	
delta-BHC	1	4.76	4.71	4.81	18.8	4.9
	2	4.26	4.21	4.31	17.9	
Heptachlor epoxide	1	5.69	5.64	5.74	17.4	1.7
	2	4.87	4.82	4.92	17.7	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PARK AVE -1MSD

Contract: UNIO02

Lab Code: CHEM **Case No.:** Q2393

SAS No.: Q2393 **SDG NO.:** Q2393

Lab Sample ID: Q2393-01MSD

Date(s) Analyzed: 06/25/2025 06/25/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 **ID:** 0.32 (mm)

GC Column:(2): ZB-MR2 **ID:** 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.07	6.02	6.12	17.6	0.6
	2	5.25	5.20	5.30	17.7	
gamma-Chlordane	1	5.94	5.89	5.99	17.2	2.9
	2	5.13	5.08	5.18	17.7	
alpha-Chlordane	1	6.03	5.98	6.08	17.5	0.6
	2	5.19	5.14	5.24	17.6	
4,4'-DDE	1	6.20	6.15	6.25	17.5	0.6
	2	5.37	5.32	5.42	17.4	
Dieldrin	1	6.35	6.30	6.40	17.4	0.6
	2	5.51	5.46	5.56	17.5	
Endrin	1	6.57	6.52	6.62	17.2	1.2
	2	5.79	5.74	5.84	17.0	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PARK AVE -3

Contract: UNIO02

Lab Code: CHEM

Case No.: Q2393

SAS No.: Q2393

SDG NO.: Q2393

Lab Sample ID: Q2393-05

Date(s) Analyzed: 06/25/2025 06/25/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	0.70	60
	2	5.93	5.88	5.98	0.38	
4,4'-DDT	1	7.02	6.97	7.07	0.40	27.9
	2	6.18	6.13	6.23	0.53	
gamma-Chlordane	1	5.95	5.90	6.00	0.73	30.5
	2	5.12	5.07	5.17	0.53	
alpha-Chlordane	1	6.03	5.98	6.08	1.10	54.9
	2	5.19	5.14	5.24	0.63	
4,4'-DDE	1	6.20	6.15	6.25	0.27	51.6
	2	5.37	5.32	5.42	0.45	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PARK AVE -4

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

Lab Sample ID: Q2393-07 Date(s) Analyzed: 06/25/2025 06/25/2025

Instrument ID (1): ECD_D Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	7.02	6.97	7.07	0.28	26
	2	6.18	6.13	6.23	0.22	
gamma-Chlordane	1	5.94	5.89	5.99	0.25	27.1
	2	5.12	5.07	5.17	0.19	
alpha-Chlordane	1	6.03	5.98	6.08	0.55	47.2
	2	5.19	5.14	5.24	0.34	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PARK AVE -5

Contract: UNIO02

Lab Code: CHEM

Case No.: Q2393

SAS No.: Q2393

SDG NO.: Q2393

Lab Sample ID: Q2393-09

Date(s) Analyzed: 06/25/2025 06/25/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	7.02	6.97	7.07	0.16	54.5
	2	6.18	6.13	6.23	0.28	
Aldrin	1	5.27	5.22	5.32	0.13	34.5
	2	4.37	4.32	4.42	0.19	
gamma-Chlordane	1	5.94	5.89	5.99	0.35	34.1
	2	5.12	5.07	5.17	0.25	
alpha-Chlordane	1	6.03	5.98	6.08	0.56	37.2
	2	5.19	5.14	5.24	0.38	
4,4'-DDE	1	6.19	6.14	6.24	0.27	29.3
	2	5.37	5.32	5.42	0.36	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PARK AVE -6

Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG NO.: Q2393

Lab Sample ID: Q2393-11 Date(s) Analyzed: 06/25/2025 06/25/2025

Instrument ID (1): ECD_D Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
gamma-Chlordane	1	5.94	5.89	5.99	0.38	41.5
	2	5.12	5.07	5.17	0.25	
alpha-Chlordane	1	6.03	5.98	6.08	0.61	28
	2	5.19	5.14	5.24	0.46	
4,4'-DDE	1	6.20	6.15	6.25	0.19	40
	2	5.37	5.32	5.42	0.29	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168608BS

Contract: UNIO02

Lab Code: CHEM

Case No.: Q2393

SAS No.: Q2393

SDG NO.: Q2393

Lab Sample ID: PB168608BS

Date(s) Analyzed: 06/25/2025

06/25/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.71	6.66	6.76	16.6	0
	2	5.93	5.88	5.98	16.6	
4,4'-DDE	1	6.20	6.15	6.25	16.2	0.6
	2	5.38	5.33	5.43	16.1	
4,4'-DDT	1	7.02	6.97	7.07	15.1	0.7
	2	6.18	6.13	6.23	15.0	
alpha-BHC	1	4.00	3.95	4.05	16.1	0.6
	2	3.39	3.34	3.44	16.2	
Aldrin	1	5.27	5.22	5.32	16.3	0
	2	4.37	4.32	4.42	16.3	
alpha-Chlordane	1	6.03	5.98	6.08	16.0	0.6
	2	5.19	5.14	5.24	16.1	
Endosulfan II	1	6.79	6.74	6.84	16.3	1.9
	2	6.08	6.03	6.13	16.0	
Endosulfan sulfate	1	7.15	7.10	7.20	16.0	0.6
	2	6.48	6.43	6.53	15.9	
beta-BHC	1	4.52	4.47	4.57	15.5	3.2
	2	4.03	3.98	4.08	16.0	
delta-BHC	1	4.76	4.71	4.81	17.0	4.2
	2	4.26	4.21	4.31	16.3	
Endosulfan I	1	6.07	6.02	6.12	16.1	1.2
	2	5.25	5.20	5.30	16.3	
Dieldrin	1	6.35	6.30	6.40	16.2	0.6
	2	5.51	5.46	5.56	16.1	
Endrin aldehyde	1	6.92	6.87	6.97	15.9	0.6
	2	6.26	6.21	6.31	15.8	
Methoxychlor	1	7.49	7.44	7.54	14.3	2.8
	2	6.75	6.70	6.80	14.7	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168608BS

Contract: UNIO02

Lab Code: CHEM

Case No.: Q2393

SAS No.: Q2393

SDG NO.: Q2393

Lab Sample ID: PB168608BS

Date(s) Analyzed: 06/25/2025

06/25/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin ketone	1	7.63	7.58	7.68	16.4	0.6
	2	6.99	6.94	7.04	16.3	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	16.0	1.2
	2	3.73	3.68	3.78	16.2	
Heptachlor	1	4.93	4.88	4.98	15.8	0.6
	2	4.08	4.03	4.13	15.7	
Heptachlor epoxide	1	5.69	5.64	5.74	16.0	1.2
	2	4.87	4.82	4.92	16.2	
gamma-Chlordane	1	5.95	5.90	6.00	16.1	1.2
	2	5.13	5.08	5.18	16.3	
Endrin	1	6.57	6.52	6.62	15.2	1.3
	2	5.79	5.74	5.84	15.0	



SAMPLE RAW DATA

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
 Data File : PD089140.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 17:54
 Operator : AR\AJ
 Sample : Q2393-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PARK AVE -1

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 06:40:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.881	50882452	321.0E6	17.834	18.934
28)	SA Decachlor...	9.072	8.071	69147188	324.0E6	17.605	16.387

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
Data File : PD089140.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 17:54
Operator : AR\AJ
Sample : Q2393-01
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PARK AVE -1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 06:40:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.550 min
Delta R.T.: 0.000 min
Response: 50882452
Conc: 17.83 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -1




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: 0.000 min
Response: 320995427
Conc: 18.93 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.072 min
Delta R.T.: 0.000 min
Response: 69147188
Conc: 17.61 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.071 min
Delta R.T.: 0.000 min
Response: 323972774
Conc: 16.39 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
 Data File : PD089143.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 18:35
 Operator : AR\AJ
 Sample : Q2393-03
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PARK AVE -2

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 06:41:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.551	2.882	47232586	296.1E6	16.555	17.465
28) SA Decachlor...	9.072	8.071	61974849	311.8E6	15.779	15.773

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
Data File : PD089143.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 18:35
Operator : AR\AJ
Sample : Q2393-03
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PARK AVE -2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 06:41:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.551 min
Delta R.T.: 0.000 min
Response: 47232586
Conc: 16.55 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -2




#1 Tetrachloro-m-xylene

R.T.: 2.882 min
Delta R.T.: 0.000 min
Response: 296094984
Conc: 17.47 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.072 min
Delta R.T.: 0.000 min
Response: 61974849
Conc: 15.78 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.071 min
Delta R.T.: 0.000 min
Response: 311831064
Conc: 15.77 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
 Data File : PD089147.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 19:57
 Operator : AR\AJ
 Sample : Q2393-05
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PARK AVE -3

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 06/26/2025
 Supervised By :mohammad ahmed 06/27/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 06:41:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.881	43157521	269.1E6	15.126	15.874
28)	SA Decachlor...	9.071	8.071	56789006	260.7E6	14.459m	13.185
Target Compounds							
10)	B gamma-Chl...	5.945	5.124	9031380	33205082	1.886	1.388m#
11)	B alpha-Chl...	6.030	5.188	13308266	37225834	2.750	1.626m#
12)	B 4,4'-DDE	6.197	5.373	3017121	26757879	0.692	1.168m#
16)	A 4,4'-DDD	6.701	5.929	6261222	18796923	1.824m	0.983m#
17)	MA 4,4'-DDT	7.021	6.182	3971320	28246117	1.048	1.393 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
Data File : PD089147.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 19:57
Operator : AR\AJ
Sample : Q2393-05
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PARK AVE -3

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 06/26/2025

Supervised By :mohammad ahmed 06/27/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 06:41:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.550 min
Delta R.T.: 0.000 min
Response: 43157521
Conc: 15.13 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -3


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: 0.000 min
Response: 269113573
Conc: 15.87 ng/ml




#10 gamma-Chlordane

R.T.: 5.945 min
Delta R.T.: 0.000 min
Response: 9031380
Conc: 1.89 ng/ml



#10 gamma-Chlordane

R.T.: 5.124 min
Delta R.T.: -0.002 min
Response: 33205082
Conc: 1.39 ng/ml m




#11 alpha-Chlordane

R.T.: 6.030 min
Delta R.T.: 0.003 min
Response: 13308266
Conc: 2.75 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -3


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025




#11 alpha-Chlordane

R.T.: 5.188 min
Delta R.T.: -0.002 min
Response: 37225834
Conc: 1.63 ng/ml m




#12 4,4'-DDE

R.T.: 6.197 min
Delta R.T.: 0.001 min
Response: 3017121
Conc: 0.69 ng/ml



#12 4,4'-DDE

R.T.: 5.373 min
Delta R.T.: -0.002 min
Response: 26757879
Conc: 1.17 ng/ml m




#16 4,4'-DDD

R.T.: 6.701 min
Delta R.T.: -0.004 min
Response: 6261222
Conc: 1.82 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -3


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025




#16 4,4'-DDD

R.T.: 5.929 min
Delta R.T.: -0.001 min
Response: 18796923
Conc: 0.98 ng/ml m




#17 4,4'-DDT

R.T.: 7.021 min
Delta R.T.: 0.000 min
Response: 3971320
Conc: 1.05 ng/ml



#17 4,4'-DDT

R.T.: 6.182 min
Delta R.T.: -0.002 min
Response: 28246117
Conc: 1.39 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.071 min
Delta R.T.: -0.001 min
Response: 56789006
Conc: 14.46 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -3

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025



#28 Decachlorobiphenyl

R.T.: 8.071 min
Delta R.T.: -0.001 min
Response: 260675039
Conc: 13.19 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
 Data File : PD089148.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 20:10
 Operator : AR\AJ
 Sample : Q2393-07
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PARK AVE -4

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 06/26/2025
 Supervised By :mohammad ahmed 06/27/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 06:41:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.550	2.881	33727008	211.9E6	11.821	12.497
28) SA Decachlor...	9.073	8.072	43753099	189.7E6	11.140	9.594
Target Compounds						
10) B gamma-Chl...	5.944	5.124	3197551	12167653	0.668m	0.509m
11) B alpha-Chl...	6.030	5.189	6986922	20422106	1.444	0.892m#
12) B 4,4'-DDE	6.196	5.373	1794840	16356893	0.412m	0.714m#
16) A 4,4'-DDD	6.703	5.929	959216	7118398	0.279m	0.372m#
17) MA 4,4'-DDT	7.019	6.182	2800688	11650801	0.739m	0.575

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
Data File : PD089148.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 20:10
Operator : AR\AJ
Sample : Q2393-07
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PARK AVE -4

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 06/26/2025

Supervised By :mohammad ahmed 06/27/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 06:41:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.550 min
Delta R.T.: 0.000 min
Response: 33727008
Conc: 11.82 ng/ml

Instrument :


ECD_D

ClientSampleId :

PARK AVE -4


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: 0.000 min
Response: 211866200
Conc: 12.50 ng/ml




#10 gamma-Chlordane

R.T.: 5.944 min
Delta R.T.: -0.002 min
Response: 3197551
Conc: 0.67 ng/ml m



#10 gamma-Chlordane

R.T.: 5.124 min
Delta R.T.: -0.001 min
Response: 12167653
Conc: 0.51 ng/ml m




#11 alpha-Chlordane

R.T.: 6.030 min
Delta R.T.: 0.003 min
Response: 6986922
Conc: 1.44 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -4


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025




#11 alpha-Chlordane

R.T.: 5.189 min
Delta R.T.: -0.002 min
Response: 20422106
Conc: 0.89 ng/ml m



#12 4,4'-DDE

R.T.: 6.196 min
Delta R.T.: 0.000 min
Response: 1794840
Conc: 0.41 ng/ml m




#12 4,4'-DDE

R.T.: 5.373 min
Delta R.T.: -0.002 min
Response: 16356893
Conc: 0.71 ng/ml m

Response_ Signal: PD089148.D\ECD1A.ch

#16 4,4'-DDD



R.T.: 6.703 min
Delta R.T.: -0.002 min
Response: 959216
Conc: 0.28 ng/ml


Instrument :
ECD_D
ClientSampleId :
PARK AVE -4

**Manual Integrations
APPROVED**

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025

Time Response_ Signal: PD089148.D\ECD2B.ch


#16 4,4'-DDD



R.T.: 5.929 min
Delta R.T.: 0.000 min
Response: 7118398
Conc: 0.37 ng/ml m

Time Response_ Signal: PD089148.D\ECD1A.ch


#17 4,4'-DDT




R.T.: 7.019 min
Delta R.T.: 0.000 min
Response: 2800688
Conc: 0.74 ng/ml m

Time Response_ Signal: PD089148.D\ECD2B.ch

#17 4,4'-DDT



R.T.: 6.182 min
Delta R.T.: -0.002 min
Response: 11650801
Conc: 0.57 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.073 min
Delta R.T.: 0.000 min
Response: 43753099
Conc: 11.14 ng/ml

Instrument :

ECD_D

ClientSampleId :


PARK AVE -4

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 06/26/2025

Supervised By :mohammad ahmed 06/27/2025



#28 Decachlorobiphenyl

R.T.: 8.072 min
Delta R.T.: 0.000 min
Response: 189671822
Conc: 9.59 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
 Data File : PD089149.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 20:24
 Operator : AR\AJ
 Sample : Q2393-09
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PARK AVE -5

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 06/26/2025
 Supervised By : mohammad ahmed 06/27/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 06:41:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.881	29062907	182.0E6	10.186	10.733
28)	SA Decachlor...	9.071	8.071	35658731	166.9E6	9.079	8.441
Target Compounds							
5)	MB Aldrin	5.269	4.368	1918409	12351970	0.361	0.510 #
10)	B gamma-Chl...	5.943	5.124	4601524	16244417	0.961m	0.679m#
11)	B alpha-Chl...	6.029	5.188	7409527	23992526	1.531	1.048m#
12)	B 4,4'-DDE	6.194	5.373	3181857	22374537	0.730m	0.976m#
16)	A 4,4'-DDD	6.702	5.927	1153743	5915960	0.336m	0.309m
17)	MA 4,4'-DDT	7.019	6.180	1666922	15577952	0.440	0.768 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
Data File : PD089149.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 20:24
Operator : AR\AJ
Sample : Q2393-09
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PARK AVE -5

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 06/26/2025

Supervised By :mohammad ahmed 06/27/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 06:41:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.549 min
Delta R.T.: 0.000 min
Response: 29062907
Conc: 10.19 ng/ml

Instrument :

ECD_D

ClientSampleId :


PARK AVE -5

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 06/26/2025

Supervised By :mohammad ahmed 06/27/2025




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: 0.000 min
Response: 181958809
Conc: 10.73 ng/ml




#5 Aldrin

R.T.: 5.269 min
Delta R.T.: -0.002 min
Response: 1918409
Conc: 0.36 ng/ml



#5 Aldrin

R.T.: 4.368 min
Delta R.T.: 0.000 min
Response: 12351970
Conc: 0.51 ng/ml




#10 gamma-Chlordane

R.T.: 5.943 min
Delta R.T.: -0.002 min
Response: 4601524
Conc: 0.96 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -5


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025




#10 gamma-Chlordane

R.T.: 5.124 min
Delta R.T.: -0.002 min
Response: 16244417
Conc: 0.68 ng/ml m




#11 alpha-Chlordane

R.T.: 6.029 min
Delta R.T.: 0.002 min
Response: 7409527
Conc: 1.53 ng/ml



#11 alpha-Chlordane

R.T.: 5.188 min
Delta R.T.: -0.002 min
Response: 23992526
Conc: 1.05 ng/ml m




#12 4,4'-DDE

R.T.: 6.194 min
Delta R.T.: -0.001 min
Response: 3181857
Conc: 0.73 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -5


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025




#12 4,4'-DDE

R.T.: 5.373 min
Delta R.T.: -0.002 min
Response: 22374537
Conc: 0.98 ng/ml m



#16 4,4'-DDD

R.T.: 6.702 min
Delta R.T.: -0.002 min
Response: 1153743
Conc: 0.34 ng/ml m




#16 4,4'-DDD

R.T.: 5.927 min
Delta R.T.: -0.003 min
Response: 5915960
Conc: 0.31 ng/ml m

Response_ Signal: PD089149.D\ECD1A.ch

#17 4,4'-DDT



R.T.: 7.019 min
Delta R.T.: -0.001 min
Response: 1666922
Conc: 0.44 ng/ml


Instrument :
ECD_D
ClientSampleId :
PARK AVE -5

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025

Response_ Signal: PD089149.D\ECD2B.ch


#17 4,4'-DDT



R.T.: 6.180 min
Delta R.T.: -0.004 min
Response: 15577952
Conc: 0.77 ng/ml

Response_ Signal: PD089149.D\ECD1A.ch


#28 Decachlorobiphenyl



R.T.: 9.071 min
Delta R.T.: -0.001 min
Response: 35658731
Conc: 9.08 ng/ml

Response_ Signal: PD089149.D\ECD2B.ch

#28 Decachlorobiphenyl



R.T.: 8.071 min
Delta R.T.: -0.001 min
Response: 166881994
Conc: 8.44 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
 Data File : PD089150.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 20:37
 Operator : AR\AJ
 Sample : Q2393-11
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PARK AVE -6

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 06/26/2025
 Supervised By : mohammad ahmed 06/27/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 06:41:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.550	2.881	29970356	187.6E6	10.504	11.068
28) SA Decachlor...	9.072	8.072	40778081	168.5E6	10.382	8.523
Target Compounds						
10) B gamma-Chl...	5.944	5.124	4751162	15573126	0.992m	0.651m#
11) B alpha-Chl...	6.030	5.188	7699487	27405449	1.591	1.197m
12) B 4,4'-DDE	6.196	5.373	2163514	17169392	0.496	0.749m#
17) MA 4,4'-DDT	7.019	6.178	1078162	13588966	0.285m	0.670m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
Data File : PD089150.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 20:37
Operator : AR\AJ
Sample : Q2393-11
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PARK AVE -6

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 06/26/2025

Supervised By :mohammad ahmed 06/27/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 06:41:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.550 min
Delta R.T.: 0.000 min
Response: 29970356
Conc: 10.50 ng/ml

Instrument :

ECD_D

ClientSampleId :


PARK AVE -6

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 06/26/2025

Supervised By :mohammad ahmed 06/27/2025




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: 0.000 min
Response: 187644495
Conc: 11.07 ng/ml




#10 gamma-Chlordane

R.T.: 5.944 min
Delta R.T.: -0.002 min
Response: 4751162
Conc: 0.99 ng/ml m



#10 gamma-Chlordane

R.T.: 5.124 min
Delta R.T.: -0.002 min
Response: 15573126
Conc: 0.65 ng/ml m



#11 alpha-Chlordane

R.T.: 6.030 min
Delta R.T.: 0.003 min
Response: 7699487
Conc: 1.59 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -6

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025

#11 alpha-Chlordane


R.T.: 5.188 min
Delta R.T.: -0.003 min
Response: 27405449
Conc: 1.20 ng/ml m

#12 4,4'-DDE

R.T.: 6.196 min
Delta R.T.: 0.000 min
Response: 2163514
Conc: 0.50 ng/ml

#12 4,4'-DDE

R.T.: 5.373 min
Delta R.T.: -0.002 min
Response: 17169392
Conc: 0.75 ng/ml m




#17 4,4'-DDT

R.T.: 7.019 min
Delta R.T.: -0.001 min
Response: 1078162
Conc: 0.28 ng/ml

Instrument :
ECD_D
ClientSampleId :
PARK AVE -6

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/26/2025
Supervised By :mohammad ahmed 06/27/2025



#17 4,4'-DDT

R.T.: 6.178 min
Delta R.T.: -0.006 min
Response: 13588966
Conc: 0.67 ng/ml m

#28 Decachlorobiphenyl

R.T.: 9.072 min
Delta R.T.: 0.000 min
Response: 40778081
Conc: 10.38 ng/ml

#28 Decachlorobiphenyl

R.T.: 8.072 min
Delta R.T.: 0.000 min
Response: 168492445
Conc: 8.52 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
 Data File : PD089136.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 16:59
 Operator : AR\AJ
 Sample : PB168608BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB168608BL

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 06:39:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.881	48891301	318.4E6	17.136	18.782
28)	SA Decachlor...	9.073	8.072	73640650	380.9E6	18.749	19.265

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
Data File : PD089136.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 16:59
Operator : AR\AJ
Sample : PB168608BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB168608BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 06:39:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.550 min
Delta R.T.: 0.000 min
Response: 48891301
Conc: 17.14 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB168608BL




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: 0.000 min
Response: 318421112
Conc: 18.78 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.073 min
Delta R.T.: 0.000 min
Response: 73640650
Conc: 18.75 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.072 min
Delta R.T.: 0.000 min
Response: 380875865
Conc: 19.27 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
 Data File : PD089137.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 17:13
 Operator : AR\AJ
 Sample : PB168608BS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB168608BS

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 06:40:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.881	60368991	393.7E6	21.159	23.224
28)	SA Decachlor...	9.072	8.072	92733608	479.6E6	23.611	24.259
Target Compounds							
2)	A alpha-BHC	3.999	3.393	281.1E6	1305.3E6	48.224	48.710
3)	MA gamma-BHC...	4.330	3.729	268.9E6	1205.7E6	48.174	48.702
4)	MA Heptachlor	4.929	4.083	259.0E6	1181.1E6	47.502	47.171
5)	MB Aldrin	5.271	4.368	259.3E6	1181.8E6	48.807	48.799
6)	B beta-BHC	4.516	4.025	102.8E6	517.8E6	46.595	48.183
7)	B delta-BHC	4.764	4.262	262.5E6	1215.4E6	51.142	48.863
8)	B Heptachlo...	5.691	4.872	231.3E6	1062.7E6	47.939	48.528
9)	A Endosulfan I	6.074	5.246	217.7E6	1021.4E6	48.309	49.005
10)	B gamma-Chl...	5.946	5.125	231.6E6	1169.0E6	48.373	48.868
11)	B alpha-Chl...	6.027	5.190	233.2E6	1106.8E6	48.194	48.337
12)	B 4,4'-DDE	6.196	5.375	212.5E6	1106.2E6	48.718	48.273
13)	MA Dieldrin	6.347	5.512	234.0E6	1124.1E6	48.778	48.436
14)	MA Endrin	6.574	5.789	188.9E6	981.0E6	45.777	45.178
15)	B Endosulfa...	6.785	6.080	197.8E6	971.7E6	49.006	48.113
16)	A 4,4'-DDD	6.705	5.929	171.3E6	954.6E6	49.909	49.903
17)	MA 4,4'-DDT	7.021	6.183	171.3E6	912.3E6	45.231	44.985
18)	B Endrin al...	6.915	6.258	146.0E6	727.0E6	47.640	47.482
19)	B Endosulfa...	7.149	6.482	184.3E6	938.3E6	48.028	47.822
20)	A Methoxychlor	7.493	6.754	86659387	473.8E6	42.887	44.002
21)	B Endrin ke...	7.630	6.991	200.7E6	1060.1E6	49.286	49.073
22)	Mirex	8.113	7.185	144.3E6	784.4E6	47.470	47.030



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
Data File : PD089137.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 17:13
Operator : AR\AJ
Sample : PB168608BS
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB168608BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 06:40:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
 Data File : PD089141.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 18:07
 Operator : AR\AJ
 Sample : Q2393-01MS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PARK AVE -1MS

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 06/26/2025
 Supervised By :mohammad ahmed 06/27/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 06:40:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.881	43687418	290.7E6	15.312	17.146
28)	SA Decachlor...	9.071	8.071	64713848	308.7E6	16.477	15.615
Target Compounds							
2)	A alpha-BHC	3.999	3.393	262.2E6	1206.9E6	44.982	45.040
3)	MA gamma-BHC...	4.330	3.730	251.2E6	1114.9E6	45.007	45.033
4)	MA Heptachlor	4.929	4.083	236.2E6	1089.7E6	43.322	43.520
5)	MB Aldrin	5.271	4.369	236.1E6	1085.2E6	44.430	44.810
6)	B beta-BHC	4.515	4.025	95684739	477.1E6	43.352	44.401
7)	B delta-BHC	4.763	4.262	241.4E6	1117.5E6	47.031	44.927
8)	B Heptachlo...	5.690	4.872	209.7E6	965.7E6	43.479	44.098
9)	A Endosulfan I	6.074	5.247	195.2E6	915.7E6	43.298	43.933
10)	B gamma-Chl...	5.945	5.125	204.6E6	1054.3E6	42.728	44.072
11)	B alpha-Chl...	6.026	5.190	209.0E6	1005.7E6	43.187	43.923
12)	B 4,4'-DDE	6.196	5.375	187.2E6	991.0E6	42.914	43.245
13)	MA Dieldrin	6.347	5.512	208.3E6	1013.8E6	43.417	43.683
14)	MA Endrin	6.574	5.789	171.4E6	944.5E6	41.537	43.496
15)	B Endosulfa...	6.785	6.080	167.8E6	852.6E6	41.568	42.214
16)	A 4,4'-DDD	6.705	5.928	150.9E6	826.6E6	43.981	43.210m
17)	MA 4,4'-DDT	7.021	6.183	149.7E6	810.4E6	39.526	39.963
18)	B Endrin al...	6.915	6.258	131.9E6	642.2E6	43.042	41.941
19)	B Endosulfa...	7.149	6.482	160.5E6	815.4E6	41.834	41.557
20)	A Methoxychlor	7.492	6.754	79072173	415.4E6	39.132	38.577
21)	B Endrin ke...	7.630	6.991	173.8E6	886.5E6	42.661	41.039
22)	Mirex	8.113	7.185	128.4E6	677.8E6	42.243	40.639

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
Data File : PD089141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 18:07
Operator : AR\AJ
Sample : Q2393-01MS
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PARK AVE -1MS

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 06/26/2025

Supervised By :mohammad ahmed 06/27/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 06:40:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
 Data File : PD089142.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 18:21
 Operator : AR\AJ
 Sample : Q2393-01MSD
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PARK AVE -1MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 06:41:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.881	44146990	291.0E6	15.473	17.167
28)	SA Decachlor...	9.071	8.071	64212523	316.9E6	16.349	16.027
Target Compounds							
2)	A alpha-BHC	3.999	3.393	262.1E6	1208.3E6	44.962	45.090
3)	MA gamma-BHC...	4.330	3.729	250.8E6	1115.5E6	44.940	45.059
4)	MA Heptachlor	4.929	4.083	238.0E6	1095.5E6	43.662	43.753
5)	MB Aldrin	5.271	4.368	237.7E6	1085.9E6	44.741	44.841
6)	B beta-BHC	4.515	4.025	96643636	478.7E6	43.787	44.550
7)	B delta-BHC	4.763	4.262	242.0E6	1116.6E6	47.148	44.889
8)	B Heptachlo...	5.690	4.872	211.1E6	969.5E6	43.753	44.272
9)	A Endosulfan I	6.074	5.246	198.6E6	922.8E6	44.053	44.277
10)	B gamma-Chl...	5.944	5.125	206.3E6	1064.7E6	43.084	44.504
11)	B alpha-Chl...	6.026	5.189	211.8E6	1011.2E6	43.773	44.165
12)	B 4,4'-DDE	6.195	5.374	191.2E6	998.5E6	43.852	43.571
13)	MA Dieldrin	6.346	5.512	209.8E6	1018.7E6	43.726	43.895
14)	MA Endrin	6.574	5.788	178.2E6	925.5E6	43.206	42.624
15)	B Endosulfa...	6.785	6.079	168.9E6	867.9E6	41.853	42.972
16)	A 4,4'-DDD	6.704	5.928	153.4E6	803.3E6	44.687	41.993
17)	MA 4,4'-DDT	7.020	6.182	152.5E6	829.0E6	40.245	40.882
18)	B Endrin al...	6.914	6.257	133.5E6	655.7E6	43.572	42.826
19)	B Endosulfa...	7.148	6.481	162.7E6	825.6E6	42.388	42.078
20)	A Methoxychlor	7.492	6.753	79740477	425.1E6	39.463	39.478
21)	B Endrin ke...	7.629	6.990	175.0E6	906.3E6	42.964	41.956
22)	Mirex	8.113	7.184	129.2E6	692.0E6	42.506	41.492



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062525\
Data File : PD089142.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 18:21
Operator : AR\AJ
Sample : Q2393-01MSD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PARK AVE -1MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 06:41:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PD061825	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD088991.D	4,4"-DDD	Abdul	6/18/2025 8:40:15 AM	mohammad	6/19/2025 2:43:55	Peak Integrated by Software
PEM	PD088991.D	4,4"-DDD #2	Abdul	6/18/2025 8:40:15 AM	mohammad	6/19/2025 2:43:55	Peak Integrated by Software
PEM	PD088991.D	beta-BHC	Abdul	6/18/2025 8:40:15 AM	mohammad	6/19/2025 2:43:55	Peak Integrated by Software
PEM	PD088991.D	Endrin aldehyde	Abdul	6/18/2025 8:40:15 AM	mohammad	6/19/2025 2:43:55	Peak Integrated by Software
PEM	PD088991.D	Endrin aldehyde #2	Abdul	6/18/2025 8:40:15 AM	mohammad	6/19/2025 2:43:55	Peak Integrated by Software
PEM	PD088991.D	Endrin ketone	Abdul	6/18/2025 8:40:15 AM	mohammad	6/19/2025 2:43:55	Peak Integrated by Software
PEM	PD088991.D	Endrin ketone #2	Abdul	6/18/2025 8:40:15 AM	mohammad	6/19/2025 2:43:55	Peak Integrated by Software
PSTDICC025	PD088996.D	Endrin #2	Abdul	6/18/2025 8:40:24 AM	mohammad	6/19/2025 2:43:55	Peak Integrated by Software
PSTDICC005	PD088997.D	Endrin #2	Abdul	6/18/2025 8:40:30 AM	mohammad	6/19/2025 2:43:55	Peak Integrated by Software
PSTDICC005	PD088997.D	Endrin ketone #2	Abdul	6/18/2025 8:40:30 AM	mohammad	6/19/2025 2:43:55	Peak Integrated by Software

Manual Integration Report

Sequence:	pd062525	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD089118.D	4,4"-DDE	Abdul	6/26/2025 11:16:45 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
PEM	PD089118.D	4,4"-DDE #2	Abdul	6/26/2025 11:16:45 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
PCHLORCCC5 00	PD089120.D	Chlordane-1	Abdul	6/26/2025 11:16:48 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
PCHLORCCC5 00	PD089135.D	Chlordane-1	Abdul	6/26/2025 11:17:16 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-01MS	PD089141.D	4,4"-DDD #2	Abdul	6/26/2025 11:17:24 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
PEM	PD089145.D	4,4"-DDE	Abdul	6/26/2025 11:17:27 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
PEM	PD089145.D	4,4"-DDE #2	Abdul	6/26/2025 11:17:27 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
PEM	PD089145.D	Endrin ketone	Abdul	6/26/2025 11:17:27 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
PEM	PD089145.D	Endrin ketone #2	Abdul	6/26/2025 11:17:27 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-05	PD089147.D	4,4"-DDD	Abdul	6/26/2025 11:17:30 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-05	PD089147.D	4,4"-DDD #2	Abdul	6/26/2025 11:17:30 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-05	PD089147.D	4,4"-DDE #2	Abdul	6/26/2025 11:17:30 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-05	PD089147.D	alpha-Chlordane #2	Abdul	6/26/2025 11:17:30 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software

Manual Integration Report

Sequence:	pd062525	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q2393-05	PD089147.D	Decachlorobiphenyl	Abdul	6/26/2025 11:17:30 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-05	PD089147.D	gamma-Chlordane #2	Abdul	6/26/2025 11:17:30 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-07	PD089148.D	4,4"-DDD	Abdul	6/26/2025 2:43:20 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-07	PD089148.D	4,4"-DDD #2	Abdul	6/26/2025 2:43:20 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-07	PD089148.D	4,4"-DDE	Abdul	6/26/2025 2:43:20 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-07	PD089148.D	4,4"-DDE #2	Abdul	6/26/2025 2:43:20 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-07	PD089148.D	4,4"-DDT	Abdul	6/26/2025 2:43:20 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-07	PD089148.D	alpha-Chlordane #2	Abdul	6/26/2025 2:43:20 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-07	PD089148.D	gamma-Chlordane	Abdul	6/26/2025 2:43:20 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-07	PD089148.D	gamma-Chlordane #2	Abdul	6/26/2025 2:43:20 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-09	PD089149.D	4,4"-DDD	Abdul	6/26/2025 11:17:36 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-09	PD089149.D	4,4"-DDD #2	Abdul	6/26/2025 11:17:36 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-09	PD089149.D	4,4"-DDE	Abdul	6/26/2025 11:17:36 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software

Manual Integration Report

Sequence:	pd062525	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q2393-09	PD089149.D	4,4"-DDE #2	Abdul	6/26/2025 11:17:36 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-09	PD089149.D	alpha-Chlordane #2	Abdul	6/26/2025 11:17:36 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-09	PD089149.D	gamma-Chlordane	Abdul	6/26/2025 11:17:36 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-09	PD089149.D	gamma-Chlordane #2	Abdul	6/26/2025 11:17:36 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-11	PD089150.D	4,4"-DDE #2	Abdul	6/26/2025 2:43:23 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-11	PD089150.D	4,4"-DDT	Abdul	6/26/2025 2:43:23 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-11	PD089150.D	4,4"-DDT #2	Abdul	6/26/2025 2:43:23 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-11	PD089150.D	alpha-Chlordane #2	Abdul	6/26/2025 2:43:23 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-11	PD089150.D	gamma-Chlordane	Abdul	6/26/2025 2:43:23 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
Q2393-11	PD089150.D	gamma-Chlordane #2	Abdul	6/26/2025 2:43:23 PM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
PSTDCCC050	PD089155.D	Endrin	Abdul	6/26/2025 11:17:50 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software
PSTDCCC050	PD089155.D	Methoxychlor	Abdul	6/26/2025 11:17:50 AM	mohammad	6/27/2025 2:49:12	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

pd062525

Instrument

ECD_d

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD061825

Review By	Abdul	Review On	6/18/2025 8:40:56 AM
Supervise By	mohammad	Supervise On	6/19/2025 2:43:55 AM
SubDirectory	PD061825	HP Acquire Method	HP Processing Method pd061825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD088989.D	17 Jun 2025 14:58	AR\AJ	Ok
2	I.BLK	PD088990.D	17 Jun 2025 15:11	AR\AJ	Ok
3	PEM	PD088991.D	17 Jun 2025 15:25	AR\AJ	Ok,M
4	RESCHK	PD088992.D	17 Jun 2025 15:39	AR\AJ	Ok
5	PSTDICC100	PD088993.D	17 Jun 2025 15:52	AR\AJ	Ok
6	PSTDICC075	PD088994.D	17 Jun 2025 16:06	AR\AJ	Ok
7	PSTDICC050	PD088995.D	17 Jun 2025 16:20	AR\AJ	Ok
8	PSTDICC025	PD088996.D	17 Jun 2025 16:33	AR\AJ	Ok,M
9	PSTDICC005	PD088997.D	17 Jun 2025 16:47	AR\AJ	Ok,M
10	PCHLORICC1000	PD088998.D	17 Jun 2025 17:00	AR\AJ	Ok
11	PCHLORICC750	PD088999.D	17 Jun 2025 17:14	AR\AJ	Ok
12	PCHLORICC500	PD089000.D	17 Jun 2025 17:28	AR\AJ	Ok
13	PCHLORICC250	PD089001.D	17 Jun 2025 17:41	AR\AJ	Ok
14	PCHLORICC050	PD089002.D	17 Jun 2025 17:55	AR\AJ	Ok
15	PTOXICC1000	PD089003.D	17 Jun 2025 18:09	AR\AJ	Ok
16	PTOXICC750	PD089004.D	17 Jun 2025 18:22	AR\AJ	Ok
17	PTOXICC500	PD089005.D	17 Jun 2025 18:36	AR\AJ	Ok
18	PTOXICC250	PD089006.D	17 Jun 2025 18:50	AR\AJ	Ok
19	PTOXICC100	PD089007.D	17 Jun 2025 19:03	AR\AJ	Ok,M
20	PSTDICV050	PD089008.D	17 Jun 2025 19:17	AR\AJ	Ok
21	PCHLORICV500	PD089009.D	17 Jun 2025 19:31	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD061825

Review By	Abdul	Review On	6/18/2025 8:40:56 AM
Supervise By	mohammad	Supervise On	6/19/2025 2:43:55 AM
SubDirectory	PD061825	HP Acquire Method	HP Processing Method pd061825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PD089010.D	17 Jun 2025 19:44	AR\AJ	Ok
----	------------	------------	-------------------	-------	----

M : Manual Integration



Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD062525

Review By	Abdul	Review On	6/26/2025 11:18:16 AM
Supervise By	mohammad	Supervise On	6/27/2025 2:49:12 AM
SubDirectory	PD062525	HP Acquire Method	HP Processing Method pd061825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD089116.D	25 Jun 2025 08:58	AR\AJ	Ok
2	I.BLK	PD089117.D	25 Jun 2025 09:11	AR\AJ	Ok
3	PEM	PD089118.D	25 Jun 2025 09:25	AR\AJ	Ok,M
4	PSTDCCC050	PD089119.D	25 Jun 2025 09:39	AR\AJ	Ok
5	PCHLORCCC500	PD089120.D	25 Jun 2025 09:52	AR\AJ	Ok,M
6	PTOXCCC500	PD089121.D	25 Jun 2025 10:06	AR\AJ	Ok
7	PB168531BL	PD089122.D	25 Jun 2025 10:20	AR\AJ	Ok
8	Q2126-01	PD089123.D	25 Jun 2025 10:36	AR\AJ	Ok,M
9	Q2126-07	PD089124.D	25 Jun 2025 10:50	AR\AJ	Ok,M
10	Q2126-01	PD089125.D	25 Jun 2025 13:52	AR\AJ	Not Ok
11	Q2126-07	PD089126.D	25 Jun 2025 14:06	AR\AJ	Not Ok
12	Q2126-02	PD089127.D	25 Jun 2025 14:19	AR\AJ	Ok,M
13	Q2126-08	PD089128.D	25 Jun 2025 14:33	AR\AJ	Ok
14	Q2126-02	PD089129.D	25 Jun 2025 14:47	AR\AJ	Not Ok
15	Q2126-08	PD089130.D	25 Jun 2025 15:00	AR\AJ	Not Ok
16	Q2126-01	PD089131.D	25 Jun 2025 15:14	AR\AJ	Not Ok
17	Q2381-01DL	PD089132.D	25 Jun 2025 15:28	AR\AJ	Ok,M
18	I.BLK	PD089133.D	25 Jun 2025 15:58	AR\AJ	Ok
19	PSTDCCC050	PD089134.D	25 Jun 2025 16:12	AR\AJ	Ok
20	PCHLORCCC500	PD089135.D	25 Jun 2025 16:46	AR\AJ	Ok,M
21	PB168608BL	PD089136.D	25 Jun 2025 16:59	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD062525

Review By	Abdul	Review On	6/26/2025 11:18:16 AM
Supervise By	mohammad	Supervise On	6/27/2025 2:49:12 AM
SubDirectory	PD062525	HP Acquire Method	HP Processing Method pd061825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB168608BS	PD089137.D	25 Jun 2025 17:13	AR\AJ	Ok
23	Q2392-01	PD089138.D	25 Jun 2025 17:27	AR\AJ	Ok,M
24	Q2412-01	PD089139.D	25 Jun 2025 17:40	AR\AJ	Ok,M
25	Q2393-01	PD089140.D	25 Jun 2025 17:54	AR\AJ	Ok
26	Q2393-01MS	PD089141.D	25 Jun 2025 18:07	AR\AJ	Ok,M
27	Q2393-01MSD	PD089142.D	25 Jun 2025 18:21	AR\AJ	Ok
28	Q2393-03	PD089143.D	25 Jun 2025 18:35	AR\AJ	Ok
29	I.BLK	PD089144.D	25 Jun 2025 18:48	AR\AJ	Ok
30	PEM	PD089145.D	25 Jun 2025 19:02	AR\AJ	Ok,M
31	PSTDCCC050	PD089146.D	25 Jun 2025 19:15	AR\AJ	Ok
32	Q2393-05	PD089147.D	25 Jun 2025 19:57	AR\AJ	Ok,M
33	Q2393-07	PD089148.D	25 Jun 2025 20:10	AR\AJ	Ok,M
34	Q2393-09	PD089149.D	25 Jun 2025 20:24	AR\AJ	Ok,M
35	Q2393-11	PD089150.D	25 Jun 2025 20:37	AR\AJ	Ok,M
36	Q2394-01	PD089151.D	25 Jun 2025 20:51	AR\AJ	Ok
37	Q2398-03	PD089152.D	25 Jun 2025 21:05	AR\AJ	Ok
38	Q2398-06	PD089153.D	25 Jun 2025 21:18	AR\AJ	Ok,M
39	I.BLK	PD089154.D	25 Jun 2025 21:32	AR\AJ	Ok
40	PSTDCCC050	PD089155.D	25 Jun 2025 22:40	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD061825

Review By	Abdul	Review On	6/18/2025 8:40:56 AM
Supervise By	mohammad	Supervise On	6/19/2025 2:43:55 AM
SubDirectory	PD061825	HP Acquire Method	HP Processing Method pd061825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD088989.D	17 Jun 2025 14:58		AR\AJ	Ok
2	I.BLK	I.BLK	PD088990.D	17 Jun 2025 15:11		AR\AJ	Ok
3	PEM	PEM	PD088991.D	17 Jun 2025 15:25		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD088992.D	17 Jun 2025 15:39		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PD088993.D	17 Jun 2025 15:52		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD088994.D	17 Jun 2025 16:06		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD088995.D	17 Jun 2025 16:20		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD088996.D	17 Jun 2025 16:33		AR\AJ	Ok,M
9	PSTDICC005	PSTDICC005	PD088997.D	17 Jun 2025 16:47		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD088998.D	17 Jun 2025 17:00		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD088999.D	17 Jun 2025 17:14		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD089000.D	17 Jun 2025 17:28		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD089001.D	17 Jun 2025 17:41		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD089002.D	17 Jun 2025 17:55		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PD089003.D	17 Jun 2025 18:09		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD089004.D	17 Jun 2025 18:22		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD089005.D	17 Jun 2025 18:36		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD089006.D	17 Jun 2025 18:50		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD061825

Review By	Abdul	Review On	6/18/2025 8:40:56 AM
Supervise By	mohammad	Supervise On	6/19/2025 2:43:55 AM
SubDirectory	PD061825	HP Acquire Method	HP Processing Method
			pd061825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD089007.D	17 Jun 2025 19:03		AR\AJ	Ok,M
20	PSTDICV050	ICVPD061825	PD089008.D	17 Jun 2025 19:17		AR\AJ	Ok
21	PCHLORICV500	ICVPD061825CHLOR	PD089009.D	17 Jun 2025 19:31		AR\AJ	Ok
22	PTOXICV500	ICVPD061825TOX	PD089010.D	17 Jun 2025 19:44		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD062525

Review By	Abdul	Review On	6/26/2025 11:18:16 AM
Supervise By	mohammad	Supervise On	6/27/2025 2:49:12 AM
SubDirectory	PD062525	HP Acquire Method	HP Processing Method
			pd061825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089116.D	25 Jun 2025 08:58		AR\AJ	Ok
2	I.BLK	I.BLK	PD089117.D	25 Jun 2025 09:11		AR\AJ	Ok
3	PEM	PEM	PD089118.D	25 Jun 2025 09:25		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD089119.D	25 Jun 2025 09:39		AR\AJ	Ok
5	PCHLORCCC500	PCHLORCCC500	PD089120.D	25 Jun 2025 09:52		AR\AJ	Ok,M
6	PTOXCCC500	PTOXCCC500	PD089121.D	25 Jun 2025 10:06		AR\AJ	Ok
7	PB168531BL	PB168531BL	PD089122.D	25 Jun 2025 10:20		AR\AJ	Ok
8	Q2126-01	LOD-MDL-SOIL-03-QT	PD089123.D	25 Jun 2025 10:36	CHLOR LOD 25PPB SOIL	AR\AJ	Ok,M
9	Q2126-07	LOD-MDL-WATER-01-Q	PD089124.D	25 Jun 2025 10:50	CHLOR LOD 25PPB WATER	AR\AJ	Ok,M
10	Q2126-01	LOD-MDL-SOIL-03-QT	PD089125.D	25 Jun 2025 13:52	TOX end ccc missing , TOX LOD 50 PPB SOIL	AR\AJ	Not Ok
11	Q2126-07	LOD-MDL-WATER-01-Q	PD089126.D	25 Jun 2025 14:06	TOX end ccc missing , TOX LOD 50 PPB WATER	AR\AJ	Not Ok
12	Q2126-02	LOQ-SOIL-02-QT2-202	PD089127.D	25 Jun 2025 14:19	CHLOR LOQ 50 PPB SOIL	AR\AJ	Ok,M
13	Q2126-08	LOQ-WATER-02-QT2-2	PD089128.D	25 Jun 2025 14:33	CHLOR LOQ 50 PPB WATER	AR\AJ	Ok
14	Q2126-02	LOQ-SOIL-02-QT2-202	PD089129.D	25 Jun 2025 14:47	TOX LOQ 100PPB SOIL , TOX end ccc missing	AR\AJ	Not Ok
15	Q2126-08	LOQ-WATER-02-QT2-2	PD089130.D	25 Jun 2025 15:00	TOX LOQ 100PPB WATER , TOX end ccc missing	AR\AJ	Not Ok
16	Q2126-01	LOD-MDL-SOIL-03-QT	PD089131.D	25 Jun 2025 15:14	PEST LOD 1PPB SOIL , U Flag in delta-BHC and Methoxychlor	AR\AJ	Not Ok
17	Q2381-01DL	291431DL	PD089132.D	25 Jun 2025 15:28		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD062525

Review By	Abdul	Review On	6/26/2025 11:18:16 AM
Supervise By	mohammad	Supervise On	6/27/2025 2:49:12 AM
SubDirectory	PD062525	HP Acquire Method	HP Processing Method
			pd061825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	I.BLK	I.BLK	PD089133.D	25 Jun 2025 15:58		AR\AJ	Ok
19	PSTDCCC050	PSTDCCC050	PD089134.D	25 Jun 2025 16:12		AR\AJ	Ok
20	PCHLORCCC500	PCHLORCCC500	PD089135.D	25 Jun 2025 16:46		AR\AJ	Ok,M
21	PB168608BL	PB168608BL	PD089136.D	25 Jun 2025 16:59		AR\AJ	Ok
22	PB168608BS	PB168608BS	PD089137.D	25 Jun 2025 17:13		AR\AJ	Ok
23	Q2392-01	S-1	PD089138.D	25 Jun 2025 17:27		AR\AJ	Ok,M
24	Q2412-01	TR-05-062425	PD089139.D	25 Jun 2025 17:40		AR\AJ	Ok,M
25	Q2393-01	PARK AVE -1	PD089140.D	25 Jun 2025 17:54		AR\AJ	Ok
26	Q2393-01MS	PARK AVE -1MS	PD089141.D	25 Jun 2025 18:07		AR\AJ	Ok,M
27	Q2393-01MSD	PARK AVE -1MSD	PD089142.D	25 Jun 2025 18:21		AR\AJ	Ok
28	Q2393-03	PARK AVE -2	PD089143.D	25 Jun 2025 18:35		AR\AJ	Ok
29	I.BLK	I.BLK	PD089144.D	25 Jun 2025 18:48		AR\AJ	Ok
30	PEM	PEM	PD089145.D	25 Jun 2025 19:02		AR\AJ	Ok,M
31	PSTDCCC050	PSTDCCC050	PD089146.D	25 Jun 2025 19:15		AR\AJ	Ok
32	Q2393-05	PARK AVE -3	PD089147.D	25 Jun 2025 19:57		AR\AJ	Ok,M
33	Q2393-07	PARK AVE -4	PD089148.D	25 Jun 2025 20:10		AR\AJ	Ok,M
34	Q2393-09	PARK AVE -5	PD089149.D	25 Jun 2025 20:24		AR\AJ	Ok,M
35	Q2393-11	PARK AVE -6	PD089150.D	25 Jun 2025 20:37		AR\AJ	Ok,M
36	Q2394-01	MH-K/L	PD089151.D	25 Jun 2025 20:51		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD062525

Review By	Abdul	Review On	6/26/2025 11:18:16 AM
Supervise By	mohammad	Supervise On	6/27/2025 2:49:12 AM
SubDirectory	PD062525	HP Acquire Method	HP Processing Method pd061825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	Q2398-03	M00-25-0179	PD089152.D	25 Jun 2025 21:05		AR\AJ	Ok
38	Q2398-06	ARS20-0036	PD089153.D	25 Jun 2025 21:18		AR\AJ	Ok,M
39	I.BLK	I.BLK	PD089154.D	25 Jun 2025 21:32	DCB Low in 2nd column	AR\AJ	Ok
40	PSTDCCC050	PSTDCCC050	PD089155.D	25 Jun 2025 22:40	Methoxychlor low in 1t column	AR\AJ	Ok,M

M : Manual Integration

SOP ID:	M3541-ASE Extraction-14		
Clean Up SOP #:	Florisol	Extraction Start Date :	06/25/2025
Matrix :	Solid	Extraction Start Time :	08:45
Weigh By:	EH	Extraction By:	RJ
Balance check:	RJ	Filter By:	RJ
Balance ID:	EX-SC-2	pH Meter ID:	N/A
pH Strip Lot#:	N/A	Hood ID:	3,7
		Concentration By:	EH
		Supervisor By :	RUPESH

Extraction Method:
☐ Separatory Funnel
 ☐ Continuous Liquid/Liquid
 ☐ Sonication
 ☐ Waste Dilution
 ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP24627
Surrogate	1.0ML	200 PPB	PP24597
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2613
Baked Na2SO4	N/A	EP2622
Sand	N/A	E2865
Hexane	N/A	E3945
Florisol	N/A	E3927
9:1 Hexane:Acetone Mixture	N/A	EP2596
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS723.

KD Bath ID:	N/A	Envap ID:	NE VAP-02
KD Bath Temperature:	N/A	Envap Temperature:	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
6/25/25	RS (Ext Lab)	RS (Ext Lab)
12:05	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 06/25/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB168608BL	PBLK608	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10			U2-1
PB168608BS	PLCS608	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10			2
Q2392-01	S-1	Pesticide-TCL	30.07	N/A	ritesh	Evelyn	10	B		3
Q2393-01	PARK AVE -1	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	D		4
Q2393-01MS	PARK AVE -1MS	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	D		5
Q2393-01MS D	PARK AVE -1MSD	Pesticide-TCL	30.10	N/A	ritesh	Evelyn	10	D		6
Q2393-03	PARK AVE -2	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	D		U3-1
Q2393-05	PARK AVE -3	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	D		2
Q2393-07	PARK AVE -4	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10	D		3
Q2393-09	PARK AVE -5	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10	D		4
Q2393-11	PARK AVE -6	Pesticide-TCL	30.08	N/A	ritesh	Evelyn	10	D		5
Q2394-01	MH-K/L	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	D		6
Q2398-03	M00-25-0179	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	D	Small Partical	U6-1
Q2398-06	ARS20-0036	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10	D		2
Q2399-01	TP-13	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	D		3
Q2399-05	EP-7	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	D		4
Q2403-03	LAW-25-0093	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	D	Small Partical	5
Q2403-07	ARS 20-006	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	D		6
Q2405-01	MH-M/H	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10	D		U1-1
Q2412-01	TR-05-062425	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	D		2



RS
6/25

108608
8:40

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2393P WorkList ID : 190368 Department : Extraction Date : 06-25-2025 08:40:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2392-01	S-1	Solid	Pesticide-TCL	Cool 4 deg C	EARTH03	A41	06/23/2025	8081B
Q2393-01	PARK AVE -1	Solid	Pesticide-TCL	Cool 4 deg C	UNIO02	A22	06/23/2025	8081B
Q2393-03	PARK AVE -2	Solid	Pesticide-TCL	Cool 4 deg C	UNIO02	A22	06/23/2025	8081B
Q2393-05	PARK AVE -3	Solid	Pesticide-TCL	Cool 4 deg C	UNIO02	A22	06/23/2025	8081B
Q2393-07	PARK AVE -4	Solid	Pesticide-TCL	Cool 4 deg C	UNIO02	A22	06/23/2025	8081B
Q2393-09	PARK AVE -5	Solid	Pesticide-TCL	Cool 4 deg C	UNIO02	A22	06/23/2025	8081B
Q2393-11	PARK AVE -6	Solid	Pesticide-TCL	Cool 4 deg C	UNIO02	A22	06/23/2025	8081B
Q2394-01	MH-K/L	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	A21	06/23/2025	8081B
Q2398-03	M00-25-0179	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	A33	06/23/2025	8081B
Q2398-06	ARS20-0036	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	A33	06/23/2025	8081B
Q2399-01	TP-13	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	A41	06/23/2025	8081B
Q2399-05	EP-7	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	A41	06/23/2025	8081B
Q2403-03	LAW-25-0093	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	A32	06/23/2025	8081B
Q2403-07	ARS 20-006	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03		06/23/2025	8081B
Q2405-01	MH-M/H	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	06/24/2025	8081B
Q2412-01	TR-05-062425	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	D41	06/24/2025	8081B

Date/Time 06/25/25 8:40
Raw Sample Received by: RS (LAW-25-0093)
Raw Sample Relinquished by: CP SM

Date/Time 06/25/25 9:10
Raw Sample Received by: OP SM
Raw Sample Relinquished by: RS (LAW-25-0093)



LAB CHRONICLE

OrderID:	Q2393	OrderDate:	6/23/2025 11:43:58 AM
Client:	Union Paving & Construction Company	Project:	190 Park Ave Hanover Twp NJ - PO 2556-001
Contact:	Gerard Busacco	Location:	A22,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2393-01	PARK AVE -1	SOIL	Pesticide-TCL	8081B	06/23/25	06/25/25	06/25/25	06/23/25
Q2393-03	PARK AVE -2	SOIL	Pesticide-TCL	8081B	06/23/25	06/25/25	06/25/25	06/23/25
Q2393-05	PARK AVE -3	SOIL	Pesticide-TCL	8081B	06/23/25	06/25/25	06/25/25	06/23/25
Q2393-07	PARK AVE -4	SOIL	Pesticide-TCL	8081B	06/23/25	06/25/25	06/25/25	06/23/25
Q2393-09	PARK AVE -5	SOIL	Pesticide-TCL	8081B	06/23/25	06/25/25	06/25/25	06/23/25
Q2393-11	PARK AVE -6	SOIL	Pesticide-TCL	8081B	06/23/25	06/25/25	06/25/25	06/23/25