

Surrogate Summary

SDG No.: Q2393

Client: Union Paving & Construction Company

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088990.D	PIBLK-PD088990.D	Decachlorobiphenyl	1	20	18.2	91		57	171
		Tetrachloro-m-xylene	1	20	16.1	80		61	148
		Decachlorobiphenyl	2	20	18.6	93		57	171
		Tetrachloro-m-xylene	2	20	17.7	88		61	148
I.BLK-PD089133.D	PIBLK-PD089133.D	Decachlorobiphenyl	1	20	19.9	100		57	171
		Tetrachloro-m-xylene	1	20	19.2	96		61	148
		Decachlorobiphenyl	2	20	20.8	104		57	171
		Tetrachloro-m-xylene	2	20	21.0	105		61	148
PB168608BL	PB168608BL	Decachlorobiphenyl	1	20	18.8	94		20	144
		Tetrachloro-m-xylene	1	20	17.1	86		19	148
		Decachlorobiphenyl	2	20	19.3	96		20	144
		Tetrachloro-m-xylene	2	20	18.8	94		19	148
PB168608BS	PB168608BS	Decachlorobiphenyl	1	20	23.6	118		20	144
		Tetrachloro-m-xylene	1	20	21.2	106		19	148
		Decachlorobiphenyl	2	20	24.3	121		20	144
		Tetrachloro-m-xylene	2	20	23.2	116		19	148
Q2393-01	PARK AVE -1	Decachlorobiphenyl	1	20	17.6	88		20	144
		Tetrachloro-m-xylene	1	20	17.8	89		19	148
		Decachlorobiphenyl	2	20	16.4	82		20	144
		Tetrachloro-m-xylene	2	20	18.9	95		19	148
Q2393-01MS	PARK AVE -1MS	Decachlorobiphenyl	1	20	16.5	82		20	144
		Tetrachloro-m-xylene	1	20	15.3	77		19	148
		Decachlorobiphenyl	2	20	15.6	78		20	144
		Tetrachloro-m-xylene	2	20	17.1	86		19	148
Q2393-01MSD	PARK AVE -1MSD	Decachlorobiphenyl	1	20	16.4	82		20	144
		Tetrachloro-m-xylene	1	20	15.5	77		19	148
		Decachlorobiphenyl	2	20	16.0	80		20	144
		Tetrachloro-m-xylene	2	20	17.2	86		19	148
Q2393-03	PARK AVE -2	Decachlorobiphenyl	1	20	15.8	79		20	144
		Tetrachloro-m-xylene	1	20	16.6	83		19	148
		Decachlorobiphenyl	2	20	15.8	79		20	144
		Tetrachloro-m-xylene	2	20	17.5	87		19	148
I.BLK-PD089144.D	PIBLK-PD089144.D	Decachlorobiphenyl	1	20	20.7	104		57	171
		Tetrachloro-m-xylene	1	20	19.6	98		61	148
		Decachlorobiphenyl	2	20	20.8	104		57	171
		Tetrachloro-m-xylene	2	20	21.0	105		61	148
Q2393-05	PARK AVE -3	Decachlorobiphenyl	1	20	14.5	72		20	144
		Tetrachloro-m-xylene	1	20	15.1	76		19	148
		Decachlorobiphenyl	2	20	13.2	66		20	144
		Tetrachloro-m-xylene	2	20	15.9	79		19	148
Q2393-07	PARK AVE -4	Decachlorobiphenyl	1	20	11.1	56		20	144

Surrogate Summary

SDG No.: Q2393

Client: Union Paving & Construction Company

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2393-07	PARK AVE -4	Tetrachloro-m-xylene	1	20	11.8	59		19	148
		Decachlorobiphenyl	2	20	9.59	48		20	144
Q2393-09	PARK AVE -5	Tetrachloro-m-xylene	2	20	12.5	63		19	148
		Decachlorobiphenyl	1	20	9.08	45		20	144
		Tetrachloro-m-xylene	1	20	10.2	51		19	148
		Decachlorobiphenyl	2	20	8.44	42		20	144
		Tetrachloro-m-xylene	2	20	10.7	54		19	148
Q2393-11	PARK AVE -6	Decachlorobiphenyl	1	20	10.4	52		20	144
		Tetrachloro-m-xylene	1	20	10.5	52		19	148
		Decachlorobiphenyl	2	20	8.52	43		20	144
		Tetrachloro-m-xylene	2	20	11.1	55		19	148
I.BLK-PD089154.D	PIBLK-PD089154.D	Decachlorobiphenyl	1	20	15.6	78		57	171
		Tetrachloro-m-xylene	1	20	17.6	88		61	148
		Decachlorobiphenyl	2	20	10.8	54	*	57	171
		Tetrachloro-m-xylene	2	20	21.2	106		61	148