

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: UNIO02

Lab Code: CHEM

SAS No.: Q2393 SDG NO.: Q2393

Lab File ID: BF142810.D

DFTPP Injection Date: 06/24/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	34.4
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.2 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142811.D	06/24/2025	10:16
PARK AVE -6	Q2393-11	BF142823.D	06/24/2025	17:01
PARK AVE -3	Q2393-05	BF142824.D	06/24/2025	17:33
PARK AVE -4	Q2393-07	BF142825.D	06/24/2025	18:03
PARK AVE -5	Q2393-09	BF142826.D	06/24/2025	18:34