

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: UNIO02

Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG No.: Q2393

Instrument ID: MSVOA_Y Calibration Date(s): 06/23/2025 06/23/2025

Heated Purge: (Y/N) Y Calibration Time(s): 13:38 15:31

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF005 = VY022776.D			RRF010 = VY022777.D		RRF020 = VY022778.D			
	RRF050 = VY022779.D			RRF100 = VY022780.D		RRF150 = VY022781.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Dichlorodifluoromethane	0.424	0.456	0.474	0.424	0.404	0.384	0.428	7.7	
Chloromethane	0.837	0.921	0.865	0.793	0.758	0.724	0.816	8.9	
Vinyl Chloride	0.934	1.099	1.091	1.045	0.993	0.958	1.020	6.8	
Bromomethane	0.784	0.885	0.854	0.771	0.760	0.756	0.802	6.8	
Chloroethane	0.649	0.736	0.722	0.694	0.673	0.640	0.686	5.6	
Trichlorofluoromethane	0.999	1.180	1.219	1.166	1.127	1.085	1.129	7	
1,1,2-Trichlorotrifluoroethane	0.508	0.560	0.547	0.515	0.492	0.474	0.516	6.3	
1,1-Dichloroethene	0.478	0.539	0.524	0.514	0.500	0.483	0.506	4.7	
Acetone	0.117	0.124	0.114	0.095	0.096	0.087	0.105	13.9	
Carbon Disulfide	1.516	1.705	1.731	1.667	1.625	1.566	1.635	5.1	
Methyl tert-butyl Ether	1.173	1.398	1.396	1.435	1.460	1.405	1.378	7.5	
Methyl Acetate	0.272	0.358	0.440	0.351	0.353	0.322	0.349	15.7	
Methylene Chloride	0.840	0.777	0.664	0.590	0.578	0.548	0.666	17.7	
trans-1,2-Dichloroethene	0.521	0.604	0.597	0.592	0.581	0.575	0.578	5.2	
1,1-Dichloroethane	0.949	1.075	1.079	1.077	1.055	1.030	1.044	4.8	
Cyclohexane	0.998	1.021	0.988	0.946	0.905	0.894	0.959	5.4	
2-Butanone	0.145	0.160	0.160	0.153	0.156	0.147	0.154	4.4	
Carbon Tetrachloride	0.439	0.498	0.507	0.491	0.492	0.491	0.486	5	
cis-1,2-Dichloroethene	0.606	0.689	0.687	0.685	0.687	0.678	0.672	4.8	
Bromochloromethane	0.437	0.431	0.437	0.459	0.443	0.427	0.439	2.6	
Chloroform	0.986	1.130	1.099	1.096	1.084	1.059	1.076	4.6	
1,1,1-Trichloroethane	0.847	0.945	0.973	0.950	0.939	0.923	0.929	4.7	
Methylcyclohexane	0.543	0.589	0.610	0.618	0.608	0.611	0.596	4.7	
Benzene	1.248	1.433	1.451	1.464	1.467	1.440	1.417	5.9	
1,2-Dichloroethane	0.335	0.397	0.402	0.400	0.404	0.392	0.388	6.8	
Trichloroethene	0.305	0.364	0.382	0.372	0.360	0.350	0.356	7.6	
1,2-Dichloropropane	0.289	0.339	0.345	0.339	0.341	0.337	0.332	6.4	
Bromodichloromethane	0.422	0.495	0.496	0.498	0.504	0.498	0.485	6.4	
4-Methyl-2-Pentanone	0.168	0.201	0.215	0.226	0.230	0.221	0.210	10.9	
Toluene	0.747	0.873	0.908	0.926	0.955	0.954	0.894	8.8	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: UNIO02
Lab Code: CHEM Case No.: Q2393 SAS No.: Q2393 SDG No.: Q2393
Instrument ID: MSVOA_Y Calibration Date(s): 06/23/2025 06/23/2025
Heated Purge: (Y/N) Y Calibration Time(s): 13:38 15:31
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF005 = VY022776.D		RRF010 = VY022777.D		RRF020 = VY022778.D		RRF050 = VY022779.D		RRF100 = VY022780.D		RRF150 = VY022781.D	
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD				
t-1,3-Dichloropropene	0.355	0.430	0.438	0.451	0.473	0.473	0.437	10				
cis-1,3-Dichloropropene	0.412	0.503	0.523	0.524	0.540	0.538	0.506	9.6				
1,1,2-Trichloroethane	0.207	0.249	0.249	0.253	0.255	0.250	0.244	7.4				
2-Hexanone	0.115	0.140	0.145	0.151	0.157	0.149	0.143	10.5				
Dibromochloromethane	0.260	0.315	0.321	0.329	0.336	0.329	0.315	8.8				
1,2-Dibromoethane	0.193	0.231	0.229	0.237	0.244	0.236	0.228	7.8				
Tetrachloroethene	0.399	0.465	0.535	0.515	0.473	0.446	0.472	10.3				
Chlorobenzene	0.981	1.110	1.131	1.126	1.130	1.114	1.099	5.3				
Ethyl Benzene	1.644	1.881	1.971	2.029	2.040	2.018	1.930	7.9				
m/p-Xylenes	0.624	0.722	0.759	0.782	0.800	0.791	0.746	8.8				
o-Xylene	0.578	0.674	0.708	0.734	0.759	0.765	0.703	10				
Styrene	0.926	1.108	1.165	1.249	1.309	1.309	1.178	12.5				
Bromoform	0.178	0.204	0.203	0.212	0.225	0.220	0.207	8				
Isopropylbenzene	3.354	3.764	3.823	3.778	3.709	3.759	3.698	4.7				
1,1,2,2-Tetrachloroethane	0.597	0.659	0.566	0.567	0.594	0.593	0.596	5.6				
1,3-Dichlorobenzene	1.546	1.660	1.692	1.708	1.750	1.744	1.683	4.5				
1,4-Dichlorobenzene	1.564	1.740	1.688	1.685	1.690	1.666	1.672	3.5				
1,2-Dichlorobenzene	1.395	1.488	1.502	1.499	1.515	1.502	1.483	3				
1,2-Dibromo-3-Chloropropane	0.102	0.101	0.103	0.103	0.102	0.096	0.101	2.7				
1,2,4-Trichlorobenzene	0.778	0.841	0.848	0.843	0.871	0.845	0.838	3.7				
1,2,3-Trichlorobenzene	0.679	0.723	0.735	0.728	0.751	0.727	0.724	3.3				
1,2-Dichloroethane-d4	0.568	0.550	0.557	0.559	0.571	0.545	0.558	1.8				
Dibromofluoromethane	0.306	0.297	0.295	0.304	0.314	0.308	0.304	2.3				
Toluene-d8	1.182	1.148	1.186	1.215	1.262	1.247	1.207	3.6				
4-Bromofluorobenzene	0.368	0.362	0.370	0.385	0.423	0.421	0.388	7				

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.