

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY062325

Review By	Mahesh Dadoda	Review On	6/24/2025 8:24:26 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:29:32 AM		
SubDirectory	VY062325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP134461			
Initial Calibration Stds		VP134462,VP134463,VP134464,VP134465,VP134466,VP134467			
CCC					
Internal Standard/PEM		VP133934			
ICV/I.BLK		VP134468			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY022775.D	23 Jun 2025 10:17	SY/MD	Ok
2	VSTDIC005	VY022776.D	23 Jun 2025 13:38	SY/MD	Ok,M
3	VSTDIC010	VY022777.D	23 Jun 2025 14:00	SY/MD	Ok,M
4	VSTDIC020	VY022778.D	23 Jun 2025 14:23	SY/MD	Ok,M
5	VSTDIC050	VY022779.D	23 Jun 2025 14:46	SY/MD	Ok,M
6	VSTDIC100	VY022780.D	23 Jun 2025 15:08	SY/MD	Ok,M
7	VSTDIC150	VY022781.D	23 Jun 2025 15:31	SY/MD	Ok,M
8	VIBLK	VY022782.D	23 Jun 2025 15:54	SY/MD	Ok
9	VSTDICV050	VY022783.D	23 Jun 2025 16:17	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY062425

Review By	Mahesh Dadoda	Review On	6/25/2025 12:32:22 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/25/2025 12:35:55 PM		
SubDirectory	VY062425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134483			
CCC		VP134483,VP134485			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY022784.D	24 Jun 2025 08:49	SY/MD	Ok
2	VSTDCCC050	VY022785.D	24 Jun 2025 09:20	SY/MD	Ok,M
3	VY0624SBL01	VY022786.D	24 Jun 2025 10:25	SY/MD	Ok
4	VY0624SBS01	VY022787.D	24 Jun 2025 10:58	SY/MD	Ok,M
5	VY0624SBSD01	VY022788.D	24 Jun 2025 11:21	SY/MD	Ok,M
6	Q2386-01	VY022789.D	24 Jun 2025 11:58	SY/MD	ReRun
7	Q2383-01RE	VY022790.D	24 Jun 2025 12:22	SY/MD	Not Ok
8	Q2384-01RE	VY022791.D	24 Jun 2025 12:45	SY/MD	Confirms
9	Q2355-03	VY022792.D	24 Jun 2025 13:09	SY/MD	Not Ok
10	Q2363-01	VY022793.D	24 Jun 2025 13:32	SY/MD	Dilution
11	Q2392-02	VY022794.D	24 Jun 2025 13:55	SY/MD	Ok
12	Q2393-02	VY022795.D	24 Jun 2025 14:19	SY/MD	Ok
13	Q2393-04	VY022796.D	24 Jun 2025 14:42	SY/MD	Ok
14	Q2393-06	VY022797.D	24 Jun 2025 15:06	SY/MD	Ok
15	Q2393-08	VY022798.D	24 Jun 2025 15:29	SY/MD	Ok
16	Q2393-10	VY022799.D	24 Jun 2025 15:52	SY/MD	Ok
17	Q2393-12	VY022800.D	24 Jun 2025 16:16	SY/MD	Ok
18	Q2394-03	VY022801.D	24 Jun 2025 16:39	SY/MD	Ok
19	Q2399-03	VY022802.D	24 Jun 2025 17:03	SY/MD	Ok
20	Q2399-07	VY022803.D	24 Jun 2025 17:26	SY/MD	Ok
21	Q2403-03	VY022804.D	24 Jun 2025 17:50	SY/MD	ReRun

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY062425

Review By	Maresh Dadoda	Review On	6/25/2025 12:32:22 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/25/2025 12:35:55 PM		
SubDirectory	VY062425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134483			
CCC		VP134483,VP134485			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q2403-07	VY022805.D	24 Jun 2025 18:13	SY/MD	Ok
23	Q2398-03	VY022806.D	24 Jun 2025 18:36	SY/MD	Ok
24	Q2398-06	VY022807.D	24 Jun 2025 19:00	SY/MD	Not Ok
25	VIBLK	VY022808.D	24 Jun 2025 19:23	SY/MD	Ok
26	VSTDCCC050	VY022809.D	24 Jun 2025 20:09	SY/MD	Ok,M

M : Manual Integration