

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062325

Review By	Maresh Dadoda	Review On	6/24/2025 8:24:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:29:32 AM
SubDirectory	VY062325	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134461
Initial Calibration Stds	VP134462,VP134463,VP134464,VP134465,VP134466,VP134467
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134468
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022775.D	23 Jun 2025 10:17		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY022776.D	23 Jun 2025 13:38	LR- 58	SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY022777.D	23 Jun 2025 14:00		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY022778.D	23 Jun 2025 14:23		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY022779.D	23 Jun 2025 14:46		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY022780.D	23 Jun 2025 15:08		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY022781.D	23 Jun 2025 15:31		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022782.D	23 Jun 2025 15:54		SY/MD	Ok
9	VSTDICV050	ICVVY062325	VY022783.D	23 Jun 2025 16:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062425

Review By	Mahesh Dadoda	Review On	6/25/2025 12:32:22 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/25/2025 12:35:55 PM
SubDirectory	VY062425	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134483
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134483,VP134485 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022784.D	24 Jun 2025 08:49		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022785.D	24 Jun 2025 09:20		SY/MD	Ok,M
3	VY0624SBL01	VY0624SBL01	VY022786.D	24 Jun 2025 10:25		SY/MD	Ok
4	VY0624SBS01	VY0624SBS01	VY022787.D	24 Jun 2025 10:58		SY/MD	Ok,M
5	VY0624SBSD01	VY0624SBSD01	VY022788.D	24 Jun 2025 11:21		SY/MD	Ok,M
6	Q2386-01	SU-1-062025	VY022789.D	24 Jun 2025 11:58	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q2383-01RE	RT-5376RE	VY022790.D	24 Jun 2025 12:22	vial-B Internal standard fail;Not match with first run	SY/MD	Not Ok
8	Q2384-01RE	OK-01-6202025RE	VY022791.D	24 Jun 2025 12:45	vial-B Internal Standard Fail	SY/MD	Confirms
9	Q2355-03	20071	VY022792.D	24 Jun 2025 13:09	vial-A not purge	SY/MD	Not Ok
10	Q2363-01	GCAP1	VY022793.D	24 Jun 2025 13:32	vial-A Surrogate fail ;Need MeOH	SY/MD	Dilution
11	Q2392-02	S-1A	VY022794.D	24 Jun 2025 13:55	vial-A	SY/MD	Ok
12	Q2393-02	PARK AVE -1	VY022795.D	24 Jun 2025 14:19	vial-A	SY/MD	Ok
13	Q2393-04	PARK AVE -2	VY022796.D	24 Jun 2025 14:42	vial-A	SY/MD	Ok
14	Q2393-06	PARK AVE -3	VY022797.D	24 Jun 2025 15:06	vial-A	SY/MD	Ok
15	Q2393-08	PARK AVE -4	VY022798.D	24 Jun 2025 15:29	vial-A	SY/MD	Ok
16	Q2393-10	PARK AVE -5	VY022799.D	24 Jun 2025 15:52	vial-A	SY/MD	Ok
17	Q2393-12	PARK AVE -6	VY022800.D	24 Jun 2025 16:16	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062425

Review By	Mahesh Dadoda	Review On	6/25/2025 12:32:22 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/25/2025 12:35:55 PM		
SubDirectory	VY062425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134483			
CCC		VP134483,VP134485			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2394-03	MH-K/L VOC	VY022801.D	24 Jun 2025 16:39	vial-A	SY/MD	Ok
19	Q2399-03	TP-13-VOC	VY022802.D	24 Jun 2025 17:03	vial-A	SY/MD	Ok
20	Q2399-07	EP-7-VOC	VY022803.D	24 Jun 2025 17:26	vial-A	SY/MD	Ok
21	Q2403-03	LAW-25-0093	VY022804.D	24 Jun 2025 17:50	vial-A Internal Standard Fail	SY/MD	ReRun
22	Q2403-07	ARS 20-0006	VY022805.D	24 Jun 2025 18:13	vial-A	SY/MD	Ok
23	Q2398-03	M00-25-0179	VY022806.D	24 Jun 2025 18:36	vial-A	SY/MD	Ok
24	Q2398-06	ARS20-0036	VY022807.D	24 Jun 2025 19:00	vial-A Not purge	SY/MD	Not Ok
25	VIBLK	VIBLK	VY022808.D	24 Jun 2025 19:23		SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY022809.D	24 Jun 2025 20:09		SY/MD	Ok,M

M : Manual Integration