

Surrogate Summary

SDG No.: **Q2393**

Client: **Union Paving & Construction Company**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2393-02	PARK AVE -1	1,2-Dichloroethane-d4	50	48.6	97	63	155
		Dibromofluoromethane	50	49.0	98	70	134
		Toluene-d8	50	49.6	99	74	123
		4-Bromofluorobenzene	50	57.6	115	17	146
Q2393-04	PARK AVE -2	1,2-Dichloroethane-d4	50	55.8	112	63	155
		Dibromofluoromethane	50	52.0	104	70	134
		Toluene-d8	50	50.0	100	74	123
		4-Bromofluorobenzene	50	61.1	122	17	146
Q2393-06	PARK AVE -3	1,2-Dichloroethane-d4	50	48.8	98	63	155
		Dibromofluoromethane	50	49.3	99	70	134
		Toluene-d8	50	48.4	97	74	123
		4-Bromofluorobenzene	50	56.5	113	17	146
Q2393-08	PARK AVE -4	1,2-Dichloroethane-d4	50	48.2	96	63	155
		Dibromofluoromethane	50	48.7	97	70	134
		Toluene-d8	50	49.5	99	74	123
		4-Bromofluorobenzene	50	55.2	110	17	146
Q2393-10	PARK AVE -5	1,2-Dichloroethane-d4	50	46.8	94	63	155
		Dibromofluoromethane	50	49.1	98	70	134
		Toluene-d8	50	48.5	97	74	123
		4-Bromofluorobenzene	50	51.6	103	17	146
Q2393-12	PARK AVE -6	1,2-Dichloroethane-d4	50	50.0	100	63	155
		Dibromofluoromethane	50	50.2	100	70	134
		Toluene-d8	50	49.4	99	74	123
		4-Bromofluorobenzene	50	54.2	108	17	146
VY0624SBL01	VY0624SBL01	1,2-Dichloroethane-d4	50	54.5	109	63	155
		Dibromofluoromethane	50	52.4	105	70	134
		Toluene-d8	50	49.2	98	74	123
		4-Bromofluorobenzene	50	42.8	86	17	146
VY0624SBS01	VY0624SBS01	1,2-Dichloroethane-d4	50	50.7	101	63	155
		Dibromofluoromethane	50	51.1	102	70	134
		Toluene-d8	50	50.7	101	74	123
		4-Bromofluorobenzene	50	48.5	97	17	146
VY0624SBSD01	VY0624SBSD01	1,2-Dichloroethane-d4	50	48.6	97	63	155
		Dibromofluoromethane	50	49.7	99	70	134
		Toluene-d8	50	50.5	101	74	123
		4-Bromofluorobenzene	50	47.7	95	17	146