

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2393

Client: Union Paving & Construction Company

Analytical Method: SW8260D

Datafile : VY022788.D

| Lab Sample ID | Parameter                      | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|---------------|--------------------------------|-------|--------|-------|-----|-----|------|-----|----------------|-----|
| VY0624SBSD01  | Dichlorodifluoromethane        | 20    | 21.6   | ug/Kg | 108 | 4   |      | 64  | 136            | 20  |
|               | Chloromethane                  | 20    | 19.9   | ug/Kg | 100 | 2   |      | 52  | 151            | 20  |
|               | Vinyl chloride                 | 20    | 20.6   | ug/Kg | 103 | 1   |      | 56  | 148            | 20  |
|               | Bromomethane                   | 20    | 20.6   | ug/Kg | 103 | 4   |      | 58  | 141            | 20  |
|               | Chloroethane                   | 20    | 20.0   | ug/Kg | 100 | 2   |      | 69  | 130            | 20  |
|               | Trichlorofluoromethane         | 20    | 20.7   | ug/Kg | 104 | 3   |      | 69  | 134            | 20  |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 21.0   | ug/Kg | 105 | 2   |      | 81  | 123            | 20  |
|               | 1,1-Dichloroethene             | 20    | 20.0   | ug/Kg | 100 | 1   |      | 79  | 121            | 20  |
|               | Acetone                        | 100   | 94.2   | ug/Kg | 94  | 6   |      | 40  | 171            | 20  |
|               | Carbon disulfide               | 20    | 20.5   | ug/Kg | 103 | 0   |      | 59  | 130            | 20  |
|               | Methyl tert-butyl Ether        | 20    | 19.1   | ug/Kg | 96  | 5   |      | 77  | 129            | 20  |
|               | Methyl Acetate                 | 20    | 17.5   | ug/Kg | 88  | 7   |      | 69  | 149            | 20  |
|               | Methylene Chloride             | 20    | 17.9   | ug/Kg | 90  | 2   |      | 72  | 131            | 20  |
|               | trans-1,2-Dichloroethene       | 20    | 19.8   | ug/Kg | 99  | 2   |      | 80  | 123            | 20  |
|               | 1,1-Dichloroethane             | 20    | 20.3   | ug/Kg | 102 | 1   |      | 82  | 123            | 20  |
|               | Cyclohexane                    | 20    | 20.2   | ug/Kg | 101 | 0   |      | 76  | 122            | 20  |
|               | 2-Butanone                     | 100   | 93.6   | ug/Kg | 94  | 6   |      | 69  | 131            | 20  |
|               | Carbon Tetrachloride           | 20    | 20.6   | ug/Kg | 103 | 3   |      | 76  | 129            | 20  |
|               | cis-1,2-Dichloroethene         | 20    | 19.5   | ug/Kg | 98  | 2   |      | 82  | 123            | 20  |
|               | Bromochloromethane             | 20    | 20.2   | ug/Kg | 101 | 2   |      | 80  | 127            | 20  |
|               | Chloroform                     | 20    | 19.7   | ug/Kg | 99  | 1   |      | 82  | 125            | 20  |
|               | 1,1,1-Trichloroethane          | 20    | 20.4   | ug/Kg | 102 | 1   |      | 80  | 126            | 20  |
|               | Methylcyclohexane              | 20    | 20.3   | ug/Kg | 102 | 1   |      | 77  | 123            | 20  |
|               | Benzene                        | 20    | 20.2   | ug/Kg | 101 | 0   |      | 84  | 121            | 20  |
|               | 1,2-Dichloroethane             | 20    | 20.3   | ug/Kg | 102 | 1   |      | 81  | 126            | 20  |
|               | Trichloroethene                | 20    | 20.6   | ug/Kg | 103 | 3   |      | 83  | 122            | 20  |
|               | 1,2-Dichloropropane            | 20    | 20.0   | ug/Kg | 100 | 2   |      | 83  | 122            | 20  |
|               | Bromodichloromethane           | 20    | 19.9   | ug/Kg | 100 | 2   |      | 82  | 123            | 20  |
|               | 4-Methyl-2-Pentanone           | 100   | 95.0   | ug/Kg | 95  | 5   |      | 70  | 135            | 20  |
|               | Toluene                        | 20    | 20.0   | ug/Kg | 100 | 1   |      | 83  | 122            | 20  |
|               | t-1,3-Dichloropropene          | 20    | 19.3   | ug/Kg | 97  | 1   |      | 78  | 124            | 20  |
|               | cis-1,3-Dichloropropene        | 20    | 19.6   | ug/Kg | 98  | 3   |      | 81  | 122            | 20  |
|               | 1,1,2-Trichloroethane          | 20    | 19.8   | ug/Kg | 99  | 3   |      | 82  | 125            | 20  |
|               | 2-Hexanone                     | 100   | 90.5   | ug/Kg | 91  | 8   |      | 66  | 138            | 20  |
|               | Dibromochloromethane           | 20    | 19.6   | ug/Kg | 98  | 0   |      | 79  | 125            | 20  |
|               | 1,2-Dibromoethane              | 20    | 19.1   | ug/Kg | 96  | 4   |      | 80  | 125            | 20  |
|               | Tetrachloroethene              | 20    | 21.6   | ug/Kg | 108 | 15  |      | 83  | 125            | 20  |
|               | Chlorobenzene                  | 20    | 20.1   | ug/Kg | 101 | 2   |      | 84  | 122            | 20  |
|               | Ethyl Benzene                  | 20    | 20.1   | ug/Kg | 101 | 3   |      | 82  | 124            | 20  |
|               | m/p-Xylenes                    | 40    | 39.5   | ug/Kg | 99  | 1   |      | 83  | 124            | 20  |
|               | o-Xylene                       | 20    | 19.3   | ug/Kg | 97  | 2   |      | 83  | 123            | 20  |
|               | Styrene                        | 20    | 19.3   | ug/Kg | 97  | 1   |      | 82  | 124            | 20  |
|               | Bromoform                      | 20    | 19.0   | ug/Kg | 95  | 1   |      | 75  | 127            | 20  |
|               | Isopropylbenzene               | 20    | 20.5   | ug/Kg | 103 | 2   |      | 82  | 124            | 20  |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 18.7   | ug/Kg | 94  | 15  |      | 77  | 127            | 20  |
|               | 1,3-Dichlorobenzene            | 20    | 20.1   | ug/Kg | 101 | 2   |      | 83  | 122            | 20  |
|               | 1,4-Dichlorobenzene            | 20    | 19.7   | ug/Kg | 99  | 1   |      | 84  | 121            | 20  |
|               | 1,2-Dichlorobenzene            | 20    | 19.8   | ug/Kg | 99  | 2   |      | 83  | 124            | 20  |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2393

**Client:** Union Paving & Construction Company

**Analytical Method:** SW8260D

**Datafile :** VY022788.D

| Lab Sample ID | Parameter                   | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|---------------|-----------------------------|-------|--------|-------|-----|-----|------|-----|----------------|-----|
| VY0624SBSD01  | 1,2-Dibromo-3-Chloropropane | 20    | 19.1   | ug/Kg | 96  | 1   |      | 66  | 134            | 20  |
|               | 1,2,4-Trichlorobenzene      | 20    | 20.0   | ug/Kg | 100 | 4   |      | 78  | 127            | 20  |
|               | 1,2,3-Trichlorobenzene      | 20    | 19.5   | ug/Kg | 98  | 3   |      | 70  | 137            | 20  |