

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture
Project: Amtrak Sawtooth Bridges 2025

SDG No.: Q2400
RunNo.: LB136245

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.495	0.5	99	90-110	06/24/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	06/24/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	06/24/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.495	0.5	99	90-110	06/24/2025