

Laboratory Control Sample Summary

Client: Portal Partners Tri-Venture

SDG No.: Q2400

Project: Amtrak Sawtooth Bridges 2025

Run No.: LB136245

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168573BS							
Hexavalent Chromium	mg/Kg	20	20.0		100	1	84-110	06/24/2025