

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture **SDG No.:** Q2400
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2400 **SAS No.:** Q2400
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	240000	250000	96	216000	294000	06/27/2025	13:04	LB136319
	Antimony	21.2			-50	50	06/27/2025	13:04	LB136319
	Arsenic	1.80			-20	20	06/27/2025	13:04	LB136319
	Barium	5.72	6.0	95	-94	106	06/27/2025	13:04	LB136319
	Beryllium	1.16			-6	6	06/27/2025	13:04	LB136319
	Cadmium	-1.37	1.0	137	-5	7	06/27/2025	13:04	LB136319
	Calcium	237000	240000	99	208000	282000	06/27/2025	13:04	LB136319
	Chromium	49.2	52.0	95	42	62	06/27/2025	13:04	LB136319
	Cobalt	29.0			-30	30	06/27/2025	13:04	LB136319
	Copper	-5.04	2.0	252	-18	22	06/27/2025	13:04	LB136319
	Iron	101000	100000	101	85600	116500	06/27/2025	13:04	LB136319
	Lead	-0.17			-12	12	06/27/2025	13:04	LB136319
	Magnesium	257000	260000	99	216000	294000	06/27/2025	13:04	LB136319
	Manganese	5.41	7.0	77	-13	27	06/27/2025	13:04	LB136319
	Nickel	6.13	2.0	306	-38	42	06/27/2025	13:04	LB136319
	Potassium	10.4			0	0	06/27/2025	13:04	LB136319
	Selenium	4.19			-20	20	06/27/2025	13:04	LB136319
	Silver	4.58			-10	10	06/27/2025	13:04	LB136319
	Sodium	137			0	0	06/27/2025	13:04	LB136319
	Thallium	17.3			-40	40	06/27/2025	13:04	LB136319
	Vanadium	5.71			-40	40	06/27/2025	13:04	LB136319
	Zinc	3.40			-40	40	06/27/2025	13:04	LB136319
ICSAB01	Aluminum	236000	250000	94	209000	285000	06/27/2025	13:09	LB136319
	Antimony	633	620	102	525	711	06/27/2025	13:09	LB136319
	Arsenic	100	100	100	88.4	120	06/27/2025	13:09	LB136319
	Barium	518	540	96	437	637	06/27/2025	13:09	LB136319
	Beryllium	504	500	101	420	570	06/27/2025	13:09	LB136319
	Cadmium	994	970	102	826	1120	06/27/2025	13:09	LB136319
	Calcium	233000	230000	101	199000	271000	06/27/2025	13:09	LB136319
	Chromium	521	540	96	460	624	06/27/2025	13:09	LB136319
	Cobalt	519	480	108	404	548	06/27/2025	13:09	LB136319
	Copper	479	510	94	434	588	06/27/2025	13:09	LB136319
	Iron	99300	99000	100	84400	114500	06/27/2025	13:09	LB136319
	Lead	47.5	49.0	97	37	61	06/27/2025	13:09	LB136319
	Magnesium	252000	250000	101	210000	286000	06/27/2025	13:09	LB136319
	Manganese	502	510	98	430	584	06/27/2025	13:09	LB136319
	Nickel	975	950	103	810	1100	06/27/2025	13:09	LB136319
	Potassium	6.48			0	0	06/27/2025	13:09	LB136319
	Selenium	55.8	46.0	121	26	66	06/27/2025	13:09	LB136319
	Silver	195	200	98	170	232	06/27/2025	13:09	LB136319
	Sodium	135			0	0	06/27/2025	13:09	LB136319
	Thallium	102	110	93	68	148	06/27/2025	13:09	LB136319

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture **SDG No.:** Q2400
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2400 **SAS No.:** Q2400
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	487	490	99	417	565	06/27/2025	13:09	LB136319
	Zinc	1030	950	108	809	1095	06/27/2025	13:09	LB136319