

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2400 SAS No.: Q2400 SDG NO.: Q2400

Continuing Calib Date: 06/25/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 21:27 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1242-1 (1)   | 5.64       | 5.64      | 5.54      | 5.74  | 0.00       |
| Aroclor-1242-2 (2)   | 5.67       | 5.67      | 5.57      | 5.77  | 0.01       |
| Aroclor-1242-3 (3)   | 5.73       | 5.73      | 5.63      | 5.83  | 0.00       |
| Aroclor-1242-4 (4)   | 5.82       | 5.83      | 5.73      | 5.93  | 0.01       |
| Aroclor-1242-5 (5)   | 6.55       | 6.56      | 6.46      | 6.66  | 0.01       |
| Tetrachloro-m-xylene | 4.49       | 4.49      | 4.39      | 4.59  | 0.00       |
| Decachlorobiphenyl   | 10.18      | 10.19     | 10.09     | 10.29 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2400 SAS No.: Q2400 SDG NO.: Q2400

Continuing Calib Date: 06/25/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 21:27 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1242-1       | (1) | 4.87       | 4.87      | 4.77      | 4.97 | 0.01       |
| Aroclor-1242-2       | (2) | 4.88       | 4.89      | 4.79      | 4.99 | 0.01       |
| Aroclor-1242-3       | (3) | 5.06       | 5.06      | 4.96      | 5.16 | 0.00       |
| Aroclor-1242-4       | (4) | 5.14       | 5.15      | 5.05      | 5.25 | 0.01       |
| Aroclor-1242-5       | (5) | 5.67       | 5.67      | 5.57      | 5.77 | 0.01       |
| Tetrachloro-m-xylene |     | 3.79       | 3.79      | 3.69      | 3.89 | 0.00       |
| Decachlorobiphenyl   |     | 8.79       | 8.80      | 8.70      | 8.90 | 0.01       |