

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2400 SAS No.: Q2400 SDG NO.: Q2400

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PO111837.D Time Analyzed: 10:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.764	4.666	4.866	478.580	500.000	-4.3
Aroclor-1016-2	4.783	4.684	4.884	502.080	500.000	0.4
Aroclor-1016-3	4.840	4.741	4.941	493.430	500.000	-1.3
Aroclor-1016-4	4.959	4.861	5.061	498.230	500.000	-0.4
Aroclor-1016-5	5.217	5.118	5.318	505.260	500.000	1.1
Aroclor-1260-1	6.255	6.157	6.357	501.740	500.000	0.3
Aroclor-1260-2	6.444	6.346	6.546	535.120	500.000	7.0
Aroclor-1260-3	6.811	6.713	6.913	557.390	500.000	11.5
Aroclor-1260-4	7.071	6.972	7.172	549.840	500.000	10.0
Aroclor-1260-5	7.313	7.216	7.416	549.690	500.000	9.9
Decachlorobiphenyl	8.709	8.612	8.812	48.930	50.000	-2.1
Tetrachloro-m-xylene	3.676	3.577	3.777	52.330	50.000	4.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2400 SAS No.: Q2400 SDG NO.: Q2400

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PO111837.D Time Analyzed: 10:12

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.747	4.651	4.851	495.700	500.000	-0.9
Aroclor-1016-2	4.766	4.669	4.869	491.990	500.000	-1.6
Aroclor-1016-3	4.940	4.844	5.044	491.700	500.000	-1.7
Aroclor-1016-4	4.983	4.887	5.087	490.880	500.000	-1.8
Aroclor-1016-5	5.195	5.099	5.299	488.140	500.000	-2.4
Aroclor-1260-1	6.225	6.129	6.329	489.490	500.000	-2.1
Aroclor-1260-2	6.413	6.316	6.516	501.980	500.000	0.4
Aroclor-1260-3	6.565	6.469	6.669	485.440	500.000	-2.9
Aroclor-1260-4	7.035	6.939	7.139	480.340	500.000	-3.9
Aroclor-1260-5	7.276	7.181	7.381	482.370	500.000	-3.5
Decachlorobiphenyl	8.657	8.561	8.761	47.900	50.000	-4.2
Tetrachloro-m-xylene	3.670	3.573	3.773	49.130	50.000	-1.7