

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2400 SAS No.: Q2400 SDG NO.: Q2400

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL02 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PO111852.D Time Analyzed: 17:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.764	4.666	4.866	466.320	500.000	-6.7
Aroclor-1016-2	4.783	4.684	4.884	490.250	500.000	-2.0
Aroclor-1016-3	4.840	4.741	4.941	477.440	500.000	-4.5
Aroclor-1016-4	4.959	4.861	5.061	476.840	500.000	-4.6
Aroclor-1016-5	5.217	5.118	5.318	506.280	500.000	1.3
Aroclor-1260-1	6.255	6.157	6.357	472.500	500.000	-5.5
Aroclor-1260-2	6.444	6.346	6.546	506.380	500.000	1.3
Aroclor-1260-3	6.811	6.713	6.913	539.280	500.000	7.9
Aroclor-1260-4	7.070	6.972	7.172	558.080	500.000	11.6
Aroclor-1260-5	7.314	7.216	7.416	535.440	500.000	7.1
Decachlorobiphenyl	8.709	8.612	8.812	44.490	50.000	-11.0
Tetrachloro-m-xylene	3.676	3.577	3.777	51.140	50.000	2.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2400 SAS No.: Q2400 SDG NO.: Q2400

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL02 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PO111852.D Time Analyzed: 17:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.748	4.651	4.851	488.370	500.000	-2.3
Aroclor-1016-2	4.767	4.669	4.869	489.250	500.000	-2.2
Aroclor-1016-3	4.942	4.844	5.044	487.830	500.000	-2.4
Aroclor-1016-4	4.984	4.887	5.087	485.500	500.000	-2.9
Aroclor-1016-5	5.197	5.099	5.299	487.430	500.000	-2.5
Aroclor-1260-1	6.226	6.129	6.329	480.520	500.000	-3.9
Aroclor-1260-2	6.415	6.316	6.516	495.430	500.000	-0.9
Aroclor-1260-3	6.565	6.469	6.669	468.960	500.000	-6.2
Aroclor-1260-4	7.036	6.939	7.139	463.180	500.000	-7.4
Aroclor-1260-5	7.278	7.181	7.381	468.310	500.000	-6.3
Decachlorobiphenyl	8.657	8.561	8.761	45.710	50.000	-8.6
Tetrachloro-m-xylene	3.672	3.573	3.773	48.490	50.000	-3.0