

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2400 SAS No.: Q2400 SDG NO.: Q2400

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP073230.D Time Analyzed: 09:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.642	5.546	5.746	497.850	500.000	-0.4
Aroclor-1016-2	5.663	5.567	5.767	531.200	500.000	6.2
Aroclor-1016-3	5.726	5.629	5.829	527.920	500.000	5.6
Aroclor-1016-4	5.823	5.727	5.927	553.300	500.000	10.7
Aroclor-1016-5	6.115	6.019	6.219	524.910	500.000	5.0
Aroclor-1260-1	7.231	7.136	7.336	558.340	500.000	11.7
Aroclor-1260-2	7.485	7.389	7.589	543.860	500.000	8.8
Aroclor-1260-3	7.843	7.747	7.947	553.660	500.000	10.7
Aroclor-1260-4	8.067	7.971	8.171	560.880	500.000	12.2
Aroclor-1260-5	8.385	8.288	8.488	533.220	500.000	6.6
Decachlorobiphenyl	10.182	10.087	10.287	51.670	50.000	3.3
Tetrachloro-m-xylene	4.491	4.394	4.594	51.870	50.000	3.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2400 SAS No.: Q2400 SDG NO.: Q2400

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP073230.D Time Analyzed: 09:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.865	4.768	4.968	478.630	500.000	-4.3
Aroclor-1016-2	4.882	4.786	4.986	499.110	500.000	-0.2
Aroclor-1016-3	5.059	4.962	5.162	500.330	500.000	0.1
Aroclor-1016-4	5.101	5.004	5.204	498.360	500.000	-0.3
Aroclor-1016-5	5.314	5.218	5.418	534.890	500.000	7.0
Aroclor-1260-1	6.345	6.249	6.449	506.310	500.000	1.3
Aroclor-1260-2	6.534	6.438	6.638	518.190	500.000	3.6
Aroclor-1260-3	6.685	6.589	6.789	499.370	500.000	-0.1
Aroclor-1260-4	7.155	7.060	7.260	493.440	500.000	-1.3
Aroclor-1260-5	7.397	7.302	7.502	497.480	500.000	-0.5
Decachlorobiphenyl	8.792	8.697	8.897	53.360	50.000	6.7
Tetrachloro-m-xylene	3.785	3.688	3.888	49.360	50.000	-1.3