

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2400 SAS No.: Q2400 SDG NO.: Q2400

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP073245.D Time Analyzed: 15:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.644	5.546	5.746	491.050	500.000	-1.8
Aroclor-1016-2	5.666	5.567	5.767	517.830	500.000	3.6
Aroclor-1016-3	5.728	5.629	5.829	515.440	500.000	3.1
Aroclor-1016-4	5.826	5.727	5.927	548.500	500.000	9.7
Aroclor-1016-5	6.117	6.019	6.219	510.890	500.000	2.2
Aroclor-1260-1	7.233	7.136	7.336	549.470	500.000	9.9
Aroclor-1260-2	7.487	7.389	7.589	535.750	500.000	7.2
Aroclor-1260-3	7.845	7.747	7.947	541.180	500.000	8.2
Aroclor-1260-4	8.069	7.971	8.171	540.740	500.000	8.1
Aroclor-1260-5	8.386	8.288	8.488	517.600	500.000	3.5
Decachlorobiphenyl	10.183	10.087	10.287	50.590	50.000	1.2
Tetrachloro-m-xylene	4.493	4.394	4.594	50.850	50.000	1.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2400 SAS No.: Q2400 SDG NO.: Q2400

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP073245.D Time Analyzed: 15:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.866	4.768	4.968	493.090	500.000	-1.4
Aroclor-1016-2	4.883	4.786	4.986	489.210	500.000	-2.2
Aroclor-1016-3	5.060	4.962	5.162	496.280	500.000	-0.7
Aroclor-1016-4	5.102	5.004	5.204	491.290	500.000	-1.7
Aroclor-1016-5	5.315	5.218	5.418	539.600	500.000	7.9
Aroclor-1260-1	6.346	6.249	6.449	499.230	500.000	-0.2
Aroclor-1260-2	6.535	6.438	6.638	517.590	500.000	3.5
Aroclor-1260-3	6.686	6.589	6.789	495.890	500.000	-0.8
Aroclor-1260-4	7.156	7.060	7.260	493.520	500.000	-1.3
Aroclor-1260-5	7.398	7.302	7.502	497.660	500.000	-0.5
Decachlorobiphenyl	8.792	8.697	8.897	52.400	50.000	4.8
Tetrachloro-m-xylene	3.786	3.688	3.888	49.630	50.000	-0.7