

Surrogate Summary

SDG No.: **Q2400**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO111586.D	PIBLK-PO111586.D	Tetrachloro-m-xylene	1	20	20.3	102		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.7	104		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.0	95		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.5	103		70 (60)	130 (140)
I.BLK-PO111841.D	PIBLK-PO111841.D	Tetrachloro-m-xylene	1	20	20.0	100		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.3	101		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	18.2	91		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	19.6	98		70 (60)	130 (140)
Q2400-01	B-156-SB01	Tetrachloro-m-xylene	1	20	12.6	63		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	8.33	42		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	10.8	54		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	9.21	46		30 (32)	150 (175)
Q2400-02	B-134-SB01	Tetrachloro-m-xylene	1	20	20.8	104		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	15.0	75		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.6	93		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	15.8	79		30 (32)	150 (175)
I.BLK-PO111856.D	PIBLK-PO111856.D	Tetrachloro-m-xylene	1	20	19.8	99		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.7	94		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	18.0	90		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	19.0	95		70 (60)	130 (140)
I.BLK-PP072990.D	PIBLK-PP072990.D	Tetrachloro-m-xylene	1	20	17.3	86		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.3	91		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	18.0	90		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.2	86		70 (60)	130 (140)
I.BLK-PP073234.D	PIBLK-PP073234.D	Tetrachloro-m-xylene	1	20	19.6	98		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	19.6	98		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.2	96		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	21.1	105		70 (60)	130 (140)
PB168607BL	PB168607BL	Tetrachloro-m-xylene	1	20	19.8	99		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.6	98		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.5	103		30 (32)	150 (175)
PB168607BS	PB168607BS	Tetrachloro-m-xylene	1	20	20.2	101		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.4	102		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.1	96		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.6	108		30 (32)	150 (175)
Q2392-01MS	S-1MS	Tetrachloro-m-xylene	1	20	20.8	104		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.4	102		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.0	95		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.0	110		30 (32)	150 (175)
Q2392-01MSD	S-1MSD	Tetrachloro-m-xylene	1	20	19.2	96		30 (32)	150 (144)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q2400

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2392-01MSD	S-1MSD	Decachlorobiphenyl	1	20	18.4	92		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	17.3	86		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.6	98		30 (32)	150 (175)
I.BLK-PP073249.D	PIBLK-PP073249.D	Tetrachloro-m-xylene	1	20	18.5	93		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.9	95		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	18.5	93		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.2	101		70 (60)	130 (140)