

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2400</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Portal Partners Tri-Venture</u>	Datafile :	<u>PP073237.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168607BS (Column 1)	AR1016	166.6	165	ug/kg	99				40 (71)	140 (120)	
	AR1260	166.6	165	ug/kg	99				40 (65)	140 (130)	
PB168607BS (Column 2)	AR1016	166.6	155	ug/kg	93				40 (71)	140 (120)	
	AR1260	166.6	152	ug/kg	91				40 (65)	140 (130)	