

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2401 SAS No.: Q2401 SDG NO.: Q2401
 Lab File ID: VN087207.D Date Analyzed: 06/27/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	196333	8.23	346797	9.11	266770	11.87
UPPER LIMIT	392666	8.73	693594	9.606	533540	12.365
LOWER LIMIT	98166.5	7.73	173399	8.606	133385	11.365
EPA SAMPLE NO.						
MW2	141577	8.23	317471	9.11	253799	11.87
VN0627WBL01	196165	8.23	398772	9.10	333801	11.87
VN0627WBS01	176379	8.23	277531	9.11	254005	11.87
VN0627WBSD01	143301	8.23	249904	9.11	238840	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.