

Surrogate Summary

SDG No.: Q2402

Client: G Environmental

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2402-01	OBSI	1,2-Dichloroethane-d4	50	61.1	122	70 (74)	130 (125)
		Dibromofluoromethane	50	50.2	100	70 (75)	130 (124)
		Toluene-d8	50	52.8	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.4	99	70 (77)	130 (121)
VN0624WBL01	VN0624WBL01	1,2-Dichloroethane-d4	50	56.3	113	70 (74)	130 (125)
		Dibromofluoromethane	50	53.1	106	70 (75)	130 (124)
		Toluene-d8	50	53.0	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.4	105	70 (77)	130 (121)
VN0624WBS01	VN0624WBS01	1,2-Dichloroethane-d4	50	52.5	105	70 (74)	130 (125)
		Dibromofluoromethane	50	51.6	103	70 (75)	130 (124)
		Toluene-d8	50	47.1	94	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.6	105	70 (77)	130 (121)
VN0624WBSD0	VN0624WBSD01	1,2-Dichloroethane-d4	50	54.1	108	70 (74)	130 (125)
		Dibromofluoromethane	50	54.7	109	70 (75)	130 (124)
		Toluene-d8	50	47.6	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.8	102	70 (77)	130 (121)

