

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2402 SAS No.: Q2402 SDG NO.: Q2402
 Lab File ID: VN087149.D Date Analyzed: 06/24/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:14
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	137487	8.23	272967	9.11	232699	11.87
UPPER LIMIT	274974	8.73	545934	9.606	465398	12.365
LOWER LIMIT	68743.5	7.73	136484	8.606	116350	11.365
EPA SAMPLE NO.						
OBSI	181938	8.23	387109	9.11	336293	11.87
VN0624WBL01	149121	8.23	287883	9.11	275522	11.87
VN0624WBS01	160941	8.23	295023	9.11	279201	11.87
VN0624WBSD01	148340	8.23	264363	9.11	239185	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.