

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2402 SAS No.: Q2402 SDG NO.: Q2402
 Lab File ID: VN087149.D Date Analyzed: 06/24/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:14
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	123154	13.788				
UPPER LIMIT	246308	14.288				
LOWER LIMIT	61577	13.288				
EPA SAMPLE NO.						
OBSI	160118	13.79				
VN0624WBL01	132163	13.79				
VN0624WBS01	141896	13.79				
VN0624WBSD01	121327	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.