



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Total Quality Environmental Inc.

Project Name: Flushing Ave

Project # N/A

Order ID # Q2404

Test Name: Metals Group4

A. Number of Samples and Date of Receipt:

12 Water samples were received on 06/23/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Metals Group4 and Metals ICP-Group1. This data package contains results for Metals Group4.

C. Analytical Techniques:

The analysis and digestion of Metals Group4 was based on method 200.7.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all parameters.

The Duplicate analysis met criteria for all parameters.

The Matrix Spike analysis met criteria for all parameters.

The Matrix Spike Duplicate analysis met criteria for all parameters.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

The temperature of the samples at the time of receipt was 26.4°C.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____