

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2404
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2404 **SAS No.:** Q2404
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3870	4000	97	95 - 105	P	06/26/2025	17:07	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2404
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2404 **SAS No.:** Q2404
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	13.5	12.0	112	80 - 120	P	06/26/2025	17:19	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2404
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2404 **SAS No.:** Q2404
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4910	5000	98	90 - 110	P	06/26/2025	17:55	LB136305
CCV02	Lead	4900	5000	98	90 - 110	P	06/26/2025	18:48	LB136305
CCV03	Lead	4840	5000	97	90 - 110	P	06/26/2025	19:27	LB136305
CCV04	Lead	4900	5000	98	90 - 110	P	06/26/2025	20:02	LB136305
CCV05	Lead	4950	5000	99	90 - 110	P	06/26/2025	20:46	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2404
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2404 **SAS No.:** Q2404
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3950	4000	99	95 - 105	P	06/27/2025	11:45	LB136319

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2404
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2404 **SAS No.:** Q2404
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	10.3	12.0	86	80 - 120	P	06/27/2025	12:03	LB136319

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2404
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2404 **SAS No.:** Q2404
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4980	5000	100	90 - 110	P	06/27/2025	13:22	LB136319
CCV02	Lead	5070	5000	101	90 - 110	P	06/27/2025	14:19	LB136319
CCV03	Lead	5120	5000	102	90 - 110	P	06/27/2025	15:12	LB136319
CCV04	Lead	5120	5000	102	90 - 110	P	06/27/2025	16:04	LB136319
CCV05	Lead	4960	5000	99	90 - 110	P	06/27/2025	16:56	LB136319
CCV06	Lead	5000	5000	100	90 - 110	P	06/27/2025	17:49	LB136319
CCV07	Lead	5120	5000	102	90 - 110	P	06/27/2025	18:41	LB136319
CCV08	Lead	5040	5000	101	90 - 110	P	06/27/2025	19:29	LB136319

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2404
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2404 **SAS No.:** Q2404
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3960	4000	99	95 - 105	P	07/03/2025	14:50	LB136377

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2404
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2404 **SAS No.:** Q2404
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	10.9	12.0	91	80 - 120	P	07/03/2025	14:55	LB136377

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q2404

Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2404 **SAS No.:** Q2404

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4690	5000	94	90 - 110	P	07/03/2025	16:06	LB136377
CCV02	Lead	4740	5000	95	90 - 110	P	07/03/2025	16:53	LB136377
CCV03	Lead	4720	5000	94	90 - 110	P	07/03/2025	17:28	LB136377