

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Total Quality Environmental Inc. **SDG No.:** Q2404
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q2404 **SAS No.:** Q2404
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	2.20			-12	12	06/26/2025	17:37	LB136305
ICSAB01	Lead	49.4	49.0	101	37	61	06/26/2025	17:42	LB136305
ICSA	Lead	35.9			-12	12	06/26/2025	17:46	LB136305
ICSAB	Lead	70.9	49.0	145	37	61	06/26/2025	17:50	LB136305
ICSA01	Lead	-0.17			-12	12	06/27/2025	13:04	LB136319
ICSAB01	Lead	47.5	49.0	97	37	61	06/27/2025	13:09	LB136319
ICSA	Lead	11.8			-12	12	06/27/2025	13:14	LB136319
ICSAB	Lead	61.7	49.0	126	37	61	06/27/2025	13:18	LB136319
ICSA01	Lead	1.07			-12	12	07/03/2025	15:41	LB136377
ICSAB01	Lead	44.5	49.0	91	37	61	07/03/2025	15:45	LB136377
ICSA	Lead	53.6			-12	12	07/03/2025	15:50	LB136377
ICSAB	Lead	87.4	49.0	178	37	61	07/03/2025	15:54	LB136377