

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Coppola Services</u>	SDG No.: <u>Q2409</u>
Contract: <u>COPP02</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	245000	255000	96	216000	294000	10/02/2025	11:42	LB137406
	Antimony	1.99			-50	50	10/02/2025	11:42	LB137406
	Arsenic	5.85			-20	20	10/02/2025	11:42	LB137406
	Barium	18.7	6.0	312	-94	106	10/02/2025	11:42	LB137406
	Beryllium	1.98			-6	6	10/02/2025	11:42	LB137406
	Cadmium	-1.23	1.0	123	-5	7	10/02/2025	11:42	LB137406
	Calcium	236000	245000	96	208000	282000	10/02/2025	11:42	LB137406
	Chromium	50.8	52.0	98	42	62	10/02/2025	11:42	LB137406
	Cobalt	0.27			-30	30	10/02/2025	11:42	LB137406
	Copper	3.53	2.0	176	-18	22	10/02/2025	11:42	LB137406
	Iron	96000	101000	95	85600	116500	10/02/2025	11:42	LB137406
	Lead	-7.87			-12	12	10/02/2025	11:42	LB137406
	Magnesium	251000	255000	98	216000	294000	10/02/2025	11:42	LB137406
	Manganese	0.23	7.0	3	-13	27	10/02/2025	11:42	LB137406
	Nickel	3.41	2.0	170	-38	42	10/02/2025	11:42	LB137406
	Potassium	-102			0	0	10/02/2025	11:42	LB137406
	Selenium	-8.13			-20	20	10/02/2025	11:42	LB137406
	Silver	-3.77			-10	10	10/02/2025	11:42	LB137406
	Sodium	-181			0	0	10/02/2025	11:42	LB137406
	Thallium	14.0			-40	40	10/02/2025	11:42	LB137406
	Vanadium	2.54			-40	40	10/02/2025	11:42	LB137406
	Zinc	-0.077			-40	40	10/02/2025	11:42	LB137406
ICSAB01	Aluminum	249000	247000	101	209000	285000	10/02/2025	11:46	LB137406
	Antimony	622	618	101	525	711	10/02/2025	11:46	LB137406
	Arsenic	110	104	106	88.4	120	10/02/2025	11:46	LB137406
	Barium	515	537	96	437	637	10/02/2025	11:46	LB137406
	Beryllium	548	495	111	420	570	10/02/2025	11:46	LB137406
	Cadmium	1040	972	107	826	1120	10/02/2025	11:46	LB137406
	Calcium	240000	235000	102	199000	271000	10/02/2025	11:46	LB137406
	Chromium	594	542	110	460	624	10/02/2025	11:46	LB137406
	Cobalt	515	476	108	404	548	10/02/2025	11:46	LB137406
	Copper	510	511	100	434	588	10/02/2025	11:46	LB137406
	Iron	98300	99300	99	84400	114500	10/02/2025	11:46	LB137406
	Lead	42.6	49.0	87	37	61	10/02/2025	11:46	LB137406
	Magnesium	255000	248000	103	210000	286000	10/02/2025	11:46	LB137406
	Manganese	497	507	98	430	584	10/02/2025	11:46	LB137406
	Nickel	1030	954	108	810	1100	10/02/2025	11:46	LB137406
	Potassium	-265			0	0	10/02/2025	11:46	LB137406
	Selenium	46.7	46.0	102	26	66	10/02/2025	11:46	LB137406
	Silver	212	201	106	170	232	10/02/2025	11:46	LB137406
	Sodium	-229			0	0	10/02/2025	11:46	LB137406
	Thallium	115	108	106	68	148	10/02/2025	11:46	LB137406

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Coppola Services

SDG No.: Q2409

Contract: COPP02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	499	491	102	417	565	10/02/2025	11:46	LB137406
	Zinc	1060	952	111	809	1095	10/02/2025	11:46	LB137406
ICSA01	Aluminum	245000	255000	96	216000	294000	10/03/2025	13:20	LB137426
	Antimony	3.52			-50	50	10/03/2025	13:20	LB137426
	Arsenic	8.41			-20	20	10/03/2025	13:20	LB137426
	Barium	1.66	6.0	28	-94	106	10/03/2025	13:20	LB137426
	Beryllium	1.42			-6	6	10/03/2025	13:20	LB137426
	Cadmium	0.91	1.0	91	-5	7	10/03/2025	13:20	LB137426
	Calcium	236000	245000	96	208000	282000	10/03/2025	13:20	LB137426
	Chromium	46.6	52.0	90	42	62	10/03/2025	13:20	LB137426
	Cobalt	0.68			-30	30	10/03/2025	13:20	LB137426
	Copper	3.06	2.0	153	-18	22	10/03/2025	13:20	LB137426
	Iron	99600	101000	99	85600	116500	10/03/2025	13:20	LB137426
	Lead	0.97			-12	12	10/03/2025	13:20	LB137426
	Magnesium	247000	255000	97	216000	294000	10/03/2025	13:20	LB137426
	Manganese	-1.76	7.0	25	-13	27	10/03/2025	13:20	LB137426
	Nickel	3.93	2.0	196	-38	42	10/03/2025	13:20	LB137426
	Potassium	187			0	0	10/03/2025	13:20	LB137426
	Selenium	9.27			-20	20	10/03/2025	13:20	LB137426
	Silver	1.41			-10	10	10/03/2025	13:20	LB137426
	Sodium	127			0	0	10/03/2025	13:20	LB137426
	Thallium	11.3			-40	40	10/03/2025	13:20	LB137426
	Vanadium	4.84			-40	40	10/03/2025	13:20	LB137426
	Zinc	1.66			-40	40	10/03/2025	13:20	LB137426
ICSAB01	Aluminum	247000	247000	100	209000	285000	10/03/2025	13:25	LB137426
	Antimony	609	618	98	525	711	10/03/2025	13:25	LB137426
	Arsenic	110	104	106	88.4	120	10/03/2025	13:25	LB137426
	Barium	496	537	92	437	637	10/03/2025	13:25	LB137426
	Beryllium	516	495	104	420	570	10/03/2025	13:25	LB137426
	Cadmium	980	972	101	826	1120	10/03/2025	13:25	LB137426
	Calcium	235000	235000	100	199000	271000	10/03/2025	13:25	LB137426
	Chromium	553	542	102	460	624	10/03/2025	13:25	LB137426
	Cobalt	491	476	103	404	548	10/03/2025	13:25	LB137426
	Copper	484	511	95	434	588	10/03/2025	13:25	LB137426
	Iron	97500	99300	98	84400	114500	10/03/2025	13:25	LB137426
	Lead	49.1	49.0	100	37	61	10/03/2025	13:25	LB137426
	Magnesium	248000	248000	100	210000	286000	10/03/2025	13:25	LB137426
	Manganese	492	507	97	430	584	10/03/2025	13:25	LB137426
	Nickel	981	954	103	810	1100	10/03/2025	13:25	LB137426
	Potassium	15.2			0	0	10/03/2025	13:25	LB137426
	Selenium	56.4	46.0	123	26	66	10/03/2025	13:25	LB137426
	Silver	208	201	104	170	232	10/03/2025	13:25	LB137426

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Coppola Services

SDG No.: Q2409

Contract: COPP02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Sodium	92.2			0	0	10/03/2025	13:25	LB137426
	Thallium	105	108	97	68	148	10/03/2025	13:25	LB137426
	Vanadium	485	491	99	417	565	10/03/2025	13:25	LB137426
	Zinc	1030	952	108	809	1095	10/03/2025	13:25	LB137426