

LB136295


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

33630



A= 0.0000e+000
 B= 2.8328e-004
 C= -1.8782e-002
 Rho= 0.9953127
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	-0.015	-0.015	12	0.000	12	0				-
0.2	0.200	0.160	-0.040	630	0.0 %	630					-20
2.5	2.500	2.605	0.105	9261	0.0 %	9261					4
5.0	5.000	4.768	-0.232	16898	0.0 %	16898					-5
7.5	7.500	8.175	0.675	28924	0.0 %	28924					9
10.0	10.000	9.508	-0.492	33630	0.0 %	33630					-3

LB136295 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jun 2025 11:58:16

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	12	-	0.0000	7470A	PPB	26 Jun 2025 12:08:59	S	Std
0.2 - 1	50.2	630	-	0.2000	7470A	PPB	26 Jun 2025 12:11:16	S	Std
2.5 - 1	52.5	9261	-	2.5000	7470A	PPB	26 Jun 2025 12:13:33	S	Std
5.0 - 1	55	16898	-	5.0000	7470A	PPB	26 Jun 2025 12:15:51	S	Std
7.5 - 1	57.5	28924	-	7.5000	7470A	PPB	26 Jun 2025 12:18:10	S	Std
10.0 - 1	510	33630	-	10.0000	7470A	PPB	26 Jun 2025 12:23:10	S	Std
ICV07 - 1	ICV07	13308	3.7511	-	7470A	PPB	26 Jun 2025 12:28:24	U	SMPL
ICB07 - 1	ICB07	-235	-0.0854	-	7470A	PPB	26 Jun 2025 12:30:39	U	SMPL
CCV20 - 1	CCV20	16683	4.7072	-	7470A	PPB	26 Jun 2025 12:32:56	U	SMPL
CCB20 - 1	CCB20	-242	-0.0873	-	7470A	PPB	26 Jun 2025 12:35:11	U	SMPL
CRA - 1	CRA	739	0.1906	-	7470A	PPB	26 Jun 2025 12:37:28	U	SMPL
HighStd - 1	HighStd	33974	9.6055	-	7470A	PPB	26 Jun 2025 12:39:43	U	SMPL
ChkStd - 1	ChkStd	23510	6.6412	-	7470A	PPB	26 Jun 2025 12:41:58	U	SMPL
PB168615BL - 1	PBW	-54	-0.0341	-	7470A	PPB	26 Jun 2025 12:47:24	U	SMPL
PB168615BS - 1	LCSW	13495	3.8041	-	7470A	PPB	26 Jun 2025 12:52:24	U	SMPL
Q2386-02 - 1	SU-1-062025	-174	-0.0681	-	7470A	PPB	26 Jun 2025 12:54:40	U	SMPL
Q2388-04 - 1	TP-12	-21	-0.0247	-	7470A	PPB	26 Jun 2025 12:56:58	U	SMPL
Q2389-04 - 1	MH-J-I	-27	-0.0264	-	7470A	PPB	26 Jun 2025 12:59:14	U	SMPL
Q2391-01 - 1	AUD-1623	31	-0.0100	-	7470A	PPB	26 Jun 2025 13:01:32	U	SMPL
Q2394-04 - 1	MH-K/L	13	-0.0151	-	7470A	PPB	26 Jun 2025 13:03:47	U	SMPL
CCV21 - 1	CCV21	17990	5.0775	-	7470A	PPB	26 Jun 2025 13:06:02	U	SMPL
CCB21 - 1	CCB21	-215	-0.0797	-	7470A	PPB	26 Jun 2025 13:08:17	U	SMPL
Q2399-04 - 1	TP-13	-25	-0.0259	-	7470A	PPB	26 Jun 2025 13:10:35	U	SMPL
Q2399-08 - 1	EP-13	-1	-0.0191	-	7470A	PPB	26 Jun 2025 13:12:50	U	SMPL
Q2405-04 - 1	MH-M/N	-6	-0.0205	-	7470A	PPB	26 Jun 2025 13:15:06	U	SMPL
Q2409-02 - 1	COP-SOIL-PILE	-24	-0.0256	-	7470A	PPB	26 Jun 2025 13:17:21	U	SMPL
Q2409-02DUP - 1	COP-SOIL-PILEDUP	19	-0.0134	-	7470A	PPB	26 Jun 2025 13:19:37	U	SMPL
Q2409-02MS - 1	COP-SOIL-PILEMS	13439	3.7882	-	7470A	PPB	26 Jun 2025 13:21:53	U	SMPL
Q2409-02MSD - 1	COP-SOIL-PILEMSD	13433	3.7865	-	7470A	PPB	26 Jun 2025 13:24:09	U	SMPL
PB168571TB - 1	PB168571TB	-185	-0.0712	-	7470A	PPB	26 Jun 2025 13:26:28	U	SMPL
Q2409-02LX5 - 1		-11	-0.0219	-	7470A	PPB	26 Jun 2025 13:28:47	U	SMPL
Q2409-02A - 1		13119	3.6976	-	7470A	PPB	26 Jun 2025 13:31:03	U	SMPL
CCV22 - 1	CCV22	16972	4.7891	-	7470A	PPB	26 Jun 2025 13:33:18	U	SMPL
CCB22 - 1	CCB22	-215	-0.0797	-	7470A	PPB	26 Jun 2025 13:35:36	U	SMPL