



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Coppola Services SDG No.: Q2409
Contract: COPP02 Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	16.5	20.0	82	65 - 135	P	06/25/2025	15:24	LB136273
	Barium	99.4	100	99	65 - 135	P	06/25/2025	15:24	LB136273
	Cadmium	5.59	6.0	93	65 - 135	P	06/25/2025	15:24	LB136273
	Chromium	8.98	10.0	90	65 - 135	P	06/25/2025	15:24	LB136273
	Copper	22.1	20.0	110	65 - 135	P	06/25/2025	15:24	LB136273
	Lead	13.2	12.0	110	65 - 135	P	06/25/2025	15:24	LB136273
	Nickel	38.6	40.0	96	65 - 135	P	06/25/2025	15:24	LB136273
	Selenium	20.3	20.0	102	65 - 135	P	06/25/2025	15:24	LB136273
	Silver	9.71	10.0	97	65 - 135	P	06/25/2025	15:24	LB136273
	Zinc	39.7	40.0	99	65 - 135	P	06/25/2025	15:24	LB136273
CRA	Mercury	0.19	0.2	95	70 - 130	CV	06/26/2025	12:37	LB136295
CRI01	Arsenic	18.2	20.0	91	65 - 135	P	06/26/2025	17:28	LB136305
	Barium	100	100	100	65 - 135	P	06/26/2025	17:28	LB136305
	Cadmium	5.73	6.0	96	65 - 135	P	06/26/2025	17:28	LB136305
	Chromium	11.3	10.0	113	65 - 135	P	06/26/2025	17:28	LB136305
	Copper	22.8	20.0	114	65 - 135	P	06/26/2025	17:28	LB136305
	Lead	11.9	12.0	99	65 - 135	P	06/26/2025	17:28	LB136305
	Nickel	39.3	40.0	98	65 - 135	P	06/26/2025	17:28	LB136305
	Selenium	20.5	20.0	102	65 - 135	P	06/26/2025	17:28	LB136305
	Silver	8.68	10.0	87	65 - 135	P	06/26/2025	17:28	LB136305
	Zinc	40.9	40.0	102	65 - 135	P	06/26/2025	17:28	LB136305