

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Coppola Services **Level:** LOW **SDG No.:** Q2409
Contract: COPP02 **Lab Code:** CHEM **Case No.:** Q2409 **SAS No.:** Q2409
Matrix: Water **Sample ID:** Q2409-02 **Client ID:** COP-SOIL-PILEDUP
Percent Solids for Sample: NA **Duplicate ID** Q2409-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	765		737		4		P
Cadmium	ug/L	20	65.9		64.0		3		P
Chromium	ug/L	20	11.8	J	70.9		143		P
Copper	ug/L	20	85.3	J	78.4	J	8		P
Lead	ug/L	20	155		161		4		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Nickel	ug/L	20	200	U	48.6	J	200.0		P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P
Zinc	ug/L	20	580		557		4		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Coppola Services **Level:** LOW **SDG No.:** Q2409
Contract: COPP02 **Lab Code:** CHEM **Case No.:** Q2409 **SAS No.:** Q2409
Matrix: Water **Sample ID:** Q2409-02MS **Client ID:** COP-SOIL-PILEMSD
Percent Solids for Sample: NA **Duplicate ID** Q2409-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4180		4250		2		P
Barium	ug/L	20	1740		1730		1		P
Cadmium	ug/L	20	1100		1110		1		P
Chromium	ug/L	20	2110		2150		2		P
Copper	ug/L	20	1530		1540		1		P
Lead	ug/L	20	5040		5110		1		P
Mercury	ug/L	20	37.9		37.9		0		CV
Nickel	ug/L	20	2550		2600		2		P
Selenium	ug/L	20	9720		9860		1		P
Silver	ug/L	20	351		357		2		P
Zinc	ug/L	20	1600		1620		1		P