

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** Coppola Services **SDG No.:** Q2409  
**Contract:** COPP02 **Lab Code:** CHEM **Case No.:** Q2409 **SAS No.:** Q2409

| Analyte    | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M |
|------------|-------|------------|--------|---|---------------|----------------------|---|
| PB168613BS |       |            |        |   |               |                      |   |
| Arsenic    | ug/L  | 4000       | 3490   |   | 87            | 80 - 120             | P |
| Barium     | ug/L  | 1000       | 980    |   | 98            | 80 - 120             | P |
| Cadmium    | ug/L  | 1000       | 901    |   | 90            | 80 - 120             | P |
| Chromium   | ug/L  | 2000       | 1950   |   | 98            | 80 - 120             | P |
| Copper     | ug/L  | 1500       | 1530   |   | 102           | 80 - 120             | P |
| Lead       | ug/L  | 5000       | 4490   |   | 90            | 80 - 120             | P |
| Nickel     | ug/L  | 2500       | 2300   |   | 92            | 80 - 120             | P |
| Selenium   | ug/L  | 10000      | 9320   |   | 93            | 80 - 120             | P |
| Silver     | ug/L  | 380        | 341    |   | 90            | 80 - 120             | P |
| Zinc       | ug/L  | 1000       | 1020   |   | 102           | 80 - 120             | P |

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** Coppola Services

**SDG No.:** Q2409

**Contract:** COPP02

**Lab Code:** CHEM

**Case No.:** Q2409

**SAS No.:** Q2409

| Analyte               | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M  |
|-----------------------|-------|------------|--------|---|---------------|----------------------|----|
| PB168615BS<br>Mercury | ug/L  | 4.0        | 3.80   |   | 95            | 80 - 120             | CV |