

CALIBRATION VERIFICATION SUMMARY

Contract: COPP02

Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409 SDG NO.: Q2409

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 19:30 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.64	5.65	5.55	5.75	0.01
Aroclor-1016-2	(2)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-3	(3)	5.72	5.73	5.63	5.83	0.01
Aroclor-1016-4	(4)	5.82	5.83	5.73	5.93	0.01
Aroclor-1016-5	(5)	6.11	6.12	6.02	6.22	0.01
Aroclor-1260-1	(1)	7.23	7.24	7.14	7.34	0.01
Aroclor-1260-2	(2)	7.48	7.49	7.39	7.59	0.01
Aroclor-1260-3	(3)	7.84	7.85	7.75	7.95	0.01
Aroclor-1260-4	(4)	8.07	8.07	7.97	8.17	0.01
Aroclor-1260-5	(5)	8.38	8.39	8.29	8.49	0.01
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.19	10.09	10.29	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: COPP02

Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409 SDG NO.: Q2409

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 19:30 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.88	4.87	4.77	4.97	-0.01
Aroclor-1016-2	(2)	4.90	4.89	4.79	4.99	-0.01
Aroclor-1016-3	(3)	5.08	5.06	4.96	5.16	-0.02
Aroclor-1016-4	(4)	5.12	5.10	5.00	5.20	-0.02
Aroclor-1016-5	(5)	5.33	5.32	5.22	5.42	-0.01
Aroclor-1260-1	(1)	6.37	6.35	6.25	6.45	-0.01
Aroclor-1260-2	(2)	6.55	6.54	6.44	6.64	-0.01
Aroclor-1260-3	(3)	6.71	6.69	6.59	6.79	-0.02
Aroclor-1260-4	(4)	7.18	7.16	7.06	7.26	-0.02
Aroclor-1260-5	(5)	7.42	7.40	7.30	7.50	-0.02
Tetrachloro-m-xylene		3.81	3.79	3.69	3.89	-0.02
Decachlorobiphenyl		8.81	8.80	8.70	8.90	-0.01