

CALIBRATION VERIFICATION SUMMARY

Contract: COPP02

Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409 SDG NO.: Q2409

Continuing Calib Date: 07/01/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 16:50 Initial Calibration Time(s): 15:52 16:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endrin	6.57	6.58	6.48	6.68	0.01
Methoxychlor	7.49	7.49	7.39	7.59	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: COPP02

Lab Code: CHEM **Case No.:** Q2409 **SAS No.:** Q2409 **SDG NO.:** Q2409

Continuing Calib Date: 07/01/2025 **Initial Calibration Date(s):** 06/17/2025 06/17/2025

Continuing Calib Time: 16:50 **Initial Calibration Time(s):** 15:52 16:47

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.07	8.07	7.97	8.17	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.08	4.08	3.98	4.18	0.00
Heptachlor epoxide	4.87	4.87	4.77	4.97	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Methoxychlor	6.75	6.76	6.66	6.86	0.01