

CALIBRATION VERIFICATION SUMMARY

Contract: COPP02

Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409 SDG NO.: Q2409

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 07/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD089277.D Time Analyzed: 16:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.073	8.972	9.172	48.690	50.000	-2.6
Endrin	6.574	6.475	6.675	47.220	50.000	-5.6
gamma-BHC (Lindane)	4.331	4.230	4.430	51.800	50.000	3.6
Heptachlor	4.930	4.829	5.029	50.200	50.000	0.4
Heptachlor epoxide	5.691	5.591	5.791	49.950	50.000	-0.1
Methoxychlor	7.493	7.393	7.593	50.180	50.000	0.4
Tetrachloro-m-xylene	3.550	3.450	3.650	51.880	50.000	3.8

CALIBRATION VERIFICATION SUMMARY

Contract: COPP02

Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409 SDG NO.: Q2409

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 07/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD089277.D Time Analyzed: 16:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.072	7.972	8.172	47.700	50.000	-4.6
Endrin	5.789	5.689	5.889	50.440	50.000	0.9
gamma-BHC (Lindane)	3.728	3.630	3.830	51.080	50.000	2.2
Heptachlor	4.083	3.983	4.183	49.000	50.000	-2.0
Heptachlor epoxide	4.871	4.773	4.973	50.730	50.000	1.5
Methoxychlor	6.754	6.655	6.855	51.030	50.000	2.1
Tetrachloro-m-xylene	2.881	2.780	2.980	53.140	50.000	6.3