

CALIBRATION VERIFICATION SUMMARY

Contract: COPP02

Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409 SDG NO.: Q2409

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 07/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD089289.D Time Analyzed: 19:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.072	8.972	9.172	46.970	50.000	-6.1
Endrin	6.575	6.475	6.675	46.290	50.000	-7.4
gamma-BHC (Lindane)	4.331	4.230	4.430	51.280	50.000	2.6
Heptachlor	4.930	4.829	5.029	49.310	50.000	-1.4
Heptachlor epoxide	5.692	5.591	5.791	48.920	50.000	-2.2
Methoxychlor	7.493	7.393	7.593	48.080	50.000	-3.8
Tetrachloro-m-xylene	3.550	3.450	3.650	51.730	50.000	3.5

CALIBRATION VERIFICATION SUMMARY

Contract: COPP02

Lab Code: CHEM Case No.: Q2409 SAS No.: Q2409 SDG NO.: Q2409

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 07/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD089289.D Time Analyzed: 19:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.072	7.972	8.172	44.010	50.000	-12.0
Endrin	5.789	5.689	5.889	48.800	50.000	-2.4
gamma-BHC (Lindane)	3.730	3.630	3.830	49.230	50.000	-1.5
Heptachlor	4.083	3.983	4.183	47.790	50.000	-4.4
Heptachlor epoxide	4.871	4.773	4.973	49.490	50.000	-1.0
Methoxychlor	6.753	6.655	6.855	48.100	50.000	-3.8
Tetrachloro-m-xylene	2.881	2.780	2.980	52.340	50.000	4.7